

## Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission  
Equipment Authorization Division, Application  
Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

October 6, 2008

TO WHOM IT MAY CONCERN

### Modular Approval Request

The *M3002-66480 WLAN Module* consists of a transmitter and a receiver in the 2.4 GHz and 5GHz band.

1. It does have its own RF shielding.
2. It has digital input buffers.
3. It has its own power supply regulation.
4. The module has unique antenna plugs to be used with approved antennas only.
5. The module has been tested in a stand-alone configuration.
6. The module will be labelled with its own FCC ID number.
7. The module complies with all applicable FCC rules. Instructions for maintaining compliance are provided in documents accompanying the host devices in which the module will be incorporated (Instructions for Use or Service Guide).
8. The module complies with applicable SAR and RF exposure requirements.

If you have any questions, please don't hesitate to contact us at the address shown below.

Sincerely,



Name: Andreas Suchi  
Company: Philips Medizin Systeme Böblingen GmbH  
Address: Hewlett-Packard-Str. 2  
  
Phone: +49 7031 463 1291  
Fax: +49 7031 463 2442  
Email: [andreas.suchi@philips.com](mailto:andreas.suchi@philips.com)

