

RF Exposure Statement

Philips Medical Systems hereby declares that the

Wireless Medical Telemetry Device,
M4841A (FCC ID: PQCM4841A)

is compliant to 47 CFR Part 95H, §95.1125 and §2.1093 of the FCC rules.

As per the FCC/TCB training of October 15, 2004, this statement of compliance is sufficient for devices with sourced based time average power of less than 1 mW. This applies to both TCB and FCC filings.

The output power for this device based on sourced based time average power is less than 0.1 mW. The duty cycle is 1/64.

Barry Wyshogrod, Regulatory Engineer
Signature already on-file at the FCC

Submitted 3-November-2004