

APPENDIX A: SAR TEST DATA

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Right Head, Slide In, Touch, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.03 V/m

Peak SAR (extrapolated) = 0.452 W/kg

SAR(1 g) = 0.326 mW/g; SAR(10 g) = 0.247 mW/g



0 dB = 0.361mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Right Head, Slide In, Tilt, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 9.56 V/m

Peak SAR (extrapolated) = 0.284 W/kg

SAR(1 g) = 0.227 mW/g; SAR(10 g) = 0.175 mW/g



0 dB = 0.248mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Right Head, Slide Out, Touch, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.21 V/m

Peak SAR (extrapolated) = 0.610 W/kg

SAR(1 g) = 0.496 mW/g; SAR(10 g) = 0.375 mW/g



0 dB = 0.538mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Right Head, Slide Out, Tilt, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 9.11 V/m

Peak SAR (extrapolated) = 0.341 W/kg

SAR(1 g) = 0.266 mW/g; SAR(10 g) = 0.199 mW/g



0 dB = 0.296mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Left Head, Slide In, Touch, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.47 V/m

Peak SAR (extrapolated) = 0.317 W/kg

SAR(1 g) = 0.247 mW/g; SAR(10 g) = 0.197 mW/g



0 dB = 0.267mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Left Head, Slide In, Tilt, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 11.7 V/m

Peak SAR (extrapolated) = 0.273 W/kg

SAR(1 g) = 0.218 mW/g; SAR(10 g) = 0.167 mW/g



0 dB = 0.240mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Left Head, Slide Out, Touch, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.92 V/m

Peak SAR (extrapolated) = 0.536 W/kg

SAR(1 g) = 0.442 mW/g; SAR(10 g) = 0.335 mW/g



0 dB = 0.481mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Left Head, Slide Out, Tilt, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 10.1 V/m

Peak SAR (extrapolated) = 0.314 W/kg

SAR(1 g) = 0.247 mW/g; SAR(10 g) = 0.186 mW/g



0 dB = 0.272mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Muscle ($\sigma = 0.94$ mho/m, $\epsilon_r = 53.12$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-26-2009; Ambient Temp: 22.6°C; Tissue Temp: 22.1 °C

Probe: EX3DV4 - SN3589; ConvF(8.32, 8.32, 8.32); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Body-Worn Data SAR, Back side, Slide In, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 23.6 V/m

Peak SAR (extrapolated) = 0.734 W/kg

SAR(1 g) = 0.532 mW/g; SAR(10 g) = 0.374 mW/g



0 dB = 0.601mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Muscle ($\sigma = 0.94$ mho/m, $\epsilon_r = 53.12$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-26-2009; Ambient Temp: 22.6°C; Tissue Temp: 22.1 °C

Probe: EX3DV4 - SN3589; ConvF(8.32, 8.32, 8.32); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Body-Worn Speech SAR, Back side, Slide In, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 22.1 V/m

Peak SAR (extrapolated) = 0.660 W/kg

SAR(1 g) = 0.440 mW/g; SAR(10 g) = 0.336 mW/g



0 dB = 0.543mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Right Head, Slide In, Touch, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.06 V/m

Peak SAR (extrapolated) = 0.505 W/kg

SAR(1 g) = 0.315 mW/g; SAR(10 g) = 0.191 mW/g



0 dB = 0.368mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Right Head, Slide In, Tilt, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 9.65 V/m

Peak SAR (extrapolated) = 0.296 W/kg

SAR(1 g) = 0.187 mW/g; SAR(10 g) = 0.114 mW/g



0 dB = 0.219mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934 FCC GSM; Type: 850CDMA/1900GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Right Head, Slide Out, Touch, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 4.11 V/m

Peak SAR (extrapolated) = 0.143 W/kg

SAR(1 g) = 0.083 mW/g; SAR(10 g) = 0.048 mW/g



0 dB = 0.100mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Right Head, Slide Out, Tilt, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.29 V/m

Peak SAR (extrapolated) = 0.136 W/kg

SAR(1 g) = 0.089 mW/g; SAR(10 g) = 0.055 mW/g



0 dB = 0.104mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Left Head, Slide In, Touch, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.77 V/m

Peak SAR (extrapolated) = 0.775 W/kg

SAR(1 g) = 0.474 mW/g; SAR(10 g) = 0.273 mW/g



0 dB = 0.568mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Left Head, Slide In, Tilt, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.93 V/m

Peak SAR (extrapolated) = 0.249 W/kg

SAR(1 g) = 0.165 mW/g; SAR(10 g) = 0.101 mW/g



0 dB = 0.187mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Left Head, Slide Out, Touch, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 4.34 V/m

Peak SAR (extrapolated) = 0.138 W/kg

SAR(1 g) = 0.082 mW/g; SAR(10 g) = 0.047 mW/g



0 dB = 0.098mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Left Head, Slide Out, Tilt, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.30 V/m

Peak SAR (extrapolated) = 0.111 W/kg

SAR(1 g) = 0.073 mW/g; SAR(10 g) = 0.045 mW/g



0 dB = 0.086mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900 GPRS; 2 Tx slots; Frequency: 1880 MHz; Duty Cycle: 1: 4.15

Medium: 1900 Muscle ($\sigma = 1.58$ mho/m, $\epsilon_r = 55.87$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-22-2009; Ambient Temp: 23.3°C; Tissue Temp: 22.5°C

Probe: EX3DV4 - SN3589; ConvF(6.88, 6.88, 6.88); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GPRS, Body-Worn Data SAR, Back side, Slide In, Mid.ch, 2 Tx Slots


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.92 V/m

Peak SAR (extrapolated) = 1.06 W/kg

SAR(1 g) = 0.617 mW/g; SAR(10 g) = 0.347 mW/g



0 dB = 0.759mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Muscle ($\sigma = 1.58$ mho/m, $\epsilon_r = 55.87$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-22-2009; Ambient Temp: 23.3°C; Tissue Temp: 22.5°C

Probe: EX3DV4 - SN3589; ConvF(6.88, 6.88, 6.88); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Body-Worn Speech SAR, Back side, Slide In, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 9.06 V/m

Peak SAR (extrapolated) = 0.649 W/kg

SAR(1 g) = 0.375 mW/g; SAR(10 g) = 0.209 mW/g



0 dB = 0.463mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Right Section

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Right Head, Slide Out, Touch, Mid.ch


Area Scan (7x12x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.21 V/m

Peak SAR (extrapolated) = 0.610 W/kg

SAR(1 g) = 0.496 mW/g; SAR(10 g) = 0.375 mW/g



PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0709

Communication System: Cellular CDMA; Frequency: 836.52 MHz; Duty Cycle: 1:1

Medium: 835 Muscle ($\sigma = 0.94$ mho/m, $\epsilon_r = 53.12$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.32, 8.32, 8.32); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: Cellular CDMA, Body-Worn Data SAR, Back side, Slide In, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 23.6 V/m

Peak SAR (extrapolated) = 0.734 W/kg

SAR(1 g) = 0.532 mW/g; SAR(10 g) = 0.374 mW/g



PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Left Section

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GSM, Left Head, Slide In, Touch, Mid.ch


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.77 V/m

Peak SAR (extrapolated) = 0.775 W/kg

SAR(1 g) = 0.474 mW/g; SAR(10 g) = 0.273 mW/g



PCTEST ENGINEERING LABORATORY, INC.

DUT: PTX-934; Type: Cellular CDMA and PCS GSM/GPRS Phone with Bluetooth; Serial: 0803

Communication System: GSM1900 GPRS; 2 Tx slots; Frequency: 1880 MHz; Duty Cycle: 1:4.15

Medium: 1900 Muscle ($\sigma = 1.58$ mho/m, $\epsilon_r = 55.87$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-22-2009; Ambient Temp: 23.3°C; Tissue Temp: 22.5°C

Probe: EX3DV4 - SN3589; ConvF(6.88, 6.88, 6.88); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS GPRS, Body-Worn Data SAR, Back side, Slide In, Mid.ch, 2 Tx Slots


Area Scan (7x10x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 6.92 V/m

Peak SAR (extrapolated) = 1.06 W/kg

SAR(1 g) = 0.617 mW/g; SAR(10 g) = 0.347 mW/g



APPENDIX B: DIPOLE VALIDATION

PCTEST ENGINEERING LABORATORY, INC.

DUT: SAR Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:4d026

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.89$ mho/m, $\epsilon_r = 39.58$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.5 cm

Test Date: 01-26-2009; Ambient Temp: 22.2°C; Tissue Temp: 21.8 °C

Probe: EX3DV4 - SN3589; ConvF(8.33, 8.33, 8.33); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Sub; Type: SAM 4.0; Serial: TP-1403

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

835MHz SAR Validation


Area Scan (7x13x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Input Power = 24.0 dBm (250 mW)

SAR(1 g) = 2.43 mW/g; SAR(10 g) = 1.58 mW/g

Deviation = 7.88 %



0 dB = 2.86mW/g

PCTEST ENGINEERING LABORATORY, INC.

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: 502

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: 1900 Brain ($\sigma = 1.38$ mho/m, $\epsilon_r = 39.81$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section; Space: 1.0 cm

Test Date: 01-22-2009; Ambient Temp: 23.1°C; Tissue Temp: 22.6 °C

Probe: EX3DV4 - SN3589; ConvF(7.5, 7.5, 7.5); Calibrated: 6/26/2008

Sensor-Surface: 3mm (Mechanical Surface Detection)

Electronics: DAE4 Sn704; Calibrated: 6/26/2008

Phantom: SAM Main; Type: SAM 4.0; Serial: TP-1406

Measurement SW: DASY4, V4.7 Build 71; Postprocessing SW: SEMCAD, V1.8 Build 172

1900 MHz SAR Validation


Area Scan (6x6x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Input Power = 20.0 dBm (100 mW)

SAR(1 g) = 4.03 mW/g; SAR(10 g) = 2.04 mW/g

Deviation = 1.00%



0 dB = 5.17mW/g

APPENDIX C: PROBE CALIBRATION



Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **PC Test**

Certificate No: **EX3-3589_Jun08**

CALIBRATION CERTIFICATE

Object **EX3DV4 - SN:3589**

Calibration procedure(s) **QA CAL-01.v6, QA CAL-14.v3 and QA CAL-23.v3
 Calibration procedure for dosimetric E-field probes**

Calibration date: **June 26, 2008**

Condition of the calibrated item **In Tolerance**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter E4419B	GB41293874	1-Apr-08 (No. 217-00788)	Apr-09
Power sensor E4412A	MY41495277	1-Apr-08 (No. 217-00788)	Apr-09
Power sensor E4412A	MY41498087	1-Apr-08 (No. 217-00788)	Apr-09
Reference 3 dB Attenuator	SN: S5054 (3c)	8-Aug-07 (No. 217-00719)	Aug-08
Reference 20 dB Attenuator	SN: S5086 (20b)	31-Mar-08 (No. 217-00787)	Apr-09
Reference 30 dB Attenuator	SN: S5129 (30b)	8-Aug-07 (No. 217-00720)	Aug-08
Reference Probe ES3DV2	SN: 3013	2-Jan-08 (No. ES3-3013_Jan08)	Jan-09
DAE4	SN: 660	3-Sep-07 (No. DAE4-660_Sep07)	Sep-08

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
RF generator HP 8648C	US3642U01700	4-Aug-99 (in house check Oct-07)	In house check: Oct-09
Network Analyzer HP 8753E	US37390585	18-Oct-01 (in house check Oct-07)	In house check: Oct-08

	Name	Function	Signature
Calibrated by:	Katja Pokovic	Technical Manager	
Approved by:	Niels Kuster	Quality Manager	

Issued: June 26, 2008

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
Polarization ϕ	ϕ rotation around probe axis
Polarization ϑ	ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}**: Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E²-field uncertainty inside TSL (see below *ConvF*).
- NORM(f)_{x,y,z}** = NORM_{x,y,z} * *frequency_response* (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCP_{x,y,z}**: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * *ConvF* whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Probe EX3DV4

SN:3589

Manufactured:	March 30, 2006
Last calibrated:	May 28, 2007
Recalibrated:	June 26, 2008

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

DASY - Parameters of Probe: EX3DV4 SN:3589

Sensitivity in Free Space^A

Diode Compression^B

NormX	0.47 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP X	96 mV
NormY	0.38 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP Y	96 mV
NormZ	0.37 ± 10.1%	$\mu\text{V}/(\text{V}/\text{m})^2$	DCP Z	96 mV

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL **900 MHz** **Typical SAR gradient: 5 % per mm**

Sensor Center to Phantom Surface Distance		2.0 mm	3.0 mm
SAR _{be} [%]	Without Correction Algorithm	10.3	5.7
SAR _{be} [%]	With Correction Algorithm	0.8	0.5

TSL **1810 MHz** **Typical SAR gradient: 10 % per mm**

Sensor Center to Phantom Surface Distance		2.0 mm	3.0 mm
SAR _{be} [%]	Without Correction Algorithm	7.7	3.6
SAR _{be} [%]	With Correction Algorithm	0.3	0.3

Sensor Offset

Probe Tip to Sensor Center **1.0 mm**


The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).




^B Numerical linearization parameter: uncertainty not required.

Frequency Response of E-Field


(TEM-Cell:ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ ($k=2$)

Receiving Pattern (ϕ), $\vartheta = 0^\circ$ Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)

Dynamic Range $f(\text{SAR}_{\text{head}})$ (Waveguide R22, $f = 1800 \text{ MHz}$)




Uncertainty of Linearity Assessment: $\pm 0.6\%$ ($k=2$)

Conversion Factor Assessment

f [MHz]	Validity [MHz] ^c	TSL	Permittivity	Conductivity	Alpha	Depth	ConvF Uncertainty
835	± 50 / ± 100	Head	41.5 ± 5%	0.90 ± 5%	0.81	0.67	8.33 ± 11.0% (k=2)
1810	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.51	0.70	7.50 ± 11.0% (k=2)
2300	± 50 / ± 100	Head	39.4 ± 5%	1.71 ± 5%	0.47	0.72	6.85 ± 11.0% (k=2)
2450	± 50 / ± 100	Head	39.2 ± 5%	1.80 ± 5%	0.69	0.61	6.58 ± 11.0% (k=2)
2600	± 50 / ± 100	Head	39.0 ± 5%	1.96 ± 5%	0.36	0.83	6.51 ± 11.0% (k=2)
4950	± 50 / ± 100	Head	36.3 ± 5%	4.40 ± 5%	0.30	1.70	4.72 ± 13.1% (k=2)
5200	± 50 / ± 100	Head	36.0 ± 5%	4.66 ± 5%	0.33	1.75	4.67 ± 13.1% (k=2)
5300	± 50 / ± 100	Head	35.9 ± 5%	4.76 ± 5%	0.36	1.75	4.51 ± 13.1% (k=2)
5500	± 50 / ± 100	Head	35.6 ± 5%	4.96 ± 5%	0.38	1.75	4.42 ± 13.1% (k=2)
5600	± 50 / ± 100	Head	35.5 ± 5%	5.07 ± 5%	0.38	1.75	4.45 ± 13.1% (k=2)
5800	± 50 / ± 100	Head	35.3 ± 5%	5.27 ± 5%	0.38	1.75	4.40 ± 13.1% (k=2)
835	± 50 / ± 100	Body	55.2 ± 5%	0.97 ± 5%	0.75	0.68	8.32 ± 11.0% (k=2)
1810	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.43	0.78	6.88 ± 11.0% (k=2)
2300	± 50 / ± 100	Body	52.8 ± 5%	1.85 ± 5%	0.43	0.88	6.51 ± 11.0% (k=2)
2450	± 50 / ± 100	Body	52.7 ± 5%	1.95 ± 5%	0.46	0.86	6.27 ± 11.0% (k=2)
2600	± 50 / ± 100	Body	52.5 ± 5%	2.16 ± 5%	0.64	0.69	6.24 ± 11.0% (k=2)
4950	± 50 / ± 100	Body	49.4 ± 5%	5.01 ± 5%	0.35	1.75	4.09 ± 13.1% (k=2)
5200	± 50 / ± 100	Body	49.0 ± 5%	5.30 ± 5%	0.34	1.75	4.16 ± 13.1% (k=2)
5300	± 50 / ± 100	Body	48.5 ± 5%	5.42 ± 5%	0.42	1.75	3.86 ± 13.1% (k=2)
5500	± 50 / ± 100	Body	48.6 ± 5%	5.65 ± 5%	0.38	1.75	4.13 ± 13.1% (k=2)
5600	± 50 / ± 100	Body	48.5 ± 5%	5.77 ± 5%	0.28	1.75	4.22 ± 13.1% (k=2)
5800	± 50 / ± 100	Body	48.2 ± 5%	6.00 ± 5%	0.38	1.75	4.09 ± 13.1% (k=2)

^c The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

Deviation from Isotropy in HSL
Error (ϕ , ϑ), $f = 900 \text{ MHz}$



Uncertainty of Spherical Isotropy Assessment: $\pm 2.6\%$ ($k=2$)