

Information about the Applicant	
Grantee (Company Name)	Boston Scientific Neuromodulation
FCC Contact Person	Kaoru Adair
Address	25155 Rye Canyon Loop
City, State, Zip	Valencia, CA 91355
Job Number	BOSN0011
Model	SC-5232
FCC ID	Q4D-SC5232
Agent	
Approval Type	Original
Equipment Class	DCD- Part 15 Low Power Transmitter Below 1705 kHz
Rule Part	15.209

Overview

This application is for the original approval of Boston Scientific Neuromodulation Corporation's model SC5232 Remote Control Device. The Remote Control Device (RC) is a component of their Precision Spectra Spinal Cord Stimulator and is used to adjust the stimulation parameters of the IPG (implanted device) or ETS (external trial simulator).

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Short term confidentiality requested for photos and manuals.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature



Dave Tolman, TCB Committee
6/12/2012