




LCIE SUD EST
Laboratoire de Moirans
Z.I. Centr'Alp
170, Rue de Chatagnon
38430 MOIRANS - FRANCE

PRODUCT LABELING

FCCID: Q6BSWSILBT01

1000097161

Échelle ... différer ... dessin original (ISO 5455)



Version	KROHNE V number	(1) Product name	(2) Product type	(3) Code number	(4) Serial no.	(5) Connection type	(6) MD	(7) FCC ID	(8) IC	(9) CMIIT	Applicable logos				
G 3/4 non SIL	VFLA300001	Liquid Level Switch	LLS 4000	084H6001	Several characters alphanumeric	G 3/4"	YYYY-MM (numeric)	Q6BSWSILBT01	1991D-SWSILBT01	t.b.d.	FC	SIL 2	30	ENE	PS
G 3/4 SIL	VFLA300002	Liquid Level Switch	LLS 4000	084H6002		G 3/4"					YES	NO	YES	YES	YES
NPT 3/4 non SIL	VFLA300003	Liquid Level Switch	LLS 4000U	084H6003		NPT 3/4"					YES	YES	YES	YES	YES
NPT 3/4 SIL	VFLA300004	Liquid Level Switch	LLS 4000U	084H6004		NPT 3/4"					YES	NO	YES	YES	YES
Electronic spare	VFLA300001	Electronic top part	LLS 4000/4000U	084H6010							YES	NO	YES	YES	YES

Not available

Device label specifications	
Type	Brady B-422
Color	White
Material	Polyester
Resolution	≥ 600 dpi
Ribbon	Serie R6000
Font	Myriad Pro Bold: product name, product type, code number Myriad Pro Regular: all others
Font size	10pt: product name 7pt: product type 5pt: made in, lbl. nr., address 6pt: all others
Logo	Minimum size: 6 mm Free space around: half size
QR-code	Left side: serial number Right side: Danfoss download center

Note: the label shall meet requirements of instructions 500B0292.

Rev	Mod	Nom	Cont	Norme	Homol.
-					
Engineer			Tolérances générales		
Prod			Bords de pièces		
Cont			Etat de surface		
Norme			Matière		
Homol.			4007282601		
			Ech		
			Folio 1/1		
KROHNE		Label		Code d'article 4007312201	
		Device label		Doc. type Doc. key Rev	
		LLS 4000/4000U		FZ 4007312201 -	
KSAS (Romans, F)		DANFOSS		Status concept 000	

Toute communication ou reproduction de ce document, sous quelque forme que se soit, et toute utilisation non autorisée sans la permission écrite de la LCIE SUD EST sont formellement interdites. Toute violation de ces droits est punie par la loi. Tous droits réservés pour le cas de la délivrance d'un modèle d'usage ou d'un modèle de présentation.



LCIE SUD EST
Laboratoire de Moirans
Z.I. Centr'Alp
170, Rue de Chatagnon
38430 MOIRANS - FRANCE

Liquid Level Switch

LLS 4000

084H6001

S/N: F19IT0000001
Supply: 24V DC $\approx$ 80mA
Connection type: G 3/4"
MWP (PS): 65 bars
Process Temp.: -50°C ... +120°C
MD: 2020/07 IP66/IP67
FCC ID: Q6BSWSILBT01 PV01
IC: 1991D-SWSILBT01



Danfoss

MADE IN FRANCE

FCC CE



Danfoss A/S, 6430 Nordborg, Denmark

lbl. nr.4007312201

Liquid Level Switch

LLS 4000U

084H6003


S/N: F19IT0000003
Supply: 24V DC $\approx$ 80mA
Connection type: NPT 3/4"
MWP (PS): 65 bars
Process Temp.: -50°C ... +120°C
MD: 2020/07 IP66/IP67
FCC ID: Q6BSWSILBT01 PV01
IC: 1991D-SWSILBT01



Danfoss

MADE IN FRANCE

FCC CE



Danfoss A/S, 6430 Nordborg, Denmark

lbl. nr.4007312201

Emplacement:

