

01 HAC RF_GSM850_GSM_Voice_Ch128_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 824.2 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch128/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 160.6 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 75.106 V/m; Power Drift = -0.03 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m


Grid 1 151.4 M3	Grid 2 155.0 M3	Grid 3 142.1 M4
Grid 4 157.2 M3	Grid 5 160.6 M3	Grid 6 146.8 M4
Grid 7 155.8 M3	Grid 8 159.3 M3	Grid 9 145.1 M4

Cursor:

Total = 160.6 V/m

E Category: M3

Location: 3.5, 4, 9.7 mm



0 dB = 160.6V/m

02 HAC RF_GSM850_GSM_Voice_Ch189_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 836.4 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch189/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 202.2 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 94.968 V/m; Power Drift = 0.06 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m


Grid 1 189.9 M3	Grid 2 193.7 M3	Grid 3 178.4 M3
Grid 4 196.1 M3	Grid 5 202.2 M3	Grid 6 184.7 M3
Grid 7 193.9 M3	Grid 8 200.5 M3	Grid 9 182.9 M3

Cursor:

Total = 202.2 V/m

E Category: M3

Location: 1, 4.5, 9.7 mm



0 dB = 202.2V/m

03 HAC RF_GSM850_GSM_Voice_Ch251_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 848.8 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch251/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 197.6 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 92.899 V/m; Power Drift = -0.030 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m


Grid 1 183.5 M3	Grid 2 189.6 M3	Grid 3 174.4 M3
Grid 4 192.3 M3	Grid 5 197.6 M3	Grid 6 181.5 M3
Grid 7 191.1 M3	Grid 8 196.4 M3	Grid 9 180.0 M3

Cursor:

Total = 197.6 V/m

E Category: M3

Location: 2.5, 3, 9.7 mm



0 dB = 197.6V/m

04 HAC RF_GSM1900_GSM_Voice_Ch512_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 1850.2 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch512/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 65.324 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 19.919 V/m; Power Drift = 0.07 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m



Grid 1 53.697 M3	Grid 2 45.127 M4	Grid 3 43.868 M4
Grid 4 52.337 M3	Grid 5 65.324 M3	Grid 6 64.773 M3
Grid 7 77.224 M3	Grid 8 82.562 M3	Grid 9 77.418 M3

Cursor:

Total = 82.562 V/m

E Category: M3

Location: 0, 25, 9.7 mm



0 dB = 82.560V/m

05 HAC RF_GSM1900_GSM_Voice_Ch661_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 1880 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch661/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 68.034 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 20.571 V/m; Power Drift = 0.0096 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m



Grid 1 54.055 M3	Grid 2 47.940 M3	Grid 3 44.769 M4
Grid 4 54.764 M3	Grid 5 68.034 M3	Grid 6 67.481 M3
Grid 7 80.915 M3	Grid 8 86.764 M2	Grid 9 81.823 M3

Cursor:

Total = 86.764 V/m

E Category: M2

Location: -0.5, 25, 9.7 mm



0 dB = 86.760V/m

06 HAC RF_GSM1900_GSM_Voice_Ch810_E

DUT: 331403-03A

Communication System: General GSM; Frequency: 1909.8 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch810/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 67.783 V/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 21.257 V/m; Power Drift = -0.10 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak E-field in V/m


Grid 1 53.455 M3	Grid 2 47.347 M3	Grid 3 48.054 M3
Grid 4 53.689 M3	Grid 5 67.783 M3	Grid 6 67.229 M3
Grid 7 76.300 M3	Grid 8 82.861 M3	Grid 9 78.674 M3

Cursor:

Total = 82.861 V/m

E Category: M3

Location: -1, 25, 9.7 mm



0 dB = 82.860V/m

07 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9262_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1852.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9262/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 28.312 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 25.844 V/m; Power Drift = -0.12 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m



Grid 1 21.948 M4	Grid 2 18.038 M4	Grid 3 18.812 M4
Grid 4 22.245 M4	Grid 5 28.312 M4	Grid 6 28.109 M4
Grid 7 31.997 M4	Grid 8 34.957 M4	Grid 9 32.895 M4

Cursor:

Total = 34.957 V/m

E Category: M4

Location: -0.5, 25, 9.7 mm



0 dB = 34.960V/m

08 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9400_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1880 MHz; Duty Cycle: 1:1
 Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³
 Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9400/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 22.159 V/m
 Probe Modulation Factor = 1.000
 Device Reference Point: 0, 0, -6.3 mm
 Reference Value = 19.720 V/m; Power Drift = 0.0076 dB



Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m

Grid 1 16.821 M4	Grid 2 14.329 M4	Grid 3 14.500 M4
Grid 4 17.458 M4	Grid 5 22.159 M4	Grid 6 22.003 M4
Grid 7 25.106 M4	Grid 8 27.511 M4	Grid 9 26.071 M4

Cursor:

Total = 27.511 V/m
 E Category: M4
 Location: -1, 25, 9.7 mm



0 dB = 27.510V/m

09 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9538_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1907.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9538/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 28.414 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 25.607 V/m; Power Drift = -0.04 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m


Grid 1 21.631 M4	Grid 2 18.700 M4	Grid 3 19.435 M4
Grid 4 21.962 M4	Grid 5 28.414 M4	Grid 6 28.275 M4
Grid 7 30.970 M4	Grid 8 34.827 M4	Grid 9 33.136 M4

Cursor:

Total = 34.827 V/m

E Category: M4

Location: -1, 25, 9.7 mm



0 dB = 34.830V/m

10 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4132_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 826.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4132/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 52.627 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 70.224 V/m; Power Drift = 0.18 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m


Grid 1 49.541 M4	Grid 2 50.860 M4	Grid 3 47.020 M4
Grid 4 51.222 M4	Grid 5 52.627 M4	Grid 6 48.444 M4
Grid 7 50.772 M4	Grid 8 52.179 M4	Grid 9 47.971 M4

Cursor:

Total = 52.627 V/m

E Category: M4

Location: 2, 3.5, 9.7 mm



0 dB = 52.630V/m

11 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4182_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 836.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4182/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 47.278 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 63.471 V/m; Power Drift = 0.06 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m


Grid 1 44.169 M4	Grid 2 45.625 M4	Grid 3 42.278 M4
Grid 4 45.755 M4	Grid 5 47.278 M4	Grid 6 43.657 M4
Grid 7 45.410 M4	Grid 8 46.919 M4	Grid 9 43.239 M4

Cursor:

Total = 47.278 V/m

E Category: M4

Location: 1.5, 4, 9.7 mm



0 dB = 47.280V/m

12 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4233_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 846.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4233/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 51.872 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 69.436 V/m; Power Drift = 0.05 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m


Grid 1 48.142 M4	Grid 2 49.855 M4	Grid 3 46.423 M4
Grid 4 50.027 M4	Grid 5 51.872 M4	Grid 6 48.178 M4
Grid 7 49.690 M4	Grid 8 51.493 M4	Grid 9 47.774 M4

Cursor:

Total = 51.872 V/m

E Category: M4

Location: 1, 4.5, 9.7 mm



0 dB = 51.870V/m

13 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1312_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1712.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1312/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 26.975 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 25.090 V/m; Power Drift = -0.02 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m



Grid 1 22.884 M4	Grid 2 17.211 M4	Grid 3 17.801 M4
Grid 4 22.549 M4	Grid 5 26.975 M4	Grid 6 26.651 M4
Grid 7 33.359 M4	Grid 8 34.614 M4	Grid 9 31.481 M4

Cursor:

Total = 34.614 V/m

E Category: M4

Location: 2, 25, 9.7 mm



0 dB = 34.610V/m

14 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1413_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1732.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1413/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 29.556 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 26.454 V/m; Power Drift = 0.03 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m



Grid 1 25.782 M4	Grid 2 19.041 M4	Grid 3 19.315 M4
Grid 4 24.161 M4	Grid 5 29.556 M4	Grid 6 29.259 M4
Grid 7 36.351 M4	Grid 8 37.957 M4	Grid 9 34.845 M4

Cursor:

Total = 37.957 V/m

E Category: M4

Location: 1.5, 25, 9.7 mm



0 dB = 37.960V/m

15 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1513_E

DUT: 331403-03A

Communication System: UMTS; Frequency: 1752.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1513/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 26.396 V/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 24.276 V/m; Power Drift = 0.06 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak E-field in V/m



Grid 1 22.347 M4	Grid 2 16.837 M4	Grid 3 17.401 M4
Grid 4 21.746 M4	Grid 5 26.396 M4	Grid 6 26.071 M4
Grid 7 31.856 M4	Grid 8 33.370 M4	Grid 9 30.611 M4

Cursor:

Total = 33.370 V/m

E Category: M4

Location: 1.5, 25, 9.7 mm



0 dB = 33.370V/m

16 HAC RF_GSM850_GSM_Voice_Ch128_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 824.2 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch128/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.337 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.060 A/m; Power Drift = -0.0027 dB

Hearing Aid Near-Field Category: M4 (AWF -5 dB)

Peak H-field in A/m


Grid 1 0.304 M4	Grid 2 0.205 M4	Grid 3 0.116 M4
Grid 4 0.312 M4	Grid 5 0.220 M4	Grid 6 0.132 M4
Grid 7 0.337 M4	Grid 8 0.236 M4	Grid 9 0.149 M4

Cursor:

Total = 0.337 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.340A/m

17 HAC RF_GSM850_GSM_Voice_Ch189_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 836.4 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch189/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.418 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.073 A/m; Power Drift = -0.14 dB

Hearing Aid Near-Field Category: M4 (AWF -5 dB)

Peak H-field in A/m


Grid 1 0.386 M4	Grid 2 0.255 M4	Grid 3 0.144 M4
Grid 4 0.389 M4	Grid 5 0.267 M4	Grid 6 0.158 M4
Grid 7 0.418 M4	Grid 8 0.284 M4	Grid 9 0.179 M4

Cursor:

Total = 0.418 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.420A/m

18 HAC RF_GSM850_GSM_Voice_Ch251_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 848.8 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch251/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.393 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.072 A/m; Power Drift = -0.045 dB

Hearing Aid Near-Field Category: M4 (AWF -5 dB)

Peak H-field in A/m


Grid 1 0.379 M4	Grid 2 0.241 M4	Grid 3 0.140 M4
Grid 4 0.378 M4	Grid 5 0.254 M4	Grid 6 0.151 M4
Grid 7 0.393 M4	Grid 8 0.271 M4	Grid 9 0.170 M4

Cursor:

Total = 0.393 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.390A/m

19 HAC RF_GSM1900_GSM_Voice_Ch512_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 1850.2 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch512/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.198 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.073 A/m; Power Drift = 0.04 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak H-field in A/m


Grid 1 0.178 M3	Grid 2 0.182 M3	Grid 3 0.173 M3
Grid 4 0.201 M3	Grid 5 0.189 M3	Grid 6 0.172 M3
Grid 7 0.260 M2	Grid 8 0.198 M3	Grid 9 0.152 M3

Cursor:

Total = 0.260 A/m

H Category: M2

Location: 25, 25, 10.2 mm



0 dB = 0.260A/m

20 HAC RF_GSM1900_GSM_Voice_Ch661_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 1880 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch661/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.210 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.074 A/m; Power Drift = 0.03 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak H-field in A/m



Grid 1 0.181 M3	Grid 2 0.186 M3	Grid 3 0.177 M3
Grid 4 0.211 M3	Grid 5 0.196 M3	Grid 6 0.177 M3
Grid 7 0.277 M2	Grid 8 0.210 M3	Grid 9 0.156 M3

Cursor:

Total = 0.277 A/m

H Category: M2

Location: 25, 25, 10.2 mm



0 dB = 0.280A/m

21 HAC RF_GSM1900_GSM_Voice_Ch810_H

DUT: 331403-03A

Communication System: General GSM; Frequency: 1909.8 MHz; Duty Cycle: 1:8

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch810/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.214 A/m

Probe Modulation Factor = 2.880

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.074 A/m; Power Drift = 0.01 dB

Hearing Aid Near-Field Category: M3 (AWF -5 dB)

Peak H-field in A/m


Grid 1 0.185 M3	Grid 2 0.191 M3	Grid 3 0.183 M3
Grid 4 0.212 M3	Grid 5 0.199 M3	Grid 6 0.182 M3
Grid 7 0.279 M2	Grid 8 0.214 M3	Grid 9 0.158 M3

Cursor:

Total = 0.279 A/m

H Category: M2

Location: 25, 25, 10.2 mm



0 dB = 0.280A/m

22 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9262_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1852.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9262/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.081 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.083 A/m; Power Drift = -0.08 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.072 M4	Grid 2 0.073 M4	Grid 3 0.069 M4
Grid 4 0.081 M4	Grid 5 0.077 M4	Grid 6 0.069 M4
Grid 7 0.104 M4	Grid 8 0.081 M4	Grid 9 0.062 M4

Cursor:

Total = 0.104 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.100A/m

23 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9400_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1880 MHz; Duty Cycle: 1:1
 Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³
 Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9400/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.063 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.065 A/m; Power Drift = -0.11 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.055 M4	Grid 2 0.057 M4	Grid 3 0.054 M4
Grid 4 0.063 M4	Grid 5 0.059 M4	Grid 6 0.054 M4
Grid 7 0.082 M4	Grid 8 0.063 M4	Grid 9 0.048 M4

Cursor:

Total = 0.082 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.080A/m

24 HAC RF_WCDMA Band II_RMC12.2Kbps_Ch9538_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1907.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch9538/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.082 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.081 A/m; Power Drift = -0.06 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.071 M4	Grid 2 0.073 M4	Grid 3 0.069 M4
Grid 4 0.080 M4	Grid 5 0.076 M4	Grid 6 0.068 M4
Grid 7 0.105 M4	Grid 8 0.082 M4	Grid 9 0.061 M4

Cursor:

Total = 0.105 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.110A/m

25 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4132_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 826.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4132/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.109 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.057 A/m; Power Drift = 0.05 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.099 M4	Grid 2 0.068 M4	Grid 3 0.039 M4
Grid 4 0.102 M4	Grid 5 0.073 M4	Grid 6 0.044 M4
Grid 7 0.109 M4	Grid 8 0.078 M4	Grid 9 0.050 M4

Cursor:

Total = 0.109 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.110A/m

26 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4182_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 836.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4182/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.096 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.049 A/m; Power Drift = 0.06 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.088 M4	Grid 2 0.059 M4	Grid 3 0.034 M4
Grid 4 0.088 M4	Grid 5 0.063 M4	Grid 6 0.037 M4
Grid 7 0.096 M4	Grid 8 0.067 M4	Grid 9 0.042 M4

Cursor:

Total = 0.096 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.100A/m

27 HAC RF_WCDMA Band V_RMC12.2Kbps_Ch4233_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 846.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch4233/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.106 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.054 A/m; Power Drift = 0.02 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.097 M4	Grid 2 0.066 M4	Grid 3 0.038 M4
Grid 4 0.098 M4	Grid 5 0.069 M4	Grid 6 0.041 M4
Grid 7 0.106 M4	Grid 8 0.075 M4	Grid 9 0.047 M4

Cursor:

Total = 0.106 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.110A/m

28 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1312_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1712.4 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1312/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.082 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.087 A/m; Power Drift = 0.17 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.081 M4	Grid 2 0.081 M4	Grid 3 0.073 M4
Grid 4 0.091 M4	Grid 5 0.082 M4	Grid 6 0.072 M4
Grid 7 0.106 M4	Grid 8 0.082 M4	Grid 9 0.065 M4

Cursor:

Total = 0.106 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.110A/m

29 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1413_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1732.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1413/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.089 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.096 A/m; Power Drift = 0.10 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.085 M4	Grid 2 0.086 M4	Grid 3 0.079 M4
Grid 4 0.094 M4	Grid 5 0.089 M4	Grid 6 0.079 M4
Grid 7 0.111 M4	Grid 8 0.089 M4	Grid 9 0.071 M4

Cursor:

Total = 0.111 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.110A/m

30 HAC RF_WCDMA Band IV_RMC12.2Kbps_Ch1513_H

DUT: 331403-03A

Communication System: UMTS; Frequency: 1752.6 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: H3DV6 - SN6300; ; Calibrated: 2012-12-12
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1210; Calibrated: 2013-6-19
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.4.5 (3634)

Ch1513/Hearing Aid Compatibility Test (101x101x1): Measurement grid: dx=5mm, dy=5mm

Maximum value of peak Total field = 0.076 A/m

Probe Modulation Factor = 1.000

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.083 A/m; Power Drift = 0.0098 dB

Hearing Aid Near-Field Category: M4 (AWF 0 dB)

Peak H-field in A/m


Grid 1 0.074 M4	Grid 2 0.074 M4	Grid 3 0.067 M4
Grid 4 0.082 M4	Grid 5 0.076 M4	Grid 6 0.067 M4
Grid 7 0.098 M4	Grid 8 0.076 M4	Grid 9 0.060 M4

Cursor:

Total = 0.098 A/m

H Category: M4

Location: 25, 25, 10.2 mm



0 dB = 0.100A/m