

For patient questions
please call: 866-426-4401

Medical Equipment
If found please call:
CardioNet, Inc.
San Diego, CA 92101
866-426-4404



CARDIONET MONITOR



S/N MNV11045



Warning:
Read Patient Information
Guide before use.

Protect from water and other liquids.

FCC ID: QBI-1001

Patent Nos. 5,959,529; 6,225,901


15 Volts DC, 1.2 amps

FCC ID NBZNRM-6832(cell modem)

RF ISM band 902 - 928 MHz

Revision number 220-0001-01, A
AECG Device Type I

4	3	2	1
ECO	APV	DWN	DESCRIPTION
			DATE
			REV



NOTES:

1. Manufacturing to print as per format shown, using printer ribbon P/N: 32123015.
2. Characters: black on white background
Font size:
 A 6 point
 A 5 point
3. Spacing between next smaller or larger font size to be 1.5mm, unless otherwise specified.
4. BOM to be as follows:
BOM Description
01016529 NRM-6832 OEM module
01016587 NRM-6832-M Minstrel III module

5. Electronic Identifier (EID): 00-60-D6-XX-XX-XX
- 6 hex ASCII digit pairs.
- First 3 are manufacturers ID.
- Last 3 pairs are the unit serial number.
- Only the last 3 pairs are encoded in the bar code.
6. Software revision format: per Novatel specification eg. TA01-O1 as provided by Novatel materials group.
7. Hardware revision format: PP.P-BB.B
PP.P: PCB assembly bill of materials revision level
BB.B: Top level bill of materials revision level
8. Temporary Deviation Notice (TDN) format: YYMMDD-X
YY= year, MM= month DD=day, X= unique identifier if required
9. Industry Canada ID TBD

 NOVATEL WIRELESS Novatel Wireless Technologies Ltd. Calgary, AB	NOTES:			
	Label Product FCC/EID Vespa			
	DWN: ?	SIZE	FSCM NO	DWG NO DP-25015379
	CHK: ?	SCALE	1 : 1	SHEET 1 OF 1

4

3

2

1

D

D

C

C

B

B

72.81

53.90

label as shown on DP-25015379

Label placed on non-connector side of the module

A

A

Novatel Wireless Technologies Ltd.
Calgary, AB

NOTES:

FCC label placement

DWN: ?

SIZE

FSCM NO

DWG NO

REV

CHK: ?

SCALE

2 : 1

SHEET

1 OF 1

4

3

2

1

For patient questions
please call: 866-426-4401

Medical Equipment
If found please call:
CardioNet, Inc.
San Diego, CA 92101
866-426-4404



CARDIONET MONITOR



S/N MNV1 1045



Warning:
Read Patient Information
Guide before use.

Protect from water and other liquids.

FCC ID: OBI-1001

Patent Nos. 5,959,529; 6,225,901

15 Volts DC, 1.2 amps

FCC ID NBZNRM-6832(cell modem)

RF ISM band 902 - 928 MHz.

Revision number 220-0001-01, A
AECG Device Type 1