



510 MARKET STREET
SAN DIEGO, CA 92101
UNITED STATES

TELEPHONE: 619.243.7500
FAX: 619.243.7700

WWW.CARDIONET.COM

June 12, 2002

Joe Dichoso
FCC Application Processing Branch

Re: FCC ID QBI-1001
Applicant: Cardionet
Correspondence Reference Number: 23094
731 Confirmation Number: EA405725

Dear Mr. Dichoso,

This is to follow up with your request for clarification with regards to the test data that has been submitted to you for review. Our Monitor product Model number 1001 uses the CDPD Wireless Radio Modem produced by Novatel Wireless, Inc. The documents you received for comparison, the Novatel Minstrel III and the Minstrel V both have the NRM6832 CDPD Modem housed in their respective enclosures so you could use any of the test data for comparison purposes. However, the Minstrel V is a later developed product so I would recommend you use it for comparing the SAR test results for the NRM6832 CDPD wireless modem which is the exact one we are presently using in our Monitor product. Although our Monitor can be used in the hand we consider it a body worn device which is carried in a fanny pack around the torso. Please find attached the photos and plots which were unreadable in the copied ITS Report.

Please feel free to contact me with any concerns or question you may have to assist in you approval process.

Best regards,



Gary Thomas
Sr. Hardware Design & System Engineer
510 Market Street
San Diego, CA 92101
619.243.7530
gthomas@cardionet.com



Incl: ITS SAR Test Plots



Plots.pdf



Plots 2.pdf



Photos.pdf