

Federal Communications Commission  
Equipment Authorization Branch  
7435 Oakland Mills Road  
Columbia, MD 21046

January 12, 2022

FCC IDs: QHJ-2100A Cover Letter

Dear Examiner:

The QHJ-2100 is referred to within engineering testing units as the following and these are electronically and mechanically equivalent:

Wireless Transmitter, Wolfgang Programmer Interface, RF02, & Model 2100 Wireless Transmitter

Wolfgang Programmer Interface radio description:

- Uses transceiver: Microsemi ZL70103
- Communication: MedRadio (402.45 to 404.55 MHz)
- Communicates with: Wolfgang Implantable Pulse Generator (Microsemi ZL70321)

Type of equipment approval: Original certification

Applicable rule parts:

- CFR Title 47, Part 95, Subpart I

Wolfgang Programmer Interface FCC ID: QHJ-2100A

“Seeking original certification”

Sincerely,



Chester Burress  
MicroTransponder, Inc.  
Sr. Director of Technical Operations, GMQ