

Declaration

We use DASY4 (software version 4.6) to perform SAR tests. During the "Area Scan" step the system automatically detects the maximum spot and secondary spots if existing. If the maximum of the secondary spots is less than 2 dB below the maximum spot and more than 0.05 W/kg, they will be evaluated during the fine scan step and reported.

Signature by:

(Deputy Director of the Laboratory)

Date:

2006/08/30