



1100 Falcon Ave.  
Glencoe, MN 55336

Tele: 320-864-4444  
Fax: 320-864-6611

*"Your Agent to Product Certifications and Licenses Throughout the World"*

To: Diane Poole

February 3, 2003

From: Duane R. Bagdons

Correspondence Reference Number: 24641  
731 confirmation Number: EA955974

I have just emailed into the file the following documents to satisfy this correspondence:

Lower Band Edge Compliance graphs  
Parts List  
Tune Up Information  
User Manual (see note below)  
SAR (see note below)

Regarding the User Manual, the customer is not going to have a specific manual just for the Wireless Adapter, they are going to use the NX Patient Station operations manual for both of these units. The wireless adapter is somewhat of a subset of the NX Patient Station in that it is the remote connection to the NX Patient Station. There are references in the manual to the Wireless Adapter. I have attached the manual for the NX Patient Station for your review.

Regarding the SAR Report, this is for Specific Absorption Rate to Human Tissue in cases of cell phones and devices that are close proximity to human tissue. The Wireless Adapter is not in this category. It is mounted to the chassis of a medical input peripheral and acts as the wireless link between the NX Patient Station and the medical input peripheral. The closest the wireless adapter would ever get to Human tissue is approximately 4 feet. I do not believe the SAR testing is required for such a device.

Thank you

Duane