

Eurofins Product Service GmbH
Storkower Straße 38c
15526 Reichenwalde
Germany

Tel +49 (0) 30 68905-1291
Fax +49 (0) 30 68905-1921
dirk.koenig@biotronik.com
www.biotronik.com

Berlin, 15. April 2025

Change to CardioMessenger Smart 4G

Herewith we like to inform about a modem change of BIOTRONIK's CardioMessenger Smart 4G. The change is driven by the discontinuity of the current modem. The current ME910C1-WW LTE modem (FCC ID: RI7ME910C1WW) used in the CardioMessenger Smart 4G will be replaced by ME910G1-WW modem (FCC ID: RI7ME910G1WW). The new modem is pin-compatible and supports the same frequency bands as the current modem. The only change is an updated parts list. There are no further changes to the PCB including the integrated antenna system.

Sincerely,

BIOTRONIK SE & Co. KG



Dirk König
Regulatory Affairs Senior Manager