

**Federal Communication Commission
Authorization and Evaluation Division
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046**

Tel +49 (0) 30 68905-1291
Fax +49 (0) 30 68905-961291
Dirk.koenig@biotronik.com
www.biotronik.com

Berlin, 5 December 2022

**Statement of re-use of original test results of approved radio modules
according to KDB 996369 D02 Module Q and A v01 Q1(a)**

Reference of new FCC ID application:

Applicant: BIOTRONIK SE & Co. KG	
FCC ID:	QRI-QCONNECT

We, the undersigned, hereby declare that our radio device hosts the following approved radio module(s)

Grantee: Intel Corporation			
Type of Module: AX210NGW			
FCC ID:	PD9AX210NG	Report Reference	200611-03.TR01, Rev 01 ; 200611-03.TR02 Rev 03 200611-03.TR03, Rev. 01 ; 200611-03.TR04 Rev 01 200611-03.TR06, rev 01

Grantee: Telit Communications S.p.A.			
Type of Module: LM940			
FCC ID:	RI7LM940	Report Reference	FW750208 Rev 01 ; FG750208B Rev 01

The module(s) is/are installed according to the installation instructions of the module manufacturer(s) without modification. The original module test report(s) accurately represent all conducted measured test results under the new conditions. All radiated measurements according to the applicable radio standard(s), the RF-exposure and simultaneous transmission evaluation are performed and reported for the entire host.

Sincerely,

BIOTRONIK SE & Co. KG


Dirk König
Regulatory Affairs Senior Manager