

**FCC TEST REPORT**

FCC 47 CFR Part 15C  
Industry Canada RSS-310  
License exempt radio equipment

**Report Reference No.** .....: G0M-1412-4360-TFC209LP-V02

**Testing Laboratory** .....: Eurofins Product Service GmbH

Address .....: Storkower Str. 38c  
15526 Reichenwalde  
Germany

Accreditation .....:



A2LA Accredited Testing Laboratory, Certificate No.: 1983.01  
FCC Filed Test Laboratory, Reg.-No.: 96970  
IC OATS Filing assigned code: 3470A

**Applicant's name** .....: Biotronik SE & Co. KG

Address .....: Woermannkehre 1  
12359 Berlin  
GERMANY

**Test specification:**

Standard .....: 47 CFR Part 15C  
RSS-310, Issue 3, 2010-12  
RSS-Gen, Issue 4, 2014-11  
ANSI C63.4:2009

**Equipment under test (EUT):**

Product description	ICD / Implantable Cardioverter Defibrillator	
Model No.	Inlexa 1 HF-T	
Additional Model(s)	Inlexa 1 DR-T (SN: 60508939) Inlexa 1 VR-T (SN: 60508940)	
Brand Name(s)	Inlexa 1	
Hardware version	Rev.: 01	
Firmware / Software version	ROM: 4.1 / RAM: 3.0 FCC-ID: QRITACHBORAX IC: 4708A-TACHBORAX	
<b>Test result</b>	<b>Passed</b>	

---

Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany

**Possible test case verdicts:**

- neither assessed nor tested .....: N/N
- required by standard but not appl. to test object.....: N/A
- required by standard but not tested.....: N/T
- not required by standard for the test object .....: N/R
- test object does meet the requirement.....: P (Pass)
- test object does not meet the requirement.....: F (Fail)

**Testing:**

Test Lab Temperature .....: 20 – 23 °C

Test Lab Humidity .....: 32 – 38 %

Date of receipt of test item .....: 2015-01-26

Date (s) of performance of tests .....: 2015-01-26

Compiled by .....: Wilfried Treffke

Tested by (+ signature).....: Wilfried Treffke  
(Responsible for Test)



Approved by (+ signature) .....: Christian Weber



Date of issue .....: 2016-02-12

Total number of pages .....: 29

**General remarks:**

**The test results presented in this report relate only to the object tested.**

**The results contained in this report reflect the results for this particular model and serial number. It is the responsibility of the manufacturer to ensure that all production models meet the intent of the requirements detailed within this report.**

This report shall not be reproduced, except in full, without the written approval of the Issuing testing laboratory.

**Additional comments :**

All devices feature the two RF-Telemetry functions Home Monitoring and wireless Wand.

RF-Telemetry functions are using the MICS-Band (402MHz – 405MHz).

A „-T“ inside the name of the device represents a device containing RF-Telemetry.

HF-T are triple-chamber devices.

DR-T are dual-chamber devices.

VR-T are single chamber devices without additional atrial detection.

All variants are available with DF-1.

All of these differences are only relevant in terms of medical aspects. They do not interfere the RF performance.

Evaluation measurements were performed for worst case antenna selection and the Inlexa 1 HF-T was selected. The model Inlexa 1 HF-T, as the most complex model, was selected for the measurements.


**Tach\_Borax Family Explanation (G0M-1412-4360)**
**1. Family Letter**

	Product Name	Type	no. of chambers	Connector	max.stored energy	SN
1	Inlexa 1 HF-T	CRT	3	DF-1	40J	60508938 (Master)
2	Inlexa 1 DR-T	DR	2	DF-1	40J	60508939
3	Inlexa 1 VR-T	VR	1	DF-1	40J	60508940

**2. Family description**
**2.1 PC-Board**


All family devices are using the same electronic. This means all active and all passive electrical components are the same.

**PC Board (4190)**  
**10pol feedtrough**  
**Schematic file SCH-0160\_01.pdf**


**2.2 RF-Antenna**

**The family members are equipped with the identical RF antenna.**

*DF-1 header with antenna*



*Signature:*

*Date: 2/12/2016*

**Mark Briesemeister**  
**Junior Manager Regulatory Affairs**  
**BIOTRONIK SE & Co. KG**  
**Woermannkehre 1**  
**12359 Berlin**  
**Germany**

## Version History

<b>Version</b>	<b>Issue Date</b>	<b>Remarks</b>	<b>Revised by</b>
01	2015-01-30	Initial Release	
02	2016-02-12	Certification numbers corrected	C. Weber


## REPORT INDEX

<b>1 EQUIPMENT (TEST ITEM) DESCRIPTION</b>	<b>7</b>
1.1 Photos – Equipment External	8
1.2 Photos – Equipment internal	11
1.3 Photos – Test setup	13
1.4 Supporting Equipment Used During Testing	14
1.5 Test Modes	15
1.6 Test Equipment Used During Testing	16
1.7 Sample emission level calculation	17
1.8 Simulated human body	18
<b>2 RESULT SUMMARY</b>	<b>19</b>
<b>3 TEST CONDITIONS AND RESULTS</b>	<b>20</b>
3.1 Test Conditions and Results – Occupied Bandwidth	20
3.2 Test Conditions and Results – Fundamental field strength emissions	22
3.4 Test Conditions and Results – Receiver radiated emissions	24
ANNEX A Transmitter radiated spurious emissions	26
ANNEX B Receiver radiated spurious emissions	28

## 1 Equipment (Test item) Description

<b>Description</b>	ICD / Implantable Cardioverter Defibrillator	
<b>Model</b>	Inlexa 1 HF-T	
<b>Additional Model(s)</b>	additional models according to family letter	
<b>Brand Name(s)</b>	Inlexa 1	
<b>Serial number</b>	60508938	
<b>Hardware version</b>	Rev.: 01	
<b>Software / Firmware version</b>	ROM: 4.1 / RAM: 3.0	
<b>FCC-ID</b>	QRITACHBORAX	
<b>IC</b>	4708A-TACHBORAX	
<b>Equipment type</b>	End product	
<b>Radio type</b>	Transceiver / Inductive Loop Coil Transmitter	
<b>Radio technology</b>	ULP-AMI	
<b>Operating frequency range</b>	64 kHz	
<b>Frequency range</b>	$F_{MID}$	64 kHz
<b>Modulations</b>	OOK	
<b>Number of channels</b>	1	
<b>Channel spacing</b>	None	
<b>Number of antennas</b>	1	
<b>Antenna</b>	Type	integrated
	Model	Tach_Borax Coil
	Manufacturer	Biotronik SE & Co. KG
	Gain	unspecified
<b>Manufacturer</b>	Biotronik SE & Co. KG Woermannkehre 1 12359 Berlin GERMANY	
<b>Power supply</b>	$V_{NOM}$	3.0 VDC (Lithium-Battery)
	$V_{MIN}$	N/A
	$V_{MAX}$	N/A
<b>AC/DC-Adaptor</b>	Model	N/A
	Vendor	N/A
	Input	N/A
	Output	N/A

## 1.1 Photos – Equipment External



---

Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany

**VARIANT Inlexa 1 VR-T****VARIANT Inlexa 1 DR-T**

**AE WIRELESS WAND TelBoxII**



---

Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany

## 1.2 Photos – Equipment internal




---


Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany



### 1.3 Photos – Test setup



#### 1.4 Supporting Equipment Used During Testing

Product Type*	Device	Manufacturer	Model No.	Comments
AE	Enginnering communication box	BIOTRONIK	TelBoxII	for test mode
AE	Electrode	Biotronik	Linox Smart	-

**\*Note:** Use the following abbreviations:

AE : Auxiliary/Associated Equipment, or

SIM : Simulator (Not Subjected to Test)

CABL : Connecting cables

### 1.5 Test Modes

Mode #	Description	
Single	General conditions:	EUT powered by fully charged battery
	Radio conditions:	Mode = standalone transmit Modulation = OOK Power level = Maximum
Receive	General conditions:	EUT powered by fully charged battery
	Radio conditions:	Mode = standalone receive Modulation = OOK

### 1.6 Test Equipment Used During Testing

<b>Measurement Software</b>					
Description	Manufacturer	Name		Version	
EMC Test Software	Dare Instruments	Radimation		2014-01-15	

<b>Occupied Bandwidth</b>					
Description	Manufacturer	Model	Identifier	Cal. Date	Cal. Due
Spectrum Analyzer	R&S	FSP 30	EF00312	2014-02	2015-02

<b>Field strength emissions</b>					
Description	Manufacturer	Model	Identifier	Cal. Date	Cal. Due
Semi-anechoic chamber	Frankonia	AC 5	EF00395	-	-
Spectrum Analyzer	R&S	FSIQ26	EF00242	2014-03	2015-03
Biconical Antenna	R&S	HK 116	EF00012	2013-02	2016-02
LPD antenna	R&S	HL 223	EF00212	2013-02	2016-02
LPD Antenna	R&S	HL 025	EF00327	2013-02	2016-02

## 1.7 Sample emission level calculation

The following is a description of terms and a sample calculation, as appears in the radiated emissions data table. The numbers used in the calculation are for example only. There is no direct correlation to the specific data taken for the product described in this document:

Reading:

This is the reading obtained on the spectrum analyzer in dB $\mu$ V. Any external preamplifiers used are taken into account through internal analyzer settings.

A.F.:

This is the antenna factor for the receiving antenna. It is a conversion factor, which converts electric fields strengths to voltages, which can be measured directly on the spectrum analyzer. It is treated as a loss in dB. Cable losses have been included with the A.F. to simplify the calculations. The antenna factor is used in calculations as follows:

$$\text{Reading on Analyzer (dB}\mu\text{V)} + \text{A.F. (dB)} = \text{Net field strength (dB}\mu\text{V/m)}$$

Net:

This is the net field strength measurement (as shown above).

Limit:

This is the FCC Class B radiated emission limit (in units of dB $\mu$ V/m). The FCC limits are given in units of  $\mu$ V/m. The following formula is used to convert the units of  $\mu$ V/m to dB $\mu$ V/m:

$$\text{Limit (dB}\mu\text{V/m)} = 20 * \log (\mu\text{V/m})$$

Margin:

This is the margin of compliance below the FCC limit. The units are given in dB. A negative margin indicates the emission was below the limit. A positive margin indicates that the emission exceeds the limit.

Example only:

Reading + AF =	Net Reading :	Net reading - FCC limit = Margin
21.5 dB $\mu$ V + 26 dB =	47.5 dB $\mu$ V/m :	47.5 dB $\mu$ V/m - 57.0 dB $\mu$ V/m = -9.5 dB

### 1.8 Simulated human body

For radiated tests the implant was placed in a simulated human body.

Liquid components	
Component	percentage per weight
Deionized water	52.4
Bactericide	0.08
Hydroxy ethyl cellulose (HCE)	1.0
Sodium chloride	1.4
Sucrose	45.0

Measured tissue parameters:


Tissue parameters – 403.5MHz			
Component	Target	Measured	Tolerance [%]
Dielectric constant $\epsilon$	62.5	63.01	0.82
Conductivity $\sigma$ [ms/cm]	9.0	8.9	-1.11

## 2 Result Summary

FCC 47 CFR Part 15C, IC RSS-310				
Product Specific Standard Section	Requirement – Test	Reference Method	Result	Remarks
RSS-Gen 6.6	Occupied Bandwidth	RSS-Gen 6.6	N/R	Informational only
FCC 15.201(a), FCC 15.209 IC RSS-310 3.7	Field strength emissions	ANSI C63.4	PASS	
IC RSS-310 2.3 IC RSS-Gen 7.1	Receiver radiated spurious emissions	ANSI C63.4	PASS	
<b>Remarks:</b>				

### 3 Test Conditions and Results


#### 3.1 Test Conditions and Results – Occupied Bandwidth

Occupied Bandwidth acc. to IC RSS-Gen		Verdict: PASS
Test according to measurement reference	Reference Method	
	RSS-Gen 6.6	
Test frequency range	Tested frequencies	
	$F_{MID}$	
EUT test mode	Single	
<b>Limits</b>		
None (Informational only)		
<b>Test setup</b>		
 <pre> graph LR     SA[Spectrum Analyzer] --- EUT[EUT]   </pre>		
<b>Test procedure</b>		
<ol style="list-style-type: none"> <li>1. EUT set to test mode (Communication tester is used if needed)</li> <li>2. Span set to at least twice the emission spectrum</li> <li>3. Resolution bandwidth set to 1 % of span</li> <li>4. Occupied Bandwidth (99 %) measurement with spectrum analyzer built in measurement function</li> </ol>		
<b>Test results</b>		
Channel	Frequency [kHz]	Occupied Bandwidth [kHz]
$F_{MID}$	64	3.5
Comments: Measurement is applicable to all variants		

**Occupied Bandwidth -  $F_{MID}$** 
**Occupied Bandwidth acc. to RSS-Gen**

Project Number: G0M-1412-4360

Applicant: Biotronic SE & Co.KG  
 EUT Name: ICD / Implantable Cardioverter Defibrillator  
 Model: Inlexa 1  
 Test Site: Eurofins Product Service GmbH  
 Operator: Wilfried Treffke  
 Test Conditions: Tnom / Vnom  
 Mode: Tx 64 kHz  
 Test Date: 2015-01-28  
 Verdict: NONE (INFORMATION ONLY)  
 Note 1: A spectrum analyzer with an integrated 99% power bandwidth function is used  
 Note 2: Near-field measurement test fixture / 64 kHz system



Date: 28.JAN.2015 11:27:28


Test Report No.: G0M-1412-4360-TFC209LP-V02

 Eurofins Product Service GmbH  
 Storkower Str. 38c, D-15526 Reichenwalde, Germany


### 3.2 Test Conditions and Results – Fundamental field strength emissions

<b>Field strength emissions acc. to FCC 47 CFR 15.201 / IC RSS-310</b>				<b>Verdict: PASS</b>
Test according referenced standards		Reference Method		
		FCC 15.201(a) + 15.209 / IC RSS-310 3.7		
Test according to measurement reference		Reference Method		
		ANSI C63.4		
Test frequency range		Tested frequencies		
		9 kHz – 10 <sup>th</sup> Harmonic		
EUT test mode		Single		
<b>Limits</b>				
Frequency range [MHz]	Detector	Limit [ $\mu$ V/m]	Limit [dB $\mu$ V/m]	Limit Distance [m]
0.009 – 0.490	Quasi-Peak	2400/F[kHz]	48.5 – 13.8	300
0.490 – 1.705	Quasi-Peak	2400/F[kHz]	13.8 – 1.4	30
1.705 – 30	Quasi-Peak	30	29.5	30
30 – 88	Quasi-Peak	100	40	3
88 – 216	Quasi-Peak	150	43.5	3
216 – 960	Quasi-Peak	200	46	3
960 – 1000	Quasi-Peak	500	54	3
> 1000	Average	500	54	3

The emission limits shown in the above table are based on measurements employing a CISPR quasi-peak detector except for the frequency bands 9–90 kHz, 110–490 kHz and above 1000 MHz. Radiated emission limits in these three bands are based on measurements employing an average detector.

Test setup								
								
Test procedure								
<ol style="list-style-type: none"> <li>1. EUT set to test mode</li> <li>2. Span it set according to measurement range</li> <li>3. Resolution bandwidth below 1 GHz is set according to CISPR 16 with peak/quasi-peak detector and RBW of 1 MHz with peak/average detector is used above 1 GHz</li> <li>4. Markers are set to maximum emission levels</li> </ol>								
Test results								
Channel	Frequency [kHz]	Emission [kHz]	Level [dB $\mu$ V/m]	Detector	Pol.	Limit [dB $\mu$ V/m]	Limit distance [m]*	Margin [dB]
F <sub>MID</sub>	64	100.36	-58.90	pk	ver	27.60	300	-86.50
F <sub>MID</sub>	64	48.528	-63.5	av	ver	33.9	300	-97.40
Comments: * Physical distance between EUT and measurement antenna.								

### 3.4 Test Conditions and Results – Receiver radiated emissions

Receiver radiated emissions acc. to IC RSS-310		Verdict: PASS		
Test according referenced standards		Reference Method		
		IC RSS-310 3.7		
Test according to measurement reference		Reference Method		
		ANSI C63.4		
Test frequency range		Tested frequencies		
		9 kHz – 10 <sup>th</sup> Harmonic		
EUT test mode		Receive		
Limits				
Frequency range [MHz]	Detector	Limit [ $\mu$ V/m]	Limit [dB $\mu$ V/m]	Limit Distance [m]
0.009 – 0.490	Quasi-Peak	2400/F[kHz]	48.5 – 13.8	300
0.490 – 1.705	Quasi-Peak	2400/F[kHz]	13.8 – 1.4	30
1.705 – 30	Quasi-Peak	30	29.5	30
30 – 88	Quasi-Peak	100	40	3
88 – 216	Quasi-Peak	150	43.5	3
216 – 960	Quasi-Peak	200	46	3
960 – 1000	Quasi-Peak	500	54	3
> 1000	Average	500	54	3
Test setup				
				

Test procedure						
1. EUT set to receive mode (Communication tester is used if needed) 2. Span it set according to measurement range 3. Resolution bandwidth below 1 GHz is set according to CISPR 16 with peak/quasi-peak detector and RBW of 1 MHz with peak/average detector is used above 1 GHz 4. Markers are set to peak emission levels						
Test results						
Channel	Frequency [kHz]	Emission [kHz]	Emission Level [dB $\mu$ V/m]	Det.	Limit [dBd $\mu$ V/m]	Margin [dB $\mu$ V/m]
F <sub>MID</sub>	64	94.68	-62.90	pk	28.1	-91.0 dB
F <sub>MID</sub>	64	64.40	-66.50	av	31.4	-97.9 dB
F <sub>MID</sub>	64	417.00	-65.10	av	15.2	-80.3 dB
Comments:						


## ANNEX A Transmitter radiated spurious emissions

### Spurious emissions according to FCC 15.209

Project number: G0M-1412-4360

Applicant: BIOTRONIK SE & Co.KG  
 EUT Name: ICD / Implantable Cardioverter Defibrillator  
 Model: Inlexa 1  
 Test Site: Eurofins Product Service GmbH  
 Operator: Treffke  
 Test Conditions:  $T_{nom} = 25^{\circ}\text{C}$ ,  $V_{nom} = 3.0 \text{ VDC}$  lithium battery  
 Antenna: Rohde & Schwarz HFH 2-Z2  
 Measurement distance: 3 m converted to 300 m  
 Mode: TX; 64 kHz  
 Test Date: 2015-01-26  
 Note:

Index 7




Frequency	Peak	Peak Limit	Peak Difference	Peak Status
100.36 kHz	-58.9 dB $\mu$ V/m	27.6 dB $\mu$ V/m	-86.48 dB	Pass
48.528 kHz	Average -63.5 dB $\mu$ V/m	Average Limit 33.9 dB $\mu$ V/m	Average Difference -97.41 dB	Average Status Pass

**Spurious emissions according to FCC 15.209**

Project number: G0M-1412-4360

Applicant: BIOTRONIK SE & Co.KG  
EUT Name: ICD / Implantable Cardioverter Defibrillator  
Model: Inlexa 1  
Test Site: Eurofins Product Service GmbH  
Operator: Treffke  
Test Conditions:  $T_{nom}$ : 25°C,  $V_{nom}$ : 3.0 VDC lithium battery  
Antenna: Rohde & Schwarz HFH 2-Z2  
Measurement distance: 3 m converted to 30 m  
Mode: TX; 64 kHz  
Test Date: 2015-01-26  
Note:

Index 8



---

Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany


## ANNEX B Receiver radiated spurious emissions

### Spurious emissions according to RSS-Gen

Project number: G0M-1412-4360

Applicant: BIOTRONIK SE & Co.KG  
 EUT Name: ICD / Implantable Cardioverter Defibrillator  
 Model: Inlexa 1  
 Test Site: Eurofins Product Service GmbH  
 Operator: Treffke  
 Test Conditions:  $T_{nom} = 25^{\circ}\text{C}$ ,  $V_{nom} = 3.0 \text{ VDC}$  lithium battery  
 Antenna: Rohde & Schwarz HFH 2-Z2  
 Measurement distance: 3 m converted to 300 m  
 Mode: RX; 64 kHz  
 Test Date: 2015-01-26  
 Note:

Index 12




Frequency	Peak	Peak Limit	Peak Difference	Peak Status
94.68 kHz	-62.9 dB $\mu$ V/m	28.1 dB $\mu$ V/m	-90.96 dB	Pass
64.404 kHz	Average -66.5 dB $\mu$ V/m	Average Limit 31.4 dB $\mu$ V/m	Average Difference -97.89 dB	Average Status Pass
417 kHz	-65.1 dB $\mu$ V/m	15.2 dB $\mu$ V/m	-80.27 dB	Pass

**Spurious emissions according to RSS-Gen**

Project number: G0M-1412-4360

Applicant: BIOTRONIK SE & Co.KG  
EUT Name: ICD / Implantable Cardioverter Defibrillator  
Model: Inlexa 1  
Test Site: Eurofins Product Service GmbH  
Operator: Treffke  
Test Conditions:  $T_{nom}$ : 25°C,  $V_{nom}$ : 3.0 VDC lithium battery  
Antenna: Rohde & Schwarz HFH 2-Z2  
Measurement distance: 3 m converted to 30 m  
Mode: RX; 64 kHz  
Test Date: 2015-01-26  
Note:

Index 13



---

Test Report No.: G0M-1412-4360-TFC209LP-V02

---

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany