

## **APPENDIX C – DIPOLE VALIDATION PLOTS**

## Dipole 900 MHz

SAM I Phantom, Flat Section; Position: (90°,90°); Frequency: 900 MHz

Probe: ET3DV6 - SN1609; ConvF(6.62,6.62,6.62); Crest factor: 1.0; Brain 900 MHz:  $\sigma = 0.96 \text{ mho/m}$ ,  $\epsilon_r = 41.8$ ,  $\rho = 1.00 \text{ g/cm}^3$

Cubes (2): SAR (1g):  $11.4 \text{ mW/g} \pm 0.04 \text{ dB}$ , SAR (10g):  $7.14 \text{ mW/g} \pm 0.04 \text{ dB}$

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Peak:  $18.6 \text{ mW/g} \pm 0.03 \text{ dB}$ ; Powerdrift: 0.04 dB

Comment :

Dipole Validation (D900V2/ S.N: 112)


900MHz Brain

Antenna Input Power: 30dBm (1 W)

HCT Co., Ltd. Brain Tissue Simulating Liquid

Liquid Temperature : 21.7 °C

Date Tested : March 24, 2004



## Dipole 900 MHz

SAM I Phantom; Flat Section; Position: (90°,90°); Frequency: 900 MHz

Probe: ET3DV6 - SN1609; ConvF(6.62,6.62,6.62); Crest factor: 1.0; Brain 900 MHz:  $\sigma = 0.96 \text{ mho/m}$ ,  $\epsilon_r = 41.8$ ,  $\rho = 1.00 \text{ g/cm}^3$

Cubes (2): SAR (1g):  $11.4 \text{ mW/g} \pm 0.04 \text{ dB}$ , SAR (10g):  $7.14 \text{ mW/g} \pm 0.04 \text{ dB}$

Cube 5x5x7: Dx = 8.0, Dy = 8.0, Dz = 5.0

Peak:  $18.6 \text{ mW/g} \pm 0.03 \text{ dB}$

Comment :

Dipole Validation (D900V2/ S.N: 112)


900MHz Brain

Antenna Input Power: 30dBm (1 W)

HCT Co., Ltd. Brain Tissue Simulating Liquid

Liquid Temperature : 21.7 °C

Date Tested : March 24, 2004



# Title : HSD8000

## SubTitle : 900 MHz Brain

March 24, 2004 08:40 AM

| Frequency        | e'      | e''     |
|------------------|---------|---------|
| 800.000000 MHz   | 43.0676 | 19.4699 |
| 810.000000 MHz   | 42.9497 | 19.4365 |
| 820.000000 MHz   | 42.8088 | 19.4215 |
| 830.000000 MHz   | 42.7005 | 19.4302 |
| 840.000000 MHz   | 42.5667 | 19.3729 |
| 850.000000 MHz   | 42.4730 | 19.3683 |
| 860.000000 MHz   | 42.3363 | 19.3459 |
| 870.000000 MHz   | 42.2292 | 19.3057 |
| 880.000000 MHz   | 42.1283 | 19.2750 |
| 890.000000 MHz   | 41.9664 | 19.2638 |
| 900.000000 MHz   | 41.7769 | 19.2125 |
| 910.000000 MHz   | 41.7106 | 19.1948 |
| 920.000000 MHz   | 41.5515 | 19.2160 |
| 930.000000 MHz   | 41.4190 | 19.1521 |
| 940.000000 MHz   | 41.3160 | 19.1722 |
| 950.000000 MHz   | 41.1626 | 19.1417 |
| 960.000000 MHz   | 41.0546 | 19.0937 |
| 970.000000 MHz   | 40.9010 | 19.0992 |
| 980.000000 MHz   | 40.7631 | 19.0541 |
| 990.000000 MHz   | 40.6369 | 19.0069 |
| 1.0000000000 GHz | 40.5755 | 19.0086 |

# Title : HSD8000

## SubTitle : 900 MHz Body

March 24, 2004 09:12 AM

| Frequency      | e'      | e''     |
|----------------|---------|---------|
| 850.000000 MHz | 56.4906 | 20.6000 |
| 855.000000 MHz | 56.4602 | 20.5785 |
| 860.000000 MHz | 56.4532 | 20.5587 |
| 865.000000 MHz | 56.4634 | 20.6077 |
| 870.000000 MHz | 56.3535 | 20.5962 |
| 875.000000 MHz | 56.3469 | 20.5974 |
| 880.000000 MHz | 56.3612 | 20.5953 |
| 885.000000 MHz | 56.3008 | 20.5912 |
| 890.000000 MHz | 56.2615 | 20.5840 |
| 895.000000 MHz | 56.1923 | 20.5785 |
| 900.000000 MHz | 56.1523 | 20.4876 |
| 905.000000 MHz | 56.0948 | 20.5305 |
| 910.000000 MHz | 56.0659 | 20.4881 |
| 915.000000 MHz | 55.9860 | 20.4466 |
| 920.000000 MHz | 55.9059 | 20.4265 |
| 925.000000 MHz | 55.8393 | 20.4114 |
| 930.000000 MHz | 55.7977 | 20.3553 |
| 935.000000 MHz | 55.7523 | 20.3730 |
| 940.000000 MHz | 55.6561 | 20.3409 |
| 945.000000 MHz | 55.5434 | 20.3623 |
| 950.000000 MHz | 55.4615 | 20.3809 |

## APPENDIX D – PROBE CALIBRATION DATA

---

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

Client **Hyundai CT (Dymstec)**

**CALIBRATION CERTIFICATE**

Object(s) **ET3DV6 - SN:1609**

Calibration procedure(s) **QA CAL-01.v2  
Calibration procedure for dosimetric E-field probes**

Calibration date: **January 22, 2004**

Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $22 \pm 2$  degrees Celsius and humidity  $< 75\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Model Type                        | ID #           | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|-----------------------------------|----------------|-------------------------------------------|------------------------|
| Power meter EPM E4419B            | GB41293874     | 2-Apr-03 (METAS, No 252-0250)             | Apr-04                 |
| Power sensor E4412A               | MY41495277     | 2-Apr-03 (METAS, No 252-0250)             | Apr-04                 |
| Reference 20 dB Attenuator        | SN: 5086 (20b) | 3-Apr-03 (METAS, No. 251-0340)            | Apr-04                 |
| Fluke Process Calibrator Type 702 | SN: 6295803    | 8-Sep-03 (Sintrol SCS No. E-030020)       | Sep-04                 |
| Power sensor HP 8481A             | MY41092180     | 18-Sep-02 (SPEAG, in house check Oct-03)  | In house check: Oct 05 |
| RF generator HP 8684C             | US3642U01700   | 4-Aug-99 (SPEAG, in house check Aug-02)   | In house check: Aug-05 |
| Network Analyzer HP 8753E         | US37390585     | 18-Oct-01 (SPEAG, in house check Oct-03)  | In house check: Oct 05 |

Calibrated by: **Nico Vetterli** **Technician** **Signature**

Approved by: **Katja Pokovic** **Laboratory Director**

Date issued: January 23, 2004

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

# Probe ET3DV6

## SN:1609

|                  |                  |
|------------------|------------------|
| Manufactured:    | July 27, 2001    |
| Last calibrated: | June 20, 2002    |
| Recalibrated:    | January 22, 2004 |

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

ET3DV6 SN:1609

January 22, 2004

**DASY - Parameters of Probe: ET3DV6 SN:1609****Sensitivity in Free Space**

|       |                                          |
|-------|------------------------------------------|
| NormX | $1.72 \mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | $1.72 \mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | $1.72 \mu\text{V}/(\text{V}/\text{m})^2$ |

**Diode Compression<sup>A</sup>**

|       |    |    |
|-------|----|----|
| DCP X | 96 | mV |
| DCP Y | 96 | mV |
| DCP Z | 96 | mV |

**Sensitivity in Tissue Simulating Liquid (Conversion Factors)**

Please see Page 7.

**Boundary Effect****Head                      900 MHz      Typical SAR gradient: 5 % per mm**

|                                                         |        |        |
|---------------------------------------------------------|--------|--------|
| Sensor Cener to Phantom Surface Distance                | 3.7 mm | 4.7 mm |
| SAR <sub>be</sub> [%]      Without Correction Algorithm | 9.2    | 5.0    |
| SAR <sub>be</sub> [%]      With Correction Algorithm    | 0.1    | 0.2    |

**Head                      1800 MHz      Typical SAR gradient: 10 % per mm**

|                                                         |        |        |
|---------------------------------------------------------|--------|--------|
| Sensor to Surface Distance                              | 3.7 mm | 4.7 mm |
| SAR <sub>be</sub> [%]      Without Correction Algorithm | 13.5   | 9.3    |
| SAR <sub>be</sub> [%]      With Correction Algorithm    | 0.2    | 0.2    |

**Sensor Offset**

|                            |              |
|----------------------------|--------------|
| Probe Tip to Sensor Center | 2.7 mm       |
| Optical Surface Detection  | in tolerance |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.


<sup>A</sup> numerical linearization parameter; uncertainty not required

ET3DV6 SN:1609

January 22, 2004

## Frequency Response of E-Field



( TEM-Cell:ifi110, Waveguide R22)



ET3DV6 SN:1609

January 22, 2004

### Receiving Pattern ( $\phi$ ) , $\theta = 0^\circ$




**Axial Isotropy Error <  $\pm 0.2$  dB**

ET3DV6 SN:1609

January 22, 2004

## Dynamic Range f(SAR<sub>head</sub>) ( Waveguide R22 )




Probe Linearity <  $\pm 0.2$  dB

ET3DV6 SN:1609

January 22, 2004

## Conversion Factor Assessment




| f [MHz] | Validity [MHz] <sup>B</sup> | Tissue | Permittivity | Conductivity | Alpha | Depth | ConvF Uncertainty  |
|---------|-----------------------------|--------|--------------|--------------|-------|-------|--------------------|
| 450     | 400-500                     | Head   | 43.5 ± 5%    | 0.87 ± 5%    | 0.73  | 1.68  | 6.88 ± 15.9% (k=2) |
| 900     | 800-1000                    | Head   | 41.5 ± 5%    | 0.97 ± 5%    | 0.58  | 1.90  | 6.62 ± 11.3% (k=2) |
| 1800    | 1710-1910                   | Head   | 40.0 ± 5%    | 1.40 ± 5%    | 0.48  | 2.67  | 5.29 ± 11.7% (k=2) |
| 2450    | 2400-2500                   | Head   | 39.2 ± 5%    | 1.80 ± 5%    | 1.06  | 1.82  | 4.83 ± 9.7% (k=2)  |
| 450     | 400-500                     | Body   | 56.7 ± 5%    | 0.94 ± 5%    | 0.96  | 1.50  | 6.73 ± 15.9% (k=2) |
| 835     | 750-950                     | Body   | 55.2 ± 5%    | 0.97 ± 5%    | 0.47  | 2.20  | 6.57 ± 11.9% (k=2) |
| 1900    | 1800-2000                   | Body   | 53.3 ± 5%    | 1.52 ± 5%    | 0.60  | 2.78  | 4.69 ± 11.3% (k=2) |
| 2450    | 2400-2500                   | Body   | 52.7 ± 5%    | 1.95 ± 5%    | 1.48  | 1.47  | 4.62 ± 9.7% (k=2)  |

<sup>B</sup> The total standard uncertainty is calculated as root-sum-square of standard uncertainty of the Conversion Factor at calibration frequency and the standard uncertainty for the indicated frequency band.

ET3DV6 SN:1609

January 22, 2004

**Deviation from Isotropy in HSL**Error ( $\theta, \phi$ ),  $f = 900$  MHz**Spherical Isotropy Error <  $\pm 0.4$  dB**

## APPENDIX E – DIPOLE CALIBRATION DATA

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

Client **Pantech & Curitel (Dymstec)**

**CALIBRATION CERTIFICATE**

Object(s) **D900V2 - SN:062**

Calibration procedure(s) **QA CAL-05.v2  
Calibration procedure for dipole validation kits**

Calibration date: **March 22, 2004**

Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                | ID #       | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|---------------------------|------------|-------------------------------------------|------------------------|
| Power meter EPM E442      | GB37480704 | 6-Nov-03 (METAS, No. 252-0254)            | Nov-04                 |
| Power sensor HP 8481A     | US37292783 | 6-Nov-03 (METAS, No. 252-0254)            | Nov-04                 |
| Power sensor HP 8481A     | MY41092317 | 18-Oct-02 (Agilent, No. 20021018)         | Oct-04                 |
| RF generator R&S SML-03   | 100698     | 27-Mar-2002 (R&S, No. 20-92389)           | In house check: Mar-05 |
| Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (SPEAG, in house check Nov-03)  | In house check: Oct 05 |

Calibrated by: 

| Name           | Function   | Signature                                                                             |
|----------------|------------|---------------------------------------------------------------------------------------|
| Judith Mueller | Technician |  |

Approved by: 

| Name          | Function            | Signature                                                                             |
|---------------|---------------------|---------------------------------------------------------------------------------------|
| Katja Pokovic | Laboratory Director |  |

Date issued: March 23, 2004

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

Schmid &amp; Partner Engineering AG

**s p e a g**

---

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, <http://www.speag.com>

# DASY

## Dipole Validation Kit

Type: D900V2

Serial: 062

Manufactured: September 25, 1999  
Calibrated: March 22, 2004

## **1. Measurement Conditions**

The measurements were performed in the flat section of the SAM twin phantom filled with **head simulating solution** of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>41.4</b>       | $\pm 5\%$ |
| Conductivity           | <b>0.95 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.18 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250 mW  $\pm 3\%$ . The results are normalized to 1W input power.

## **2. SAR Measurement with DASY4 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|                                                    |                                                            |
|----------------------------------------------------|------------------------------------------------------------|
| averaged over 1 cm <sup>3</sup> (1 g) of tissue:   | <b>10.9 mW/g <math>\pm 16.8\%</math> (k=2)<sup>1</sup></b> |
| averaged over 10 cm <sup>3</sup> (10 g) of tissue: | <b>7.08 mW/g <math>\pm 16.2\%</math> (k=2)<sup>1</sup></b> |

---

<sup>1</sup> validation uncertainty

### **3. Dipole Impedance and Return Loss**

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:       **1.393 ns**   (one direction)  
Transmission factor:   **0.990**     (voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section 1 and the distance spacer was in place during impedance measurements.

Feedpoint impedance at 900 MHz:        $\text{Re}\{Z\} = 50.9 \Omega$

$\text{Im}\{Z\} = -0.3 \Omega$

Return Loss at 900 MHz               **-40.6 dB**

### **4. Handling**

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

### **5. Design**

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

### **6. Power Test**

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Page: 011  
Date/Time: 03/22/04 12:03:47

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN062**

Communication System: CW-900; Frequency: 900 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz;

Medium parameters used:  $f = 900 \text{ MHz}$ ;  $\sigma = 0.95 \text{ mho/m}$ ;  $\epsilon_r = 41.4$ ;  $\rho = 1000 \text{ kg/m}^3$ 

Phantom section: Flat Section

Measurement Standard: DASy4 (High Precision Assessment)

DASy4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.18, 6.18, 6.18); Calibrated: 1/23/2004
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn411; Calibrated: 11/6/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006;
- Measurement SW: DASy4, V4.2 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 112

**Pin = 250 mW; d = 15 mm/Area Scan (81x81x1):** Measurement grid:  $dx=15\text{mm}$ ,  $dy=15\text{mm}$ 

Reference Value = 56.8 V/m; Power Drift = 0.0 dB

Maximum value of SAR (interpolated) = 2.95 mW/g

**Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid:  $dx=5\text{mm}$ ,  $dy=5\text{mm}$ ,  $dz=5\text{mm}$ 

Reference Value = 56.8 V/m; Power Drift = 0.0 dB

Maximum value of SAR (measured) = 2.95 mW/g

Peak SAR (extrapolated) = 4.12 W/kg

SAR(1 g) = 2.73 mW/g; SAR(10 g) = 1.77 mW/g

