


APPENDIX D - PROBE CALIBRATION CERTIFICATES

 <p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn</p>		  <p>中国认可 国际互认 校准 CNAS CALIBRATION CNAS L0570</p>																																																																																
Client	BACL	Certificate No: 24J02Z000756																																																																																
<h3>CALIBRATION CERTIFICATE</h3> <table border="1"> <tr> <td>Object</td> <td>ES3DV3 - SN : 3220</td> </tr> <tr> <td>Calibration Procedure(s)</td> <td>FF-Z11-004-02 Calibration Procedures for Dosimetric E-field Probes</td> </tr> <tr> <td>Calibration date:</td> <td>October 15, 2024</td> </tr> <tr> <td colspan="2"> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> </td> </tr> <tr> <td>Primary Standards</td> <td>ID #</td> <td>Cal Date(Calibrated by, Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>19-Oct-23(CTTL, No.J23X11026)</td> <td>Oct-24</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>19-Oct-23(CTTL, No.J23X11026)</td> <td>Oct-24</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104292</td> <td>19-Oct-23(CTTL, No.J23X11026)</td> <td>Oct-24</td> </tr> <tr> <td>Reference 10dBAttenuator</td> <td>18N50W-10dB</td> <td>19-Jan-23(CTTL, No.J23X00212)</td> <td>Jan-25</td> </tr> <tr> <td>Reference 20dBAttenuator</td> <td>18N50W-20dB</td> <td>19-Jan-23(CTTL, No.J23X00211)</td> <td>Jan-25</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>28-May-24(SPEAG, No.EX-7307_May24)</td> <td>May-25</td> </tr> <tr> <td>DAE4</td> <td>SN 771</td> <td>19-Jan-24(SPEAG, No.DAE4-771_Jan24)</td> <td>Jan-25</td> </tr> <tr> <td>Secondary Standards</td> <td>ID #</td> <td>Cal Date(Calibrated by, Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>SignalGenerator MG3700A</td> <td>6201052605</td> <td>12-Jun-24(CTTL, No.24J02X005419)</td> <td>Jun-25</td> </tr> <tr> <td>SignalGenerator APSIN26G</td> <td>181-33A6D0700-1959</td> <td>26-Mar-24(CTTL, No.24J02X002468)</td> <td>Mar-25</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>25-Dec-23(CTTL, No.J23X13425)</td> <td>Dec-24</td> </tr> <tr> <td>Reference 10dBAttenuator</td> <td>BT0520</td> <td>11-May-23(CTTL, No.J23X04061)</td> <td>May-25</td> </tr> <tr> <td>Reference 20dBAttenuator</td> <td>BT0267</td> <td>11-May-23(CTTL, No.J23X04062)</td> <td>May-25</td> </tr> <tr> <td>OCP DAK-3.5</td> <td>SN 1040</td> <td>22-Jan-24(SPEAG, No.OCP-DAK3.5-1040_Jan24)</td> <td>Jan-25</td> </tr> <tr> <td>Calibrated by:</td> <td>Yu Zongying</td> <td>Function: SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Jun</td> <td>Function: SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyuan</td> <td>Function: SAR Project Leader</td> <td></td> </tr> </table>			Object	ES3DV3 - SN : 3220	Calibration Procedure(s)	FF-Z11-004-02 Calibration Procedures for Dosimetric E-field Probes	Calibration date:	October 15, 2024	<p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p>		Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	19-Oct-23(CTTL, No.J23X11026)	Oct-24	Power sensor NRP8S	104291	19-Oct-23(CTTL, No.J23X11026)	Oct-24	Power sensor NRP8S	104292	19-Oct-23(CTTL, No.J23X11026)	Oct-24	Reference 10dBAttenuator	18N50W-10dB	19-Jan-23(CTTL, No.J23X00212)	Jan-25	Reference 20dBAttenuator	18N50W-20dB	19-Jan-23(CTTL, No.J23X00211)	Jan-25	Reference Probe EX3DV4	SN 7307	28-May-24(SPEAG, No.EX-7307_May24)	May-25	DAE4	SN 771	19-Jan-24(SPEAG, No.DAE4-771_Jan24)	Jan-25	Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration	SignalGenerator MG3700A	6201052605	12-Jun-24(CTTL, No.24J02X005419)	Jun-25	SignalGenerator APSIN26G	181-33A6D0700-1959	26-Mar-24(CTTL, No.24J02X002468)	Mar-25	Network Analyzer E5071C	MY46110673	25-Dec-23(CTTL, No.J23X13425)	Dec-24	Reference 10dBAttenuator	BT0520	11-May-23(CTTL, No.J23X04061)	May-25	Reference 20dBAttenuator	BT0267	11-May-23(CTTL, No.J23X04062)	May-25	OCP DAK-3.5	SN 1040	22-Jan-24(SPEAG, No.OCP-DAK3.5-1040_Jan24)	Jan-25	Calibrated by:	Yu Zongying	Function: SAR Test Engineer		Reviewed by:	Lin Jun	Function: SAR Test Engineer		Approved by:	Qi Dianyuan	Function: SAR Project Leader	
Object	ES3DV3 - SN : 3220																																																																																	
Calibration Procedure(s)	FF-Z11-004-02 Calibration Procedures for Dosimetric E-field Probes																																																																																	
Calibration date:	October 15, 2024																																																																																	
<p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p>																																																																																		
Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration																																																																															
Power Meter NRP2	106277	19-Oct-23(CTTL, No.J23X11026)	Oct-24																																																																															
Power sensor NRP8S	104291	19-Oct-23(CTTL, No.J23X11026)	Oct-24																																																																															
Power sensor NRP8S	104292	19-Oct-23(CTTL, No.J23X11026)	Oct-24																																																																															
Reference 10dBAttenuator	18N50W-10dB	19-Jan-23(CTTL, No.J23X00212)	Jan-25																																																																															
Reference 20dBAttenuator	18N50W-20dB	19-Jan-23(CTTL, No.J23X00211)	Jan-25																																																																															
Reference Probe EX3DV4	SN 7307	28-May-24(SPEAG, No.EX-7307_May24)	May-25																																																																															
DAE4	SN 771	19-Jan-24(SPEAG, No.DAE4-771_Jan24)	Jan-25																																																																															
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration																																																																															
SignalGenerator MG3700A	6201052605	12-Jun-24(CTTL, No.24J02X005419)	Jun-25																																																																															
SignalGenerator APSIN26G	181-33A6D0700-1959	26-Mar-24(CTTL, No.24J02X002468)	Mar-25																																																																															
Network Analyzer E5071C	MY46110673	25-Dec-23(CTTL, No.J23X13425)	Dec-24																																																																															
Reference 10dBAttenuator	BT0520	11-May-23(CTTL, No.J23X04061)	May-25																																																																															
Reference 20dBAttenuator	BT0267	11-May-23(CTTL, No.J23X04062)	May-25																																																																															
OCP DAK-3.5	SN 1040	22-Jan-24(SPEAG, No.OCP-DAK3.5-1040_Jan24)	Jan-25																																																																															
Calibrated by:	Yu Zongying	Function: SAR Test Engineer																																																																																
Reviewed by:	Lin Jun	Function: SAR Test Engineer																																																																																
Approved by:	Qi Dianyuan	Function: SAR Project Leader																																																																																
Issued: October 18, 2024																																																																																		
<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>																																																																																		



In Collaboration with
s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A,B,C,D	modulation dependent linearization parameters
Polarization Φ	Φ rotation around probe axis
Polarization θ	θ rotation around an axis that is in the plane normal to probe axis (at measurement center), $\theta=0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- $NORM_{x,y,z}$: Assessed for E-field polarization $\theta=0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: waveguide). $NORM_{x,y,z}$ are only intermediate values, i.e., the uncertainties of $NORM_{x,y,z}$ does not effect the E^2 -field uncertainty inside TSL (see below ConvF).
- $NORM(f)_{x,y,z} = NORM_{x,y,z} * frequency_response$ (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- $DCP_{x,y,z}$: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
- $Ax,y,z; Bx,y,z; Cx,y,z; VRx,y,z; A,B,C$ are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters*: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty valued are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to $NORM_{x,y,z} * ConvF$ whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy)*: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset*: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle*: The angle is assessed using the information gained by determining the $NORM_x$ (no uncertainty required).



In Collaboration with
s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn http://www.caict.ac.cn

DASY/EASY – Parameters of Probe: ES3DV3 – SN:3220

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm(μ V/(V/m) ²) ^A	1.32	1.43	1.22	\pm 10.0%
DCP(mV) ^B	115.9	113.2	112.9	

Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB/ μ V	C	D dB	VR mV	Unc ^E (k=2)
0	CW	X	0.0	0.0	1.0	0.00	303.4	\pm 3.7%
		Y	0.0	0.0	1.0		306.0	
		Z	0.0	0.0	1.0		290.0	

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X, Y, Z do not affect the E²-field uncertainty inside TSL (see Page 4).

^B Numerical linearization parameter: uncertainty not required.

^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.



In Collaboration with
s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn http://www.caict.ac.cn

DASY/EASY – Parameters of Probe: ES3DV3 – SN:3220

Calibration Parameter Determined in Head Tissue Simulating Media

f [MHz] ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unct. (k=2)
750	41.9	0.89	6.68	6.68	6.68	0.36	1.46	±12.7%
900	41.5	0.97	6.43	6.43	6.43	0.35	1.57	±12.7%
1750	40.1	1.37	5.53	5.53	5.53	0.58	1.25	±12.7%
1900	40.0	1.40	5.24	5.24	5.24	0.61	1.27	±12.7%
2300	39.5	1.67	4.97	4.97	4.97	0.80	1.14	±12.7%
2450	39.2	1.80	4.83	4.83	4.83	0.86	1.12	±12.7%
2600	39.0	1.96	4.66	4.66	4.66	0.90	1.09	±12.7%

^C Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequency up to 6 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.




In Collaboration with
s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

Frequency Response of E-Field (TEM-Cell: ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: $\pm 7.4\% (k=2)$




In Collaboration with
s p e a g
CALIBRATION LABORATORY

CAICT



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

Receiving Pattern (Φ), $\theta=0^\circ$

f=600 MHz, TEM



f=1800 MHz, R22







In Collaboration with
s p e a g
CALIBRATION LABORATORY

CAICT

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

Dynamic Range f(SAR_{head}) (TEM cell, f = 900 MHz)




Uncertainty of Linearity Assessment: $\pm 0.9\% (k=2)$


Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2117
 E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

Conversion Factor Assessment


$f=750 \text{ MHz, WGLS R9(H_convF)}$



$f=1750 \text{ MHz, WGLS R22(H_convF)}$



Deviation from Isotropy in Liquid



Uncertainty of Spherical Isotropy Assessment: $\pm 3.2\% (k=2)$



In Collaboration with
s p e a g
CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2117
E-mail: emf@caict.ac.cn <http://www.caict.ac.cn>

DASY/EASY – Parameters of Probe: ES3DV3 – SN:3220

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	141.7
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disable
Probe Overall Length	337mm
Probe Body Diameter	10mm
Tip Length	10mm
Tip Diameter	4mm
Probe Tip to Sensor X Calibration Point	2mm
Probe Tip to Sensor Y Calibration Point	2mm
Probe Tip to Sensor Z Calibration Point	2mm
Recommended Measurement Distance from Surface	3mm