

Federal Communications Commission
Authorization and Evaluation Division
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

Applicant's declaration concerning RF Radiation Exposure

We hereby indicate that the product
RTX3320
is considered to be a mobile transceiver.

The internal antenna used for this mobile transmitter must provide a separation distance of at least 20 cm from all persons and must not be co-located or operating in conjunction with any other antenna or transmitter.

A safety statement concerning minimum separation distances from enclosure of the RTX3320 will be integrated in the user's manual to provide end-users with transmitter operating conditions for satisfying RF exposure compliance.

The appropriate information can be drawn from the test report no. GOM20403-8572-C-1 and the accompanying calculations.

RTX Healthcare A/S
Stroemmen 6
9400 Noerresundby
Denmark

Noerresundby 04-May-2004

