

Information about the Applicant

Grantee (Company Name)	Nonin Medical, Inc
FCC Contact Person	Robert Johnson
Address	13700 1st Avenue North
City, State, Zip	Plymouth, MN 55441-5443
Job Number	NONN0018
Model	Saber
FCC ID	RG9SABER
Agent	
Approval Type	Class II Permissive Change
Equipment Class	DSS Part 15, Spread Spectrum Transmitter
Rule Part	15C (15.247)

Overview

This application is for a Class II permissive change required because of change in spurious radiated emission levels of FCC ID: RG9SABER, model SABER installed in Nonin Medical Model 7600 RegionalOximeter. The radio is a Bluetooth module. No changes have been made to the radio nor the antenna. Test report NONN0018 from Northwest EMC, Inc. shows compliance of spurious emission levels.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature


12/20/2010

Dave Tolman, TCB Committee