



St. Jude Medical  
Implantable  
Electronic Systems  
15900 Valley View Court  
Sylmar, CA 91342 USA  
Main 818 362 6822  
Fax 818 364 5814

March 30, 2014

TUV SUD BABT FCB  
Octagon House,  
Segensworth Road,  
Fareham,  
Hampshire,  
PO15 5RL

### **Model Names and Differences**

We, St. Jude Medical IESD, hereby request Family certification for the following RF electrically identical implantable pulse generators. The differences consist of battery capacities (either 5.3 or 7.5 Ah), type of therapy (Tonic or non-Tonic SCS/DBS, controlled by software) and method of connecting patient electrodes to the device (SJM or 1.27 mm header).

Model 3660 Orion IPG, 5.3Ah, SCS, SJM header  
Model 3661 Orion IPG, 5.3Ah, SCS, 1.27mm header  
Model 3662 Orion IPG, 7.5Ah, SCS, SJM header  
Model 3663 Orion IPG, 7.5Ah, SCS, 1.27mm header  
Model 6660 Orion IPG, 5.3Ah, DBS, Tonic, SJM header  
Model 6661 Orion IPG, 5.3Ah, DBS, Tonic, 1.27mm header  
Model 6662 Orion IPG, 7.5Ah, DBS, Tonic, SJM header  
Model 6663 Orion IPG, 7.5Ah, DBS, Tonic, 1.27mm header  
Model 3665 Orion IPG, 5.3Ah, SCS, Tonic, SJM header  
Model 3667 Orion IPG, 7.5Ah, SCS, Tonic, SJM header

Sincerely,

Name: Jorge Amely-Velez  
Title: Senior Principal Engineer  
St. Jude Medical IESD  
15900 Valley View Court  
Sylmar, CA 91342