

Helen Zhao

Subject: FW: JAWON MEDICAL CO., LTD., FCC ID: RIXECHOFT, Assessment NO.: AN04T4307, Notice#1

From: schong

Sent: Monday, November 08, 2004 4:06 AM

To: Helen Zhao

Cc: gwlee01@onetech.co.kr

Subject: Re: JAWON MEDICAL CO., LTD., FCC ID: RIXECHOFT, Assessment NO.: AN04T4307, Notice#1

Dear Ms. Helen,

Thank you for your kind help and cooperation.

Regarding the assessment no. AN04T4307, we attached the revised test report that is included additional power line conducted emission test data when the EUT is connected with USB port.

And we also attached the revised test setup photos.

If you have any question, please feel free to contact us.

Thank you and have a nice day!

Best Regards,

S. C. Hong / Project Engineer

ONETECH Corp.

ONETECH Testing & Eval. Lab.

EMC Div II, Team 2

From: Helen Zhao

To: schong

Sent: Wednesday, November 03, 2004 4:04 PM

Subject: RE: JAWON MEDICAL CO., LTD., FCC ID: RIXECHOFT, Assessment NO.: AN04T4307, Notice#1

Question #1: Please provide additional power-line conducted emission test data when the EUT was power through USB port of PC.

Best regards,
Helen Zhao