## Mike Kuo

From: Yoichi Nakajima [

Sent: February26日2004年Thursday 7:26 PM

To: Mike Kuo Cc: Mika Kaneko

Subject: Re: Toshiba Medical Systems Corporation, FCC ID:ROKMED001, AN04T3672







(Revised) Exhibit G Revised) Exhibit B2 (Revised) Exhibit D User's Man... Confident... TechnicalDe...

Hi Mike

Thank you for your assessment.

A1. Please refer to the attached revised confidentiality letter.

A2. The applicant changed the user's manual and technical description.

Please refer to the attached revised user's manual and technical description.

Best Regards Yoichi Nakajima JQA EMC Dept.

---- Original Message -----

From: "Mike Kuo" <MKUO@CCSEMC.com>

To:

Cc: "Mika Kaneko" < MKaneko@CCSEMC.com> Sent: Thursday, February 26, 2004 11:57 AM

Subject: FW: Toshiba Medical Systems Corporation, FCC ID:ROKMED001, AN04T3672

> > ----Original Message---- > From: CERTADM > Sent: Wednesday, February 25, 2004 6:56 PM > To: Mike Kuo

> Subject: Toshiba Medical Systems Corporation, FCC ID:ROKMED001, AN04T3672

> . .

> Notice\_content

> -----

- > Question #1: Operational description is listed as one of item in the request
- > for confidentiality letter. The information contains in the operational
- > description ( technical description as file name ) has same information in
- > the user manual. Please remove operational description from the request for
- > confidentiality letter.

>

- > Question #2: Based upon technical description and user manual, the output
- > power is rated as 2.5mW but actual measured value is about 1mW. Please
- > explain the difference.
- > Best Regards
- > Mike Kuo
- > The items indicated above must be submitted before processing can continue
- > on the above referenced application. Failure to provide the requested
- > information within 30 days of the original e-mail date may result in
- > application dismissal and forfeiture of the filing fee. Also, please note
- > that partial responses increase processing time and should not be submitted.
- > Any questions about the content of this correspondence should be directed to
- > the e-mail address listed below the name of the sender.