

# SAR EVALUATION REPORT


For

**CyberNet Inc.**

6<sup>th</sup> Floor, Sebang Bldg.(New Bldg.) 708-8, Yoksamdong Kangnamku, Seoul, Korea

**FCC ID: RXT-JADE-AI-CNN01**

2004-04-08

|                                                                                                                                                                     |                                                 |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|
| <b>This Report Concerns:</b><br><input checked="" type="checkbox"/> Original Report                                                                                 | <b>Equipment Type:</b><br>Wireless POS Terminal |
| <b>Test Engineer:</b> Daniel Deng/                                                |                                                 |
| <b>Report No.:</b> R0402192S                                                                                                                                        |                                                 |
| <b>Test Date:</b> 2004-03-24/2004-03-30                                                                                                                             |                                                 |
| <b>Reviewed By:</b> Ling Zhang/                                                  |                                                 |
| <b>Prepared By:</b> Bay Area Compliance Laboratory Corporation (BACL)<br>230 Commercial Street<br>Sunnyvale, CA 94085<br>Tel: (408) 732-9162<br>Fax: (408) 732 9164 |                                                 |

**Note:** This test report is specially limited to the above client company and the product model only. It may not be duplicated without prior written consent of Bay Area Compliance Laboratory Corporation. This report **must not** be used by the client to claim product endorsement by NVLAP or any agency of the U.S. Government.

**TABLE OF CONTENTS**

|                                                          |           |
|----------------------------------------------------------|-----------|
| <b>SUMMARY</b> .....                                     | <b>4</b>  |
| <b>1 - REFERENCE</b> .....                               | <b>5</b>  |
| <b>2 - TESTING EQUIPMENT</b> .....                       | <b>6</b>  |
| 2.1 EQUIPMENTS LIST & CALIBRATION INFO .....             | 6         |
| 2.2 EQUIPMENT CALIBRATION CERTIFICATE .....              | 6         |
| 1900MHZ BODY LIQUID VALIDATION 3/30/2004 .....           | 43        |
| <b>3 - EUT DESCRIPTION</b> .....                         | <b>45</b> |
| <b>4 - SYSTEM TEST CONFIGURATION</b> .....               | <b>46</b> |
| 4.1 JUSTIFICATION .....                                  | 46        |
| 4.2 EUT EXERCISE PROCEDURE .....                         | 46        |
| 4.3 EQUIPMENT MODIFICATIONS .....                        | 46        |
| <b>5 – CONDUCTED OUTPUT POWER MEASUREMENTS</b> .....     | <b>47</b> |
| 5.1 PROVISION APPLICABLE.....                            | 47        |
| 5.2 TEST PROCEDURE .....                                 | 47        |
| 5.3 TEST EQUIPMENT .....                                 | 47        |
| 5.4 TEST RESULTS .....                                   | 47        |
| <b>6 - DOSIMETRIC ASSESSMENT SETUP</b> .....             | <b>50</b> |
| 6.1 MEASUREMENT SYSTEM DIAGRAM .....                     | 51        |
| 6.2. SYSTEM COMPONENTS .....                             | 52        |
| 6.3 MEASUREMENT UNCERTAINTY .....                        | 56        |
| <b>7 - EVALUATION PROCEDURE</b> .....                    | <b>57</b> |
| 7.1 SAR EVALUATION PROCEDURE.....                        | 57        |
| 7.2 EXPOSURE LIMITS .....                                | 58        |
| 7.3 SIMULATED TISSUE LIQUID PARAMETER CONFIRMATION ..... | 58        |
| 7.4 SAR MEASUREMENT .....                                | 58        |
| 7.5 SYSTEM ACCURACY VERIFICATION .....                   | 59        |
| 7.6 LIQUID MEASUREMENT RESULT .....                      | 60        |
| <b>8 - SAR TEST RESULTS</b> .....                        | <b>65</b> |
| 8.1 SAR BODY AND HEAD WORST-CASE TEST DATA .....         | 65        |
| 8.2 PLOTS OF TEST RESULT .....                           | 65        |
| <b>EXHIBIT A - SAR SETUP PHOTOGRAPHS</b> .....           | <b>72</b> |
| RIGHT SIDE TOUCHING WITH PHANTOM .....                   | 72        |
| BACK SIDE TOUCHING WITH PHANTOM .....                    | 72        |
| BOTTOM SIDE TOUCHING WITH PHANTOM .....                  | 73        |
| <b>EXHIBIT B – EUT PHOTOGRAPHS</b> .....                 | <b>74</b> |
| EUT - TOP VIEW .....                                     | 74        |
| EUT - BACK VIEW.....                                     | 74        |
| EUT – BACK SIDE WITHOUT BATTERY .....                    | 75        |
| DOCKING UNIT - TOP VIEW .....                            | 75        |
| DOCKING UNIT - BOTTOM VIEW .....                         | 76        |
| DOCKING UNIT - CONNECTOR VIEW .....                      | 76        |
| BATTERY – TOP VIEW.....                                  | 77        |
| BATTERY – BOTTOM VIEW .....                              | 77        |
| AC ADAPTER – TOP VIEW .....                              | 78        |
| CHASSIS – COVER OFF VIEW .....                           | 78        |
| EUT – COMPONENT VIEW I.....                              | 79        |
| EUT – COMPONENT VIEW II.....                             | 79        |
| EUT – COMPONENT VIEW III .....                           | 80        |
| TRANSCIEVER – TOP VIEW .....                             | 80        |
| TRANSCIEVER – BOTTOM VIEW .....                          | 81        |
| TRANSCIEVER – COMPONENT VIEW.....                        | 81        |
| TRANSCIEVER – SOLDER VIEW .....                          | 82        |
| TRANSCIEVER - SHIELDING OFF VIEW I.....                  | 82        |
| TRANSCIEVER - SHIELDING OFF VIEW II .....                | 83        |



---

## SUMMARY

---

The US Federal Communications Commission has released the report and order “Guidelines for Evaluating the Environmental Effects of RF Radiation”, ET Docket No. 93-62 in August 1996 [1].

The order requires routine SAR evaluation prior to equipment authorization of portable transmitter devices, including portable telephones. For consumer products, the applicable limit is 1.6 mW/g as recommended by the ANSI/IEEE standard C95.1-1992 [6] for an uncontrolled environment (Paragraph 65). According to the Supplement C of OET Bulletin 65 “Evaluating Compliance with FCC Guide-lines for Human Exposure to Radio frequency Electromagnetic Fields”, released on Jun 29, 2001 by the FCC, the device should be evaluated at maximum output power (radiated from the antenna) under “worst-case” conditions for normal or intended use, incorporating normal antenna operating positions, device peak performance frequencies and positions for maximum RF energy coupling.

This report describes the methodology and results of experiments performed on wireless data terminal. The objective was to determine if there is RF radiation and if radiation is found, what is the extent of radiation with respect to safety limits. SAR (Specific Absorption Rate) is the measure of RF exposure determined by the amount of RF energy absorbed by human body (or its parts) – to determine how the RF energy couples to the body or head which is a primary health concern for body worn devices. The limit below which the exposure to RF is considered safe by regulatory bodies in North America is 1.6 mW/g average over 1 gram of tissue mass.

The test configurations were laid out on a specially designed test fixture to ensure the reproducibility of measurements. Each configuration was scanned for SAR. Analysis of each scan was carried out to characterize the above effects in the device.

There was no SAR of any concern measured on the device for any of the investigated configurations.

---

## 1 - REFERENCE

---

- [1] Federal Communications Commission, "Report and order: Guidelines for evaluating the environmental effects of radiofrequency radiation", Tech. Rep. FCC 96-326, FCC, Washington, D.C. 20554, 1996.
- [2] David L. Means Kwok Chan, Robert F. Cleveland, "Evaluating compliance with FCC guidelines for human exposure to radiofrequency electromagnetic fields", Tech. Rep., Federal Communication Commission, Office of Engineering & Technology, Washington, DC, 1997.
- [3] Thomas Schmid, Oliver Egger, and Niels Kuster, "Automated E-field scanning system for dosimetric assessments", IEEE Transactions on Microwave Theory and Techniques, vol. 44, pp. 105-113, Jan. 1996.
- [4] Niels Kuster, Ralph Kastle, and Thomas Schmid, "Dosimetric evaluation of mobile communications equipment with known precision", IEEE Transactions on Communications, vol. E80-B, no. 5, pp. 645-652, May 1997.
- [5] CENELEC, "Considerations for evaluating of human exposure to electromagnetic fields (EMFs) from mobile telecommunication equipment (MTE) in the frequency range 30MHz - 6GHz", Tech. Rep., CENELEC, European Committee for Electrotechnical Standardization, Brussels, 1997.
- [6] ANSI, ANSI/IEEE C95.1-1992: IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3 kHz to 300 GHz, The Institute of Electrical and Electronics Engineers, Inc., New York, NY 10017, 1992.
- [7] Katja Pokovic, Thomas Schmid, and Niels Kuster, "Robust setup for precise calibration of E-field probes in tissue simulating liquids at mobile communications frequencies", in ICECOM '97, Dubrovnik, October 15-17, 1997, pp. 120-24.
- [8] Katja Pokovic, Thomas Schmid, and Niels Kuster, "E-field probe with improved isotropy in brain simulating liquids", in Proceedings of the ELMAR, Zadar, Croatia, 23-25 June, 1996, pp. 172-175.
- [9] Volker Hombach, Klaus Meier, Michael Burkhardt, Eberhard Kuhn, and Niels Kuster, "The dependence of EM energy absorption upon human head modeling at 900 MHz", IEEE Transactions on Microwave Theory and Techniques, vol. 44, no. 10, pp. 1865-1873, Oct. 1996.
- [10] Klaus Meier, Ralf Kastle, Volker Hombach, Roger Tay, and Niels Kuster, "The dependence of EM energy absorption upon human head modeling at 1800 MHz", IEEE Transactions on Microwave Theory and Techniques, Oct. 1997, in press.
- [11] W. Gander, Computermathematik, Birkhaeuser, Basel, 1992.
- [12] W. H. Press, S. A. Teukolsky, W. T. Vetterling, and B. P. Flannery, Numerical Recipes in C, The Art of Scientific Computing, Second Edition, Cambridge University Press, 1992. Dosimetric Evaluation of Sample device, month 1998 9
- [13] NIS81 NAMAS, "The treatment of uncertainty in EMC measurement", Tech. Rep., NAMAS Executive, National Physical Laboratory, Teddington, Middlesex, England, 1994.
- [14] Barry N. Taylor and Christ E. Kuyatt, "Guidelines for evaluating and expressing the uncertainty of NIST measurement results", Tech. Rep., National Institute of Standards and Technology, 1994. Dosimetric Evaluation of Sample device, month 1998 10

## 2 - TESTING EQUIPMENT

### 2.1 Equipments List & Calibration Info

| Type / Model                         | Cal. Date  | S/N:            |
|--------------------------------------|------------|-----------------|
| DASY3 Professional Dosimetric System | N/A        | N/A             |
| Robot RX60L                          | N/A        | F00/5H31A1/A/01 |
| Robot Controller                     | N/A        | F01/5J72A1/A/01 |
| Dell Computer Optiplex GX110         | N/A        | N/A             |
| Pentium III, Windows NT              | N/A        | N/A             |
| SPEAG EDC3                           | N/A        | N/A             |
| SPEAG DAE3                           | 2003-06    | 456             |
| SPEAG E-Field Probe ES3DV2           | 2004-04-12 | 3019            |
| SPEAG Generic Twin Phantom           | N/A        | N/A             |
| SPEAG Light Alignment Sensor         | N/A        | 278             |
| Apriel Validation Dipole D-1800-S-2  | 2003-03-06 | BCL-049         |
| Brain Equivalent Matter (800MHz)     | Daily      | N/A             |
| Brain Equivalent Matter (1900MHz)    | Daily      | N/A             |
| Muscle Equivalent Matter (800MHz)    | Daily      | N/A             |
| Muscle Equivalent Matter (1900MHz)   | Daily      | N/A             |
| Robot Table                          | Each Use   | N/A             |
| Phone Holder                         | Each Use   | N/A             |
| Phantom Cover                        | Each Use   | N/A             |
| HP Spectrum Analyzer HP8566A         | N/A        | 2240A01930      |
| Microwave Amp. 8349A                 | N/A        | 2644A02662      |
| Power Meter HP436A                   | 2004-04-02 | 2709A29209      |
| Power Sensor HP8482A                 | 2004-04-02 | 2349A08568      |
| Network Analyzer HP-8752C            | 2004-07-30 | 820079          |
| Dielectric Probe Kit HP85070A        | Each Use   | US99360201      |

### 2.2 Equipment Calibration Certificate

Please see the attached file.

**Calibration Laboratory of**  
**Schmid & Partner**  
**Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland

**Client** Bay Area Comp. Lab (BACL)

## CALIBRATION CERTIFICATE

Object(s) E33DV2 - SN:3019

Calibration procedure(s) QA CAL-01.v2  
Calibration procedure for dosimetric E-field probes

Calibration date: October 9, 2003

Condition of the calibrated item In Tolerance (according to the specific calibration document)

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                        | ID #           | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|-----------------------------------|----------------|-------------------------------------------|------------------------|
| Power meter EPM E4419B            | GB41293874     | 2-Apr-03 (METAS, No 252-0250)             | Apr-04                 |
| Power sensor E4412A               | MY41495277     | 2-Apr-03 (METAS, No 252-0250)             | Apr-04                 |
| Reference 20 dB Attenuator        | SN: 5086 (20b) | 3-Apr-03 (METAS No. 251-0340)             | Apr-04                 |
| Fluke Process Calibrator Type 702 | SN: 6295803    | 8-Sep-03 (Sintrel SCS No. E-030020)       | Sep-04                 |
| Power sensor HP 8481A             | MY41092180     | 18-Sep-02 (Agilent, No. 20020918)         | In house check: Oct 03 |
| RF generator HP 8684C             | US3642U01700   | 4-Aug-99 (SPEAG, in house check Aug-02)   | In house check: Aug-05 |
| Network Analyzer HP 8753E         | US37390585     | 18-Oct-01 (Agilent, No. 24BR1033101)      | In house check: Oct 03 |

|                | Name          | Function            | Signature                                                                             |
|----------------|---------------|---------------------|---------------------------------------------------------------------------------------|
| Calibrated by: | Nico Vetter   | Technician          |  |
| Approved by:   | Katja Rokovec | Laboratory Director |  |

Date issued: October 9, 2003

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

880-KP0301061-A

Page 1 (1)

Schmid & Partner engineering AG

**s p e a g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, <http://www.speag.com>

# Probe ES3DV2

## SN: 3019

|                   |                  |
|-------------------|------------------|
| Manufactured:     | December 5, 2002 |
| Last calibration: | July 12, 2003    |

**Calibrated for DASY Systems**

(Note: non-compatible with DASY2 system!)



ES3DV2 SN: 3019

July 12, 2003

**DASY - Parameters of Probe: ES3DV2 SN: 3019****Sensitivity in Free Space****Diode Compression**

|       |                                                 |       |           |
|-------|-------------------------------------------------|-------|-----------|
| NormX | <b>1.03</b> $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP X | <b>99</b> |
| NormY | <b>1.12</b> $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP Y | <b>99</b> |
| NormZ | <b>0.98</b> $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP Z | <b>99</b> |

**Sensitivity in Tissue Simulating Liquid**

**Head**                      **900 MHz**                       $\epsilon_r = 41.5 \pm 5\%$                        $\sigma = 0.97 \pm 5\%$  mho/m  
Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

|         |                              |                  |             |
|---------|------------------------------|------------------|-------------|
| ConvF X | <b>6.4</b> $\pm 9.5\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>6.4</b> $\pm 9.5\%$ (k=2) | Alpha            | <b>0.68</b> |
| ConvF Z | <b>6.4</b> $\pm 9.5\%$ (k=2) | Depth            | <b>1.11</b> |

**Head**                      **1800 MHz**                       $\epsilon_r = 40.0 \pm 5\%$                        $\sigma = 1.40 \pm 5\%$  mho/m  
Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

|         |                              |                  |             |
|---------|------------------------------|------------------|-------------|
| ConvF X | <b>5.0</b> $\pm 9.5\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>5.0</b> $\pm 9.5\%$ (k=2) | Alpha            | <b>0.21</b> |
| ConvF Z | <b>5.0</b> $\pm 9.5\%$ (k=2) | Depth            | <b>2.78</b> |

**Boundary Effect**

**Head**                      **900 MHz**                      Typical SAR gradient: 5 % per mm

|                                                    |             |             |
|----------------------------------------------------|-------------|-------------|
| Probe Tip to Boundary                              | <b>1 mm</b> | <b>2 mm</b> |
| SAR <sub>be</sub> [%] Without Correction Algorithm | <b>4.3</b>  | <b>1.8</b>  |
| SAR <sub>be</sub> [%] With Correction Algorithm    | <b>0.0</b>  | <b>0.1</b>  |

**Head**                      **1800 MHz**                      Typical SAR gradient: 10 % per mm


|                                                    |             |             |
|----------------------------------------------------|-------------|-------------|
| Probe Tip to Boundary                              | <b>1 mm</b> | <b>2 mm</b> |
| SAR <sub>be</sub> [%] Without Correction Algorithm | <b>7.4</b>  | <b>5.0</b>  |
| SAR <sub>be</sub> [%] With Correction Algorithm    | <b>0.0</b>  | <b>0.1</b>  |

**Sensor Offset**

|                            |            |           |
|----------------------------|------------|-----------|
| Probe Tip to Sensor Center | <b>2.1</b> | <b>mm</b> |
|----------------------------|------------|-----------|



ES3DV2 SN: 3019

July 12, 2003

Receiving Pattern ( $\phi$ ,  $\theta = 0^\circ$ )

ES3DV2 SN: 3019

July 2003

Isotropy Error ( $\phi$ ),  $\theta = 0^\circ$ 


Page

ES3DV2 SN: 3019

July 12, 2003

## Frequency Response of E-Field



( TEM-Cell:Ifi110, Waveguide R22)



ES3DV2 SN: 3019

July 12, 2003

### Dynamic Range $f(\text{SAR}_{\text{brain}})$ ( Waveguide R22 )






ES3DV2 SN: 3019

July 12, 2003

## Conversion Factor Assessment



900 MHz  $\epsilon_r = 41.5 \pm 5\%$   $\sigma = 0.97 \pm 5\%$  mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

|         |                       |                  |
|---------|-----------------------|------------------|
| ConvF X | $6.4 \pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | $6.4 \pm 9.5\%$ (k=2) | Alpha            |
| ConvF Z | $6.4 \pm 9.5\%$ (k=2) | Depth            |

1800 MHz  $\epsilon_r = 40.0 \pm 5\%$   $\sigma = 1.40 \pm 5\%$  mho/m


Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

|         |                       |                  |
|---------|-----------------------|------------------|
| ConvF X | $5.0 \pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | $5.0 \pm 9.5\%$ (k=2) | Alpha            |
| ConvF Z | $5.0 \pm 9.5\%$ (k=2) | Depth            |

ES3DV2 SN: 3019

July 12, 2003

## Conversion Factor Assessment



**Head**      **5200**      **MHz**       $\epsilon_r = 36.0 \pm 5\%$        $\sigma = 4.66 \pm 5\%$  mho/m

Valid for f=4940-5460 MHz with Head Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                               |                  |             |
|---------|-------------------------------|------------------|-------------|
| ConvF X | <b>2.3</b> $\pm 14.6\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>2.3</b> $\pm 14.6\%$ (k=2) | Alpha            | <b>1.05</b> |
| ConvF Z | <b>2.3</b> $\pm 14.6\%$ (k=2) | Depth            | <b>1.50</b> |

**Body**      **5200**      **MHz**       $\epsilon_r = 49.0 \pm 5\%$        $\sigma = 5.30 \pm 5\%$  mho/m


Valid for f=4940-5460 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                               |                  |             |
|---------|-------------------------------|------------------|-------------|
| ConvF X | <b>1.4</b> $\pm 14.6\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>1.4</b> $\pm 14.6\%$ (k=2) | Alpha            | <b>1.01</b> |
| ConvF Z | <b>1.4</b> $\pm 14.6\%$ (k=2) | Depth            | <b>1.85</b> |

ES3DV2 SN: 3019

July 12, 2003

## Conversion Factor Assessment



Head 5800 MHz  $\epsilon_r = 35.3 \pm 5\%$   $\sigma = 5.27 \pm 5\%$  mho/m

Valid for f=5510-6090 MHz with Head Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                        |                  |      |
|---------|------------------------|------------------|------|
| ConvF X | $1.8 \pm 14.6\%$ (k=2) | Boundary effect: |      |
| ConvF Y | $1.8 \pm 14.6\%$ (k=2) | Alpha            | 0.90 |
| ConvF Z | $1.8 \pm 14.6\%$ (k=2) | Depth            | 1.90 |

Body 5800 MHz  $\epsilon_r = 48.2 \pm 5\%$   $\sigma = 6.00 \pm 5\%$  mho/m

Valid for f=5510-6090 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                        |                  |      |
|---------|------------------------|------------------|------|
| ConvF X | $1.2 \pm 14.6\%$ (k=2) | Boundary effect: |      |
| ConvF Y | $1.2 \pm 14.6\%$ (k=2) | Alpha            | 1.18 |
| ConvF Z | $1.2 \pm 14.6\%$ (k=2) | Depth            | 1.65 |




ES3DV2 SN: 3019

July 12, 2003

## Deviation from Isotropy in HSL

Error ( $\theta\phi$ ),  $f = 900$  MHz



Leugnausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, <http://www.speag.com>

# Probe ES3DV2

## SN:3019

### Additional Conversion Factors

|                   |                  |
|-------------------|------------------|
| Manufactured:     | December 5, 2002 |
| Last calibration: | July 12, 2003    |
| Add. calibration: | October 9, 2003  |

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

**DASY - Parameters of Probe: ES3DV2 SN:3019****Sensitivity in Free Space**

|       |                                                 |
|-------|-------------------------------------------------|
| NormX | <b>1.05</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | <b>1.14</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | <b>0.98</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |


**Diode Compression**

|       |           |
|-------|-----------|
| DCP X | <b>99</b> |
| DCP Y | <b>99</b> |
| DCP Z | <b>99</b> |

**Sensor Offset**

|                            |            |           |
|----------------------------|------------|-----------|
| Probe Tip to Sensor Center | <b>2.1</b> | <b>mm</b> |
|----------------------------|------------|-----------|

## Conversion Factor Assessment



**Head**                      **835 MHz**                       $\epsilon_r = 41.5 \pm 5\%$                        $\sigma = 0.90 \pm 5\%$  mho/m

Valid for f=793-877 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X


|         |                              |                  |             |
|---------|------------------------------|------------------|-------------|
| ConvF X | <b>6.5</b> $\pm 9.5\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>6.5</b> $\pm 9.5\%$ (k=2) | Alpha            | <b>0.35</b> |
| ConvF Z | <b>6.5</b> $\pm 9.5\%$ (k=2) | Depth            | <b>1.46</b> |

**Head**                      **1900 MHz**                       $\epsilon_r = 40.0 \pm 5\%$                        $\sigma = 1.40 \pm 5\%$  mho/m

Valid for f=1805-1995 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

|         |                              |                  |             |
|---------|------------------------------|------------------|-------------|
| ConvF X | <b>4.7</b> $\pm 9.5\%$ (k=2) | Boundary effect: |             |
| ConvF Y | <b>4.7</b> $\pm 9.5\%$ (k=2) | Alpha            | <b>0.22</b> |
| ConvF Z | <b>4.7</b> $\pm 9.5\%$ (k=2) | Depth            | <b>3.48</b> |

## Conversion Factor Assessment



**Body**                      **835 MHz**                       $\epsilon_r = 55.2 \pm 5\%$                        $\sigma = 0.97 \pm 5\% \text{ mho/m}$

Valid for f=793-877 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


|         |                              |                   |
|---------|------------------------------|-------------------|
| ConvF X | <b>6.1</b> $\pm 9.5\%$ (k=2) | Boundary effect:  |
| ConvF Y | <b>6.1</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.24</b> |
| ConvF Z | <b>6.1</b> $\pm 9.5\%$ (k=2) | Depth <b>2.00</b> |

**Body**                      **1900 MHz**                       $\epsilon_r = 53.3 \pm 5\%$                        $\sigma = 1.52 \pm 5\% \text{ mho/m}$

Valid for f=1805-1995 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                              |                   |
|---------|------------------------------|-------------------|
| ConvF X | <b>4.6</b> $\pm 9.5\%$ (k=2) | Boundary effect:  |
| ConvF Y | <b>4.6</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.24</b> |
| ConvF Z | <b>4.6</b> $\pm 9.5\%$ (k=2) | Depth <b>2.64</b> |

## Conversion Factor Assessment



Body 900 MHz  $\epsilon_r = 55.0 \pm 5\%$   $\sigma = 1.05 \pm 5\%$  mho/m

Valid for f=855-945 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


|         |                      |                  |
|---------|----------------------|------------------|
| ConvF X | 6.1 $\pm$ 9.5% (k=2) | Boundary effect: |
| ConvF Y | 6.1 $\pm$ 9.5% (k=2) | Alpha 0.27       |
| ConvF Z | 6.1 $\pm$ 9.5% (k=2) | Depth 1.82       |

Body 1800 MHz  $\epsilon_r = 53.3 \pm 5\%$   $\sigma = 1.52 \pm 5\%$  mho/m

Valid for f=1710-1890 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                      |                  |
|---------|----------------------|------------------|
| ConvF X | 4.7 $\pm$ 9.5% (k=2) | Boundary effect: |
| ConvF Y | 4.7 $\pm$ 9.5% (k=2) | Alpha 0.23       |
| ConvF Z | 4.7 $\pm$ 9.5% (k=2) | Depth 2.99       |

## Conversion Factor Assessment



**Head**      **2450 MHz**       $\epsilon_r = 39.2 \pm 5\%$        $\sigma = 1.80 \pm 5\%$  mho/m

Valid for f=2400-2500 MHz with Head Tissue Simulating Liquid according to EN 60381, P1528-200X

|         |                              |                   |
|---------|------------------------------|-------------------|
| ConvF X | <b>4.5</b> $\pm 9.5\%$ (k=2) | Boundary effect:  |
| ConvF Y | <b>4.5</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.40</b> |
| ConvF Z | <b>4.5</b> $\pm 9.5\%$ (k=2) | Depth <b>1.62</b> |

**Body**      **2450 MHz**       $\epsilon_r = 52.7 \pm 5\%$        $\sigma = 1.95 \pm 5\%$  mho/m

Valid for f=2400-2500 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

|         |                              |                   |
|---------|------------------------------|-------------------|
| ConvF X | <b>4.2</b> $\pm 9.5\%$ (k=2) | Boundary effect:  |
| ConvF Y | <b>4.2</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.32</b> |
| ConvF Z | <b>4.2</b> $\pm 9.5\%$ (k=2) | Depth <b>1.98</b> |

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, http://www.speag.com

## Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ES3DV2

Serial Number:

3019

Place of Assessment:

Zurich

Date of Assessment:

October 13, 2003

Probe Calibration Date:

October 9, 2003

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



ES3DV2-SN:3019

October 13, 2003



Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, http://www.speag.com

### Dosimetric E-Field Probe ES3DV2 SN:3019

Conversion factor ( $\pm$  standard deviation)

|         |       |               |                                                                               |
|---------|-------|---------------|-------------------------------------------------------------------------------|
| 150 MHz | ConvF | 8.7 $\pm$ 8 % | $\epsilon_r = 52.3 \pm 5\%$<br>$\sigma = 0.76 \pm 5\%$ mho/m<br>(head tissue) |
| 150 MHz | ConvF | 8.3 $\pm$ 8 % | $\epsilon_r = 61.9 \pm 5\%$<br>$\sigma = 0.80 \pm 5\%$ mho/m<br>(body tissue) |
| 450 MHz | ConvF | 7.4 $\pm$ 8 % | $\epsilon_r = 43.5 \pm 5\%$<br>$\sigma = 0.87 \pm 5\%$ mho/m<br>(head tissue) |
| 450 MHz | ConvF | 7.3 $\pm$ 8 % | $\epsilon_r = 56.7 \pm 5\%$<br>$\sigma = 0.94 \pm 5\%$ mho/m<br>(body tissue) |

ES3DV2-SN:3019

October 13, 2003

## Certificate of Calibration Verification

|                    |                      |
|--------------------|----------------------|
| Description of EUT | Tuned Dipole Antenna |
| EUT Model Number   | D-1800-S-1           |
| EUT Serial Number  | BCL-049              |
| Center Frequency   | 1800 MHz             |

Calibration Date: 12 April 2004

### Testing conditions:

per P1528/D1.2:2003:

|                                |       |
|--------------------------------|-------|
| Ambient Temperature (18-25 °C) | 23 °C |
| Ambient Humidity               | 43%   |

|                                                                          |       |
|--------------------------------------------------------------------------|-------|
| Liquid Temperature at start of measurements:( $\leq 2^{\circ}\text{C}$ ) | 21 °C |
| Liquid temperature at end of measurements:                               | 21 °C |

|                                     |                      |
|-------------------------------------|----------------------|
| Date and time at beginning of test: | 2004-04-09-16:20 PST |
| Date and time at beginning of test: | 2004-04-09-19:40 PST |

### Equipment used for measurements

|                                |                                                      |             |             |
|--------------------------------|------------------------------------------------------|-------------|-------------|
| Network Analyzer               | HP                                                   | 8752C       | 1 Nov 2002  |
| Impedance adapter              | AGILENT                                              | 43961A      | 31 Oct 2003 |
| Short Reference                | HP                                                   | 04191-85300 | 31 Oct 2003 |
| Open Reference                 | HP                                                   | 04191-85302 | 31 Oct 2003 |
| Load Reference                 | HP                                                   | 04191-85301 | 31 Oct 2003 |
| Signal Generator               | HP                                                   | 83650B      | 29 Feb 2004 |
| Calibration Cable:             | SMA Utiflex, 3.05 meter cable S/N 99E1206 (Number 8) |             |             |
| Phantom Model:                 | SAM                                                  |             |             |
| Liquid:                        | 1800 MHz, Head Liquid                                |             |             |
| Liquid Validation Date:        | 12 April 2004                                        |             |             |
| Quantity of Liquid in Phantom: | 19.8 Liters                                          |             |             |

### Measurement Procedure

In accordance with IEEE P1528/D1.2:2003, 8.3.4, 8.2.3 through 8.2.4

## Liquid Validation

| Instrument                 | Manufacturer Model |        | Calibrated |
|----------------------------|--------------------|--------|------------|
| Network Analyzer           | HP                 | 4396B  | 1 Nov 2002 |
| Dielectric Probe Kit,      | Agilent            | 85070C | Each Use   |
| H <sub>2</sub> O, 18 M-Ohm | BACL               |        | Each Use   |
| Probe, SAR 10 kHz - 6 GHz  | SPEAG              | ES3DV2 | 9 Oct 2003 |

### Attestation:

I hereby attest that the equipment are suitable for the performance requirements of IEEE P1528/D1.2:2003 and the personnel operating the test equipment and measurements are properly trained to perform the verification of this calibration procedure set forth in IEEE P1528/D1.2:2003.

The validation antenna herein meets the minimum requirements of 20 dB insertion loss



2004-04-12

\_\_\_\_\_  
Hans T. Mellberg  
Engineering Manager

\_\_\_\_\_  
Date

## 1800 MHz Head Liquid validation

Date : 12APR2004

Ambient Temp = 23 °C

Liquid Temp = 22 °C

| Frequency       | e'      | e''     | $\sigma$ ( $\sigma = 2\pi f \epsilon_0 \epsilon''$ ) |
|-----------------|---------|---------|------------------------------------------------------|
| 1850000000.0000 | 38.8246 | 13.2534 |                                                      |
| 1852000000.0000 | 38.7736 | 13.2429 |                                                      |
| 1854000000.0000 | 38.8400 | 13.2576 |                                                      |
| 1856000000.0000 | 38.8463 | 13.2425 |                                                      |
| 1858000000.0000 | 38.8167 | 13.2672 |                                                      |
| 1860000000.0000 | 38.8129 | 13.2552 |                                                      |
| 1862000000.0000 | 38.8118 | 13.2476 |                                                      |
| 1864000000.0000 | 38.7654 | 13.2345 |                                                      |
| 1866000000.0000 | 38.7686 | 13.2633 |                                                      |
| 1868000000.0000 | 38.7997 | 13.2690 |                                                      |
| 1870000000.0000 | 38.7262 | 13.2308 |                                                      |
| 1872000000.0000 | 38.7413 | 13.2642 |                                                      |
| 1874000000.0000 | 38.7458 | 13.2802 |                                                      |
| 1876000000.0000 | 38.7127 | 13.2833 |                                                      |
| 1878000000.0000 | 38.7145 | 13.2799 |                                                      |
| 1880000000.0000 | 38.7380 | 13.2633 |                                                      |
| 1882000000.0000 | 38.7086 | 13.2820 |                                                      |
| 1884000000.0000 | 38.7111 | 13.2991 |                                                      |
| 1886000000.0000 | 38.7184 | 13.2656 |                                                      |
| 1888000000.0000 | 38.7086 | 13.2724 |                                                      |
| 1890000000.0000 | 38.6697 | 13.2703 |                                                      |
| 1892000000.0000 | 38.6773 | 13.3051 |                                                      |
| 1894000000.0000 | 38.6729 | 13.2817 |                                                      |
| 1896000000.0000 | 38.6377 | 13.2805 |                                                      |
| 1898000000.0000 | 38.6113 | 13.2648 |                                                      |
| 1900000000.0000 | 38.6019 | 13.2714 | 1.40                                                 |
| 1902000000.0000 | 38.5554 | 13.2951 |                                                      |
| 1904000000.0000 | 38.5535 | 13.2851 |                                                      |
| 1906000000.0000 | 38.5103 | 13.3424 |                                                      |
| 1908000000.0000 | 38.5402 | 13.3692 |                                                      |
| 1910000000.0000 | 38.5162 | 13.3760 |                                                      |
| 1912000000.0000 | 38.4971 | 13.3857 |                                                      |
| 1914000000.0000 | 38.5126 | 13.3651 |                                                      |
| 1916000000.0000 | 38.4920 | 13.3817 |                                                      |
| 1918000000.0000 | 38.5463 | 13.3665 |                                                      |
| 1920000000.0000 | 38.5063 | 13.3804 |                                                      |
| 1922000000.0000 | 38.4973 | 13.3868 |                                                      |
| 1924000000.0000 | 38.5244 | 13.3470 |                                                      |
| 1926000000.0000 | 38.5362 | 13.3583 |                                                      |
| 1928000000.0000 | 38.5352 | 13.3774 |                                                      |
| 1930000000.0000 | 38.5427 | 13.3676 |                                                      |
| 1932000000.0000 | 38.5433 | 13.3562 |                                                      |
| 1934000000.0000 | 38.5374 | 13.3814 |                                                      |
| 1936000000.0000 | 38.5717 | 13.4048 |                                                      |
| 1938000000.0000 | 38.5057 | 13.4235 |                                                      |
| 1940000000.0000 | 38.5314 | 13.4375 |                                                      |
| 1942000000.0000 | 38.5104 | 13.4338 |                                                      |
| 1944000000.0000 | 38.4827 | 13.4285 |                                                      |
| 1946000000.0000 | 38.4545 | 13.4411 |                                                      |
| 1948000000.0000 | 38.4227 | 13.4385 |                                                      |
| 1950000000.0000 | 38.3682 | 13.4325 |                                                      |

System Validation for 1900 MHz Head Liquid (Ambient Temp = 23 C, Liquid Temp = 22 C,  
Forward Power = 20.42 dBm, 4/12/2004)


SAM Phantom: Flat Section; Position: (90°,90°); Frequency: 1900 MHz

Probe: E55DV2 - SN3019, Conn(F(4.70,4.70,4.70), Crest factor: 1.0, Head Liquid 1900 MHz:  $\sigma = 1.40 \text{ mho/m}$ ,  $\epsilon = 40.0$ ,  $\rho = 1.00 \text{ g/cm}^3$


Obs SAR(1g): 4.18 mW/g, SAR (10g): 2.21 mW/g, (Worst-case extrapolation)

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

PowerShift: 0.01 dB



# Insertion Loss Plot S11



## Smith Chart

CH1 A/R F5c1 500 mV

C1

↑

IF BW 90 kHz  
START 1.62 GHz

POWER 0 dBm

SWP 70 MHz  
STOP 1.82 GHzSELECT  
LETTER

SPACE

BACK  
SPACEERASE  
TITLE

DONE

STOR DEV  
(015K)

CANCEL

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

Client **Bay Area Comp. Lab (BACL)**

## CALIBRATION CERTIFICATE

Object(s) **D900V2 - SN:122**

Calibration procedure(s) **QA CAL-05.v2  
Calibration procedure for dipole validation kits**

Calibration date: **October 3, 2003**


Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                | ID #       | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|---------------------------|------------|-------------------------------------------|------------------------|
| Power sensor HP 8481A     | MY41092317 | 18-Oct-02 (Agilent, No. 20021018)         | Oct-04                 |
| Power sensor HP 8481A     | US37292783 | 30-Oct-02 (METAS, No. 252-0236)           | Oct-03                 |
| Power meter EPM E442      | GB37480704 | 30-Oct-02 (METAS, No. 252-0236)           | Oct-03                 |
| RF generator R&S SML-03   | 100698     | 27-Mar-2002 (R&S, No. 20-92389)           | In house check: Mar-05 |
| Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (Agilent, No. 24BR1033101)      | In house check: Oct 03 |

|                | Name           | Function            | Signature                                                                             |
|----------------|----------------|---------------------|---------------------------------------------------------------------------------------|
| Calibrated by: | Judith Mueller | Technician          |  |
| Approved by:   | Katja Pokovic  | Laboratory Director |  |

Date issued: October 9, 2003

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.



Schmid & Partner Engineering AG

**s p e a g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, <http://www.speag.com>

# DASY

## Dipole Validation Kit

Type: D900V2

Serial: 122

Manufactured: July 4, 2001  
Calibrated: October 3, 2003

## **1. Measurement Conditions**

The measurements were performed in the flat section of the SAM twin phantom filled with **head simulating solution** of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>42.3</b>       | $\pm 5\%$ |
| Conductivity           | <b>0.96 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.6 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was  $250\text{mW} \pm 3\%$ . The results are normalized to 1W input power.

## **2. SAR Measurement with DASY4 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|                                                  |                                                            |
|--------------------------------------------------|------------------------------------------------------------|
| averaged over $1\text{ cm}^3$ (1 g) of tissue:   | <b>10.2 mW/g <math>\pm 16.8\%</math> (k=2)<sup>1</sup></b> |
| averaged over $10\text{ cm}^3$ (10 g) of tissue: | <b>6.60 mW/g <math>\pm 16.2\%</math> (k=2)<sup>1</sup></b> |

### **3. Dipole Impedance and Return Loss**

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

|                      |                 |                                       |
|----------------------|-----------------|---------------------------------------|
| Electrical delay:    | <b>1.409 ns</b> | (one direction)                       |
| Transmission factor: | <b>0.983</b>    | (voltage transmission, one direction) |

The dipole was positioned at the flat phantom sections according to section 1 and the distance spacer was in place during impedance measurements.

|                                 |                                |
|---------------------------------|--------------------------------|
| Feedpoint impedance at 900 MHz: | $\text{Re}\{Z\} = 50.8 \Omega$ |
|---------------------------------|--------------------------------|

|  |                                |
|--|--------------------------------|
|  | $\text{Im}\{Z\} = -5.7 \Omega$ |
|--|--------------------------------|

|                        |                 |
|------------------------|-----------------|
| Return Loss at 900 MHz | <b>-24.8 dB</b> |
|------------------------|-----------------|

### **4. Measurement Conditions**

The measurements were performed in the flat section of the SAM twin phantom filled with **body simulating solution** of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>54.4</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.04 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.3 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

## **5. SAR Measurement with DASY4 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm<sup>3</sup> (1 g) of tissue:      **10.7 mW/g ± 16.8 % (k=2)<sup>2</sup>**

averaged over 10 cm<sup>3</sup> (10 g) of tissue:      **6.92 mW/g ± 16.2 % (k=2)<sup>2</sup>**

## **6. Dipole Impedance and Return Loss**

The dipole was positioned at the flat phantom sections according to section 4 and the distance spacer was in place during impedance measurements.

Feedpoint impedance at 900 MHz:      **Re{Z} = 47.1 Ω**

**Im {Z} = -6.7 Ω**

Return Loss at 900 MHz      **-22.6 dB**

## **7. Handling**

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

## **8. Design**

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

## **9. Power Test**

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN122**

Communication System: CW-900; Frequency: 900 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz ( $\sigma = 0.96$  mho/m,  $\epsilon_r = 42.26$ ,  $\rho = 1000$  kg/m<sup>3</sup>)

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.6, 6.6, 6.6); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.8 Build 60

**$P_{in} = 250$  mW;  $d = 15$  mm/Area Scan (81x81x1); Measurement grid:  $dx=15$ mm,  $dy=15$ mm**

Reference Value = 55.6 V/m

Power Drift = 0.003 dB

Maximum value of SAR = 2.75 mW/g

**$P_{in} = 250$  mW;  $d = 15$  mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid:  $dx=5$ mm,  $dy=5$ mm,  $dz=5$ mm**


Peak SAR (extrapolated) = 3.81 W/kg


SAR(1 g) = 2.55 mW/g; SAR(10 g) = 1.65 mW/g

Reference Value = 55.6 V/m

Power Drift = 0.003 dB

Maximum value of SAR = 2.76 mW/g







Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN122**

Communication System: CW-900; Frequency: 900 MHz; Duty Cycle: 1:1  
Medium: Muscle 900 MHz ( $\sigma = 1.04$  mho/m,  $\epsilon_r = 54.38$ ,  $\rho = 1000$  kg/m<sup>3</sup>)

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.3, 6.3, 6.3); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.8 Build 60

**Pin = 250 mW; d = 15 mm/Area Scan (81x81x1):** Measurement grid: dx=15mm, dy=15mm

Reference Value = 55 V/m

Power Drift = 0.0 dB

Maximum value of SAR = 2.87 mW/g

**Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm


Peak SAR (extrapolated) = 3.92 W/kg

SAR(1 g) = 2.67 mW/g; SAR(10 g) = 1.73 mW/g


Reference Value = 55 V/m

Power Drift = 0.0 dB

Maximum value of SAR = 2.88 mW/g



Body





**835MHZ Body Liquid validation****Ambient Temp=23 °C, Liquid Temp=22°C 3/24/2004**

| frequency      | $\epsilon'$ | $\epsilon''$ |
|----------------|-------------|--------------|
| 815000000.0000 | 53.2565     | 21.4858      |
| 815800000.0000 | 53.2224     | 21.4201      |
| 816600000.0000 | 53.2008     | 21.4192      |
| 817400000.0000 | 53.2638     | 21.4013      |
| 818200000.0000 | 53.2472     | 21.3898      |
| 819000000.0000 | 53.2206     | 21.3509      |
| 819800000.0000 | 53.2651     | 21.3421      |
| 820600000.0000 | 53.2031     | 21.3342      |
| 821400000.0000 | 53.2569     | 21.3262      |
| 822200000.0000 | 53.2297     | 21.3126      |
| 823000000.0000 | 53.1929     | 21.3050      |
| 823800000.0000 | 53.2089     | 21.3010      |
| 824600000.0000 | 53.1688     | 21.2968      |
| 825400000.0000 | 53.2007     | 21.2836      |
| 826200000.0000 | 53.1317     | 21.2780      |
| 827000000.0000 | 53.1724     | 21.2616      |
| 827800000.0000 | 53.1547     | 21.2583      |
| 828600000.0000 | 53.1444     | 21.2514      |
| 829400000.0000 | 53.1139     | 21.2381      |
| 830200000.0000 | 53.0801     | 21.2127      |
| 831000000.0000 | 53.0866     | 20.9850      |
| 831800000.0000 | 53.0999     | 20.9766      |
| 832600000.0000 | 53.0462     | 20.7568      |
| 833400000.0000 | 53.0726     | 20.5038      |
| 834200000.0000 | 53.0510     | 20.4107      |
| 835000000.0000 | 53.0629     | 20.2467      |
| 835800000.0000 | 53.0710     | 20.0301      |
| 836600000.0000 | 53.0212     | 19.9797      |
| 837400000.0000 | 53.0293     | 19.9712      |
| 838200000.0000 | 53.0265     | 19.9690      |
| 839000000.0000 | 52.9767     | 19.9631      |
| 839800000.0000 | 52.0030     | 19.9504      |
| 840600000.0000 | 52.9823     | 19.8921      |
| 841400000.0000 | 52.9904     | 19.8836      |
| 842200000.0000 | 52.9084     | 19.8830      |
| 843000000.0000 | 52.9562     | 19.8787      |
| 843800000.0000 | 52.8925     | 19.8652      |
| 844600000.0000 | 52.9097     | 19.8547      |
| 845400000.0000 | 52.9287     | 19.8338      |
| 846200000.0000 | 52.8795     | 19.7933      |
| 847000000.0000 | 52.8465     | 19.7809      |
| 847800000.0000 | 52.9121     | 19.7758      |
| 848600000.0000 | 52.9329     | 19.7745      |
| 849400000.0000 | 52.8786     | 19.7659      |
| 850200000.0000 | 52.8850     | 19.7540      |
| 851000000.0000 | 52.8702     | 19.7495      |
| 851800000.0000 | 52.8997     | 19.6976      |
| 852600000.0000 | 52.8670     | 19.6807      |
| 853400000.0000 | 52.8648     | 19.6731      |
| 854200000.0000 | 52.8339     | 19.6687      |
| 855000000.0000 | 52.8313     | 19.6465      |

$$\sigma = \omega \epsilon_0 \epsilon'' = 2 \pi f \epsilon_0 \epsilon'' = 0.9405$$

$$\text{where } f = 835 \times 10^6$$

$$\epsilon_0 = 8.854 \times 10^{-12}$$

$$\epsilon'' = 20.2467$$

**835 MHZ Head Liquid Validation****Ambient Temp=23 °C, Liquid Temp=22 °C , 3/24/2004**

| frequency      | e'      | e''     |
|----------------|---------|---------|
| 815000000.0000 | 40.5310 | 19.3508 |
| 815800000.0000 | 40.5616 | 19.3415 |
| 816600000.0000 | 40.5384 | 19.3411 |
| 817400000.0000 | 40.5010 | 19.3403 |
| 818200000.0000 | 40.4908 | 19.3269 |
| 819000000.0000 | 40.4811 | 19.3271 |
| 819800000.0000 | 40.4675 | 19.3255 |
| 820600000.0000 | 40.4674 | 19.3210 |
| 821400000.0000 | 40.4669 | 19.3177 |
| 822200000.0000 | 40.4667 | 19.3176 |
| 823000000.0000 | 40.4622 | 19.3167 |
| 823800000.0000 | 40.4596 | 19.3079 |
| 824600000.0000 | 40.4571 | 19.3048 |
| 825400000.0000 | 40.4554 | 19.2976 |
| 826200000.0000 | 40.4546 | 19.2917 |
| 827000000.0000 | 40.4464 | 19.2883 |
| 827800000.0000 | 40.4445 | 19.2865 |
| 828600000.0000 | 40.4346 | 19.2844 |
| 829400000.0000 | 40.4337 | 19.2795 |
| 830200000.0000 | 40.4286 | 19.2791 |
| 831000000.0000 | 40.4253 | 19.2666 |
| 831800000.0000 | 40.4251 | 19.2611 |
| 832600000.0000 | 40.4146 | 19.2558 |
| 833400000.0000 | 40.4128 | 19.2515 |
| 834200000.0000 | 40.4096 | 19.2477 |
| 835000000.0000 | 40.4054 | 19.2411 |
| 835800000.0000 | 40.4045 | 19.2393 |
| 836600000.0000 | 40.4034 | 19.2341 |
| 837400000.0000 | 40.4030 | 19.2281 |
| 838200000.0000 | 40.3998 | 19.2188 |
| 839000000.0000 | 40.3986 | 19.2150 |
| 839800000.0000 | 40.3954 | 19.2141 |
| 840600000.0000 | 40.3937 | 19.2133 |
| 841400000.0000 | 40.3930 | 19.2115 |
| 842200000.0000 | 40.3967 | 19.2095 |
| 843000000.0000 | 40.3978 | 19.2080 |
| 843800000.0000 | 40.3953 | 19.2071 |
| 844600000.0000 | 40.3974 | 19.2001 |
| 845400000.0000 | 40.3961 | 19.1988 |
| 846200000.0000 | 40.3971 | 18.1974 |
| 847000000.0000 | 40.3885 | 18.1810 |
| 847800000.0000 | 40.3863 | 18.1792 |
| 848600000.0000 | 40.3882 | 18.1712 |
| 849400000.0000 | 40.3865 | 18.1698 |
| 850200000.0000 | 40.3861 | 18.1637 |
| 851000000.0000 | 40.3831 | 18.1600 |
| 851800000.0000 | 40.3826 | 18.1596 |
| 852600000.0000 | 40.3873 | 18.1502 |
| 853400000.0000 | 40.3850 | 18.1464 |
| 854200000.0000 | 40.3715 | 18.1389 |
| 855000000.0000 | 40.3645 | 18.1308 |

$$\sigma = \omega \varepsilon_0 \varepsilon'' = 2 \pi f \varepsilon_0 \varepsilon'' = 0.8938$$

$$\text{where } f = 835 \times 10^6$$

$$\varepsilon_0 = 8.854 \times 10^{-12}$$

$$\varepsilon'' = 19.2411$$