

Test Equipment List with Calibration Data

Test report number(s):
50530-30221

Date of test:
04/2003

Type	Inv.-No.	Type Designation	Serial Number	Manufacturer	Calibration Organization	Date of Calibration	
						Last	Next
Test receiver	1025	ESVP	881120/024	Rohde & Schwarz	Rohde & Schwarz	08/2002	02/2004
Test receiver	1028	ESHS10	860043/016	Rohde & Schwarz	Rohde & Schwarz	03/2003	09/2004
Test receiver	1569	ESMI	839379/013	Rohde & Schwarz	Rohde & Schwarz	08/2002	02/2004
EMI Test Receiver	1711	ESPI7	836914/0002	Rohde & Schwarz	Rohde & Schwarz	09/2002	03/2004
Spectrum analyser	1666	FSP30	100063	Rohde & Schwarz	Rohde & Schwarz	05/2002	11/2003
V-network	1059	ESH3-Z5	894785/005	Rohde & Schwarz	Rohde & Schwarz	06/2002	06/2005
V-network	1060	ESH3-Z5	862770/021	Rohde & Schwarz	Rohde & Schwarz	05/2002	05/2005
V-network	1218	ESH3-Z5	830952/025	Rohde & Schwarz	Rohde & Schwarz	05/2002	05/2005
Loop antenna	1016	HFH2-Z2	882964/0001	Rohde & Schwarz	Rohde & Schwarz	03/2003	09/2004
Double ridged waveguide horn antenna	1516	3115	9508-4553	EMCO Elektronik	ARC	01/2002	07/2003
Biconical Antenna	1518	HK116	842204/01	Rohde & Schwarz	Rohde & Schwarz	04/2002	10/2003
Logarithmic-periodic antenna	1519	HL223	841516/23	Rohde & Schwarz	Rohde & Schwarz	04/2002	10/2003

Note: Date of next calibration contains maximum tolerance of six months for devices with calibration cycles less than three years.