

Declaration of Equality

To whom it may concern,

With regard to the application for type approval services worldwide, we hereby declare that the model PERU 098 (part number 10432917) and model PERU SB (part number 10849847) are accordant in firmware.

The transmitting units of these two models are accordant in hardware.

- These two models only differ in the pre-amplifier hardware.

The detail differences are:

1. PERU 098 pre-amplifier has 4 electrode connections
2. PERU SB pre-amplifier has 3 electrode connections (the black lead "LA" is not equipped)

Declared by:



Dr. Reiner Bachus

(Head of Quality Management MR)



Dr. Leonhard Söldner

(Head of Continuous Customer Care)

Date of issue: 16.03.2015

Stamp: Siemens Aktiengesellschaft
Healthcare Sector
Magnet-Resonanz
Karl-Schall-Str. 2
D-91052 Erlangen

Siemens AG
Healthcare
Leitung: Bernd Montag

Allee am Röthelheimpark 2
91052 Erlangen
Deutschland

Tel.: +49 (9131) 84 0

Siemens Aktiengesellschaft: Vorsitzender des Aufsichtsrats: Gerhard Cromme; Vorstand: Joe Kaeser, Vorsitzender; Roland Busch, Lisa Davis, Klaus Helmrich, Janina Kugel, Siegfried Russwurm, Ralf P. Thomas
Sitz der Gesellschaft: Berlin und München, Deutschland; Registergericht: Berlin Charlottenburg, HRB 12300, München, HRB 6684
WEEE-Reg.-Nr. DE 23691322