

Response to TCB Findings

The test setup photos show ferrite type accessories on the interconnecting cables of the "Remote" unit. Please clarify the following,

- a. Exactly on which cables were they necessary?
- b. What were the part numbers of those accessories?
- c. Did the main unit need them as well, if so on which cables?
- d. The manual must be clear on the requirement of use, for these accessories per 15.27. Please point out the section in manual to look for warnings about this.

Attached are the appendices of the Veris Operator's manual addressing the references to the special cables needed for emissions compliance.

The Wireless appendix discusses operation of the wireless connection between Main Unit and Remote Display. A caution has been added calling for the cables with ferrites if the serial data or external video connections are made. There is also a reference in this appendix to the Accessories Appendix which lists the cables under Miscellaneous Accessories. These additions should satisfy the requirements of Part 15.27. The TCB was also wondering if these cables were needed on the Main Unit. The answer is no, the connectors are not accessible on the Main Unit.

See updated Manual Appendix "D" and "G".