

Proposed labeling is as follows:

- The Argus II implant contains a transmitter, but it is exempt from certification per 15.201(a).
- The Argus II Glasses and OR coil contain a transmitter, and as such, are subject to FCC labeling requirements per Section 2.925(a). Per Section 15.19(a)(4), we have placed the FCC ID on a label permanently affixed to the RF electronics housing. Refer to Figures 4 (artwork), 5 (photo showing placement of label on RF electronics housing on glasses), and 6 (photo showing placement of label on RF electronics housing of OR coil). Because of the small size of the device, the required compliance information shall be placed in the Patient Manual rather than on the label in accordance with Section 15.19(a)(5).

Photographs showing the exterior appearance, the construction, the component placement on the chassis, the chassis assembly, the antenna used with the device, and the user controls are shown in Figures 7-15.




Figure 4. Artwork for FCC ID Label



Figure 5. FCC ID marked on RF electronics housing on Argus II glasses.



Figure 6. FCC ID marked on RF electronics housing on Argus II OR Coil.