

Question from TCB:

There is reference to an external attenuator in a couple of sections of the test report: specifically on pages 44, 55 and 84. The calibration of this attenuator is not discussed within the report. Please provide documentation of all test equipment used with this equipment in accordance with the requirements of ANSI C63.4 and ISO 17025.

Answer from test lab:

The 10db external attenuator was verified prior to testing with SA AN02869, siggen AN02227 and AN2551, cable 02947.

Information of 10db attenuator:

Manufacture: Weinschel Corp

Model: 33-10-34

Serial number: BFG198

Asset #	Description	Model	Calibration Date	Cal Due Date
AN02869	Spectrum Analyzer	E4440A	2/6/2013	2/6/2015
AN02227	Generator, Signal	2024	7/6/2011	7/6/2013
AN02551	Generator, Signal	8673M	4/4/2013	4/4/2015
AN02947	Cable	32022-2-2909K- 36TC	8/8/2011	8/8/2013