



March 14, 2007

Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, MD 21046

Attn: OET Dept.

Ref: Adhoc mode disablement for FCC Class II Permissive change for FCC ID: STJ80411396001 –  
Letter of Attestation  
Applicant: Hospira, Inc.

Dear Examiner,

This is to attest that the Custom DWL-AG132 Wireless USB Adapter is to be used with medical Infusion pumps as a client device under the control of a master device in infrastructure mode only. The adhoc networking mode has been disabled in the application software and is not available to the user.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mozammil Awan".

Mozammil Awan  
03/14/07  
Manager, Advanced Engineering