

Canohès, 2015-06-18

**Subject.: Acknowledgment of FCC Modular approval requirements**

**Applicant: PYRESCOM**

**Product: 13.56 Mhz RFID Module**

**Model: B-AV2-13MHZ**

**FCC ID: SZ4-3XX090601**

To Whom It May Concern:

A protection and surveillance cabinet for automatic external defibrillators, Aivia 300, Aivia 310, Aivia 320 and Aivia330.

Modular Approval Requirement	Yes	No
(i) The radio elements of the modular transmitter must have their own shielding. The physical crystal and tuning capacitors may be located external to the shielded radio elements.		X
(ii) The modular transmitter must have buffered modulation/data inputs (if such inputs are provided) to ensure that the module will comply with part 15 requirements under conditions of excessive data rates or over-modulation.		X
(iii) The modular transmitter must have its own power supply regulation.		X
(iv) The modular transmitter must comply with the antenna and transmission system requirements of §§ 15.203, 15.204(b) and 15.204(c). The antenna must either be permanently attached or employ a "unique" antenna coupler (at all connections between the module and the antenna, including the cable). The "professional installation" provision of § 15.203 is not applicable to modules but can apply to limited modular approvals under paragraph (b) of this section.	X	
(v) The modular transmitter must be tested in a stand-alone configuration		X
(vi) The modular transmitter must be equipped with either a permanently affixed label or must be capable of electronically displaying its FCC identification number.	X	
(vii) The modular transmitter must comply with any specific rules or operating requirements that ordinarily apply to a complete transmitter and the manufacturer must provide adequate instructions along with the module to explain any such requirements. A copy of these instructions must be included in the application for equipment authorization.	X	
(viii) The modular transmitter must comply with any applicable RF exposure requirements in its final configuration.	X	

- (i) Dedicated electronic board needn't to be shield.
- (ii) RFID reader is use to read a tag. There is no excessive data rates or over-modulation.
- (iii) RFID reader is powered by host board, which have a regulated power supply.
- (v) Limited modular approval, tested in Aivia.

Regards,

GUICHET Alice, 2015-06-18



PYRESCOM - Mas des Tilleuls - B.P. 80051 - 66680 CANOHES - FRANCE Tél: +(33) 04 68 68 39 68 - Fax. +(33) 04 68 68 39 69  
direct@pyres.com - www.pyres.com

S.A.S au capital de 2 000 000€ - R.C. Perpignan 84 B 427 - A.P.E.2620 Z - SIRET 330 925 744 000 97 - N° TVA FR 92 330 925 744 - EORI : FR 33092574400097