



Philips Healthcare
1010 Murry Ridge Lane
Murrysville, PA 15668
Home Healthcare Solutions

Date: June 27, 2017

Applicant:

Respironics Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668

Subject: Request for Class II Permissive change

This is to request a Class II permissive change for **FCC ID: THO-1116426**, originally granted on 6/17/2015.

The original grant is for a Limited Single Modular Approval for a radio module. This radio module is incorporated in our host equipment DreamStation CPAP Containing **FCC ID: THO 1116426**.

With the creation of the new radio accessories to be used with the identified host equipment, no changes were made to the Bluetooth radio module output power, modulation circuit, frequency determining circuitry, or by external adaptation circuitry that impacts the validity of the results of the original module's test reports.

The major changes filled under this application are adding 2 models of modem accessories 200601C and 200602C. **The manufacturer FCC ID for the radio modules will be used for the modem accessory as "Contains QIPEHSS-US".** A single modem, either Wi-Fi or cell modem, may be used with the base Bluetooth radio.

The necessary testing reports for the collocated new modem accessory have been added to the DreamStation submittal by Cetecom.

The new cell modem accessories are identical except that the 200602C unit has a number of parts not populated on the circuit board to eliminate the pulse oximetry function. They are:

R5, R7, R56, R57, R58, R59, R60, R66, R67, R68, R69, R70, R71, R72, R73, R74

C53, C54, C69, C91, C92, C93, C94

U7, U8, U9, U10, U11

J3

Regards,

Nicholas Cormas

Tel: 724.387.4000
Tel: 800.345.6443 (toll free, US only)

www.philips.com/healthcare