



Philips Healthcare

1010 Murry Ridge Lane

Murrysville, PA 15668

Home Healthcare Solutions

Declaration of Equality and Request for Class II Permissive Change (C2PC)

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

TO WHOM IT MAY CONCERN

Pursuant to the applicable paragraphs in CFR 47 FCC part §2.1043(b)(2) we hereby declare the following changes under Class II Permissive Change to the product:

Model Number: 200603C DreamStation Modem FCC ID:

BASE UNIT FCC ID: THO-1116426

We,
Respironics Inc
1010 Murry Ridge Lane
Murrysville, PA 15668

hereby declare the following Class 2 Permissive Change to the 200603C and 200604C related to the FCC regulations in the original certificate:

- The HVIN is unchanged.
- The external mechanical and internal electrical construction are unchanged.
 - No changes to the 200603C, 200604C PCB.
- The antenna changes in the product does not affect any of the RF characteristics of the originally certified product.
 - The integrated cellular module SARA-R410M (FCC ID : XPY2AGQN4NNN) is unchanged.
 - The minor change to the modem is changing the antenna dimensions slightly
 - The transmitter output power and spurious emission is unaffected by the cellular module upgrade as verified by testing.
 -

If you have any questions, please feel free to contact me at the address shown below.

Sincerely,

Tel: 724.387.4000
Tel: 800.345.6443 (toll free, US only)

www.philips.com/healthcare