

Limited Module Approval Request Letter

Subject: Limited Modular Approval

Date: 14-02-2012

FCC ID: TUK-MIR045

We,.. MIR s.r.l. Medical International Research authorize Eurofins Product Service GmbH to have our model: BlueMir approved under limited module approval authorization. The application of this module is specified to mobile host equipment. The requirements regulated in Public Notice DA00-1407 have been fulfilled and clearly explained below.

1. The requirement of buffered inputs: Buffered data inputs stage has been integrated in BlueMir.
2. The requirements of section 15.203 and 15.204(C): The requirements of antenna connector and spurious emission have been fulfilled. Please reference the exhibition Test Report.
3. The requirement of stand-alone test configuration: Please reference exhibition Test Configuration Photo for the stand-alone test configuration.
4. The requirement of labelling: The instruction on the labelling rule of the end product has been stated in the Users manual of this module. Please also see the exhibition Label Sample.
5. The requirement of compliance on specific rule or operating requirements: The required FCC rule has been fulfilled and all the instructions for maintaining compliance has been clearly stated in the Users Manual.
6. The requirement of RF exposure: Please refer exhibition RF Exposure for the compliance of MPE RF exposure rule.

Company name: MIR s.r.l. Medical International Research

Company Address: via del Maggiolino, 125



Signature:

Name: Paolo Sacco Boschetti

Job Title: Chairman