

Smørum, 2018-09-25

## Manufacturer's Statement

## **FCC ID location justification**

FCC ID: U28AUMRTRC

## TO WHOM IT MAY CONCERN

We, the undersigned, hereby submit the justification for placing the FCC ID in the Associated User Manual for the Hearing Aids having the above-mentioned FCC IDs.

According to KDB Publication 784748 D01, when the device is so small, or for such use that it is impracticable to label with a font size that is four points or larger (and the device does not utilize electronic labeling), then the FCC ID shall be placed in the user manual.

Hearing Aids having the above-mentioned FCC IDs are so small, and considering the dimensions of the equipment and the label area, it is impracticable to label FCC ID with a font size required on the devices.

Therefore the FCC ID label is placed in the User Manual.

Mystur

Yours sincerely,

Ole Myrtue

Regulatory Affairs Officer

Oticon A/S Kongebakken 9 DK-2765 Smoerum

Denmark