

Information about the Applicant	
Grantee (Company Name)	Verathon
FCC Contact Person	Tim M Chinowsky
Address	16932 Woodinville-Redmond Rd, Bldg A, Suite 109
City, State, Zip	Woodinville, WA 98072
Job Number	VERA0015
Model	FloSensor, P/N 0570-0175
FCC ID	U740570-0175
Agent	Northwest EMC, Inc.
Approval Type	Original, Modular approval
Equipment Class	DXT Part 15, Low Power Transceiver, Rx Verified
Rule Part	15.249

Overview

This application is for a limited modular approval of the Cypress Model CYWUSB6934 wireless USB chip in the Verathon FloSensor, P/N 0570-0175. The FloSensor is a device to detect urine flow rates in a bathroom. The radio operates in the 2400-2483.5 MHz band under 47 CFR part 15.249.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature

Dave Tolman, TCB Committee
11/15/2007