Report No. : ES/2007/50009 Page : 1 of 23

# **Contents**

## **APPENDIX**

1. Photographs of Test Setup	02
2. Photographs of EUT	06
3. Photographs of Accessories of EUT	09
4. DAE & Probe Calibration certificate	10
5. Uncertainty Analysis	20
6. Phantom Description	21
7. System Validation from Original equipment supplier	22

Report No. : ES/2007/50009 Page : 2 of 23

# 1. Photographs of Test Setup



Fig.1 Photograph of the SAR measurement System




Fig.2.1 Photograph of the Tissue Simulant liquid depth 15cm for Flat position

Page: 3 of 23




Fig.2.2 Photograph of the Tissue Simulant liquid depth 15cm for Flat position




Fig.3 Photo of the Bottom face of the notebook is paralleled with flat phantom,

Page: 4 of 23




Fig.4 Photo of the Bottom face of the notebook is paralleled with flat Phantom



Fig.5 Photo of the right side of the LCD of Notebook is vertical with flat phantom and spacing between EUT and Phantom- in contact 15mm.

Report No. : ES/2007/50009 Page : 5 of 23



Fig.6 Photo of the right side of the LCD of Notebook is vertical with flat phantom and spacing between EUT and Phantom- in contact 15mm.

Report No. : ES/2007/50009 Page : 6 of 23

2. Photographs of the EUT

## otographs of the Eor



Fig.7 Front view of EUT



Fig.8 Back view of EUT

Report No.: ES/2007/50009 Page: 7 of 23



Fig.9 Left Side of EUT



Fig.10 Right Side of EUT

Report No.: ES/2007/50009 Page: 8 of 23



Fig.11 Fold up of EUT

Report No. : ES/2007/50009 Page : 9 of 23

# 3. Photographs of the Accessories of EUT



Fig.12 Notebook charger

Page: 10 of 23

#### 4. DAE & Probe Calibration certificate

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst S Service suisse d'étalonnage C Servizio svizzero di taratura **Swiss Calibration Service** 

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

SGS (Auden)

Accreditation No.: SCS 108

Certificate No: DAE4-547\_Mar07 **CALIBRATION CERTIFICATE** DAE4 - SD 000 D04 BA - SN: 547 Object QA CAL-06.v12 Calibration procedure(s) Calibration procedure for the data acquisition electronics (DAE) March 5, 2007 Calibration date: Condition of the calibrated item In Tolerance This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) ID# Scheduled Calibration Primary Standards Cal Date (Calibrated by, Certificate No.) Fluke Process Calibrator Type 702 SN: 6295803 13-Oct-06 (Elcal AG, No: 5492) Oct-07 SN: 0810278 Keithley Multimeter Type 2001 03-Oct-06 (Elcal AG, No: 5478) Oct-07 Secondary Standards Check Date (in house) Scheduled Check Calibrator Box V1.1 SE UMS 006 AB 1002 15-Jun-06 (SPEAG, in house check) In house check Jun-07 Name Function Calibrated by: Daniel Steinacher Technician R&D Director Fin Bomholt Approved by: Issued: March 5, 2007 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: DAE4-547\_Mar07

Page 1 of 5

Report No. : ES/2007/50009 Page : 11 of 23

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst S Service suisse d'étalonnage C Servizio svizzero di taratura Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

SGS (Auden)

Certificate No: EX3-3526\_Aug06

Accreditation No.: SCS 108

Object	EX3DV3 - SN:3	526	
Calibration procedure(s)	CONTRACTOR OF THE PARTY OF THE	and QA CAL-14.v3 cedure for dosimetric E-field probes	
Calibration date:	August 25, 2006	6	
Condition of the calibrated item	In Tolerance		
			d book little a 700/
		ory facility: environment temperature (22 ± 3)°C and	a numidity < 70%.
Calibration Equipment used (M&	TE critical for calibration)	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Calibration Equipment used (M& Primary Standards Power meter E4419B	TE critical for calibration)  ID #  GB41293874	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557)	Scheduled Calibration Apr-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A	TE critical for calibration)  ID #  GB41293874  MY41495277	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)	Scheduled Calibration Apr-07 Apr-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)	Scheduled Calibration Apr-07 Apr-07 Apr-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 251-00558)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Apr-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 217-00593)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Apr-07  Aug-07  Aug-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 251-00558)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Apr-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00598) 10-Aug-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06)	Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Apr-07 Aug-07 Jan-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2 DAE4 Recondary Standards RF generator HP 8648C	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #  US3642U01700	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Aug-07  Jan-07  Jun-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2 DAE4 Recondary Standards RF generator HP 8648C	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Aug-07  Jan-07  Jun-07  Scheduled Check
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2 DAE4	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #  US3642U01700	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Aug-07  Jan-07  Jun-07  Scheduled Check  In house check: Nov-07
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards RF generator HP 8648C Network Analyzer HP 8753E	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #  US3642U01700  US37390585	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 271-00592) 4-Apr-06 (METAS, No. 217-00598) 10-Aug-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Apr-07  Aug-07  Jun-07  Scheduled Check  In house check: Nov-07  In house check: Nov 06
Calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator	TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 3013  SN: 654  ID #  US3642U01700  US37390585  Name	Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00598) 10-Aug-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05)	Scheduled Calibration  Apr-07  Apr-07  Apr-07  Aug-07  Apr-07  Aug-07  Jun-07  Scheduled Check  In house check: Nov-07  In house check: Nov 06

Certificate No: EX3-3526\_Aug06

Page 1 of 9

Page: 12 of 23

#### Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service sulsse d'étalonnage
Servizio svizzero di taratura

Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL tissue simulating liquid
NORMx,y,z sensitivity in free space
ConF sensitivity in TSL / NORMx,y,z
DCP diode compression point
Polarization φ φ rotation around probe axis

Polarization 9 9 rotation around an axis that is in the plane normal to probe axis (at

measurement center), i.e.,  $\vartheta = 0$  is normal to probe axis

#### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001

#### Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not effect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This
  linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of
  the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: EX3-3526\_Aug06 Page 2 of 9

Page: 13 of 23

EX3DV3 SN:3526

August 25, 2006

# Probe EX3DV3

SN:3526

Manufactured:

Last calibrated: Recalibrated: March 19, 2004

May 24, 2004

August 25, 2006

## Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ET3-3526\_Aug06

Page 3 of 9

Page: 14 of 23

EX3DV3 SN:3526 August 25, 2006

## DASY - Parameters of Probe: EX3DV3 SN:3526

Sensitivity in Fre	ensitivity in Free Space <sup>A</sup>			ompression <sup>E</sup>	3
NormX	<b>0.92</b> ± 10.1%	$\mu V/(V/m)^2$	DCP X	95 mV	
NormY	0.87 ± 10.1%	$\mu V/(V/m)^2$	DCP Y	<b>92</b> mV	
NormZ	0.85 ± 10.1%	$\mu V/(V/m)^2$	DCP Z	<b>94</b> mV	

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

### **Boundary Effect**

TSL	900 MHz	Typical SAR	gradient: 5 % per mm

Sensor Center to	Phantom Surface Distance	3.7 mm	4.7 mm
SAR <sub>be</sub> [%]	Without Correction Algorithm	2.1	0.9
SAR <sub>be</sub> [%]	With Correction Algorithm	0.1	0.1

#### TSL 1810 MHz Typical SAR gradient: 10 % per mm

Sensor Cente	r to Phantom Surface Distance	3.7 mm	4.7 mm
SAR <sub>be</sub> [%]	Without Correction Algorithm	1.5	0.8
SAR <sub>be</sub> [%]	With Correction Algorithm	0.1	0.6

#### Sensor Offset

Probe Tip to Sensor Center 1 mm

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>&</sup>lt;sup>A</sup> The uncertainties of NormX,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Page 8).

<sup>&</sup>lt;sup>B</sup> Numerical linearization parameter: uncertainty not required.


Report No. : ES/2007/50009 Page : 15 of 23

EX3DV3 SN:3526

August 25, 2006

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


Page: 16 of 23


EX3DV3 SN:3526

August 25, 2006

Receiving Pattern ( $\phi$ ),  $\vartheta$  = 0°







Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


Page: 17 of 23

#### EX3DV3 SN:3526

August 25, 2006

## Dynamic Range f(SAR<sub>head</sub>)

(Waveguide R22, f = 1800 MHz)




Uncertainty of Linearity Assessment: ± 0.6% (k=2)


Page: 18 of 23

#### EX3DV3 SN:3526

#### August 25, 2006

### **Conversion Factor Assessment**





f [MHz]	Validity [MHz] <sup>C</sup>	TSL	Permittivity	Conductivity	Alpha	Depth	ConvF	Uncertainty
900	± 50 / ± 100	Head	41.5 ± 5%	0.97 ± 5%	0.21	0.90	11.72	± 11.0% (k=2)
1810	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.11	1.33	9.61	± 11.0% (k=2)
1950	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.11	1.33	9.32	± 11.0% (k=2)
2450	± 50 / ± 100	Head	39.2 ± 5%	1.80 ± 5%	0.42	0.80	8.29	± 11.8% (k=2)
900	± 50 / ± 100	Body	55.0 ± 5%	1.05 ± 5%	0.21	0.89	11.63	± 11.0% (k=2)
1810	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.19	1.32	9.64	± 11.0% (k=2)
1950	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.15	1.55	9.26	± 11.0% (k=2)
2450	± 50 / ± 100	Body	52.7 ± 5%	1.95 ± 5%	0.40	0.52	8.30	± 11.8% (k=2)
5200	± 50 / ± 100	Body	49.0 ± 5%	5.30 ± 5%	0.45	1.85	3.17	± 13.1% (k=2)
5800	± 50 / ± 100	Body	48.2 ± 5%	6.00 ± 5%	0.25	1.85	2.66	± 13.1% (k=2)

 $<sup>^{\</sup>rm C}$  The validity of  $\pm$  100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


Report No. : ES/2007/50009 Page : 19 of 23

EX3DV3 SN:3526

August 25, 2006

## **Deviation from Isotropy in HSL**

Error (φ, θ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

Page: 20 of 23

 $\infty$ 

 $\infty$ 

 $\infty$ 

 $\infty$ 

 $\infty$ 

 $\infty$ 

331

 $\pm 2.3 \%$ 

 $\pm 1.2 \%$ 

 $\pm 1.1\%$ 

 $\pm 1.4\,\%$ 

 $\pm 1.2 \%$ 

 $\pm 10.0 \%$ 

 $\pm 20.1\,\%$ 

## 5. Uncertainty Analysis

Phantom and Setup

Liquid Conductivity (target)

Liquid Conductivity (meas.)

Liquid Permittivity (target)

Liquid Permittivity (meas.)

Combined Std. Uncertainty

**Expanded STD Uncertainty** 

Phantom Uncertainty

 $\pm 4.0 \%$ 

±5.0 %

 $\pm 2.5 \%$ 

 $\pm 5.0\%$ 

 $\pm 2.5 \%$ 

R

R

N

R

N

 $\sqrt{3}$ 

 $\sqrt{3}$ 

 $\sqrt{3}$ 

1

1

1

0.64

0.64

0.6

0.6

#### DASY4 Uncertainty Budget According to IEEE P1528 [1] Uncertainty Prob. Div. Std. Unc. Std. Unc. $(c_i)$ $(c_i)$ $(v_i)$ Error Description value Dist. 10g(10g)1g(1g) $v_{eff}$ Measurement System Probe Calibration $\pm 4.8 \%$ N $\pm 4.8\%$ $\pm 4.8 \%$ 1 1 1 $\infty$ $\pm 4.7 \%$ Axial Isotropy R $\sqrt{3}$ 0.7 0.7 $\pm 1.9 \%$ $\pm 1.9 \%$ $\infty$ $\pm 9.6\%$ $\sqrt{3}$ 0.7 $\pm 3.9 \%$ Hemispherical Isotropy R 0.7 $\pm 3.9 \%$ $\infty$ Boundary Effects ±1.0 % R $\sqrt{3}$ $\pm 0.6 \%$ $\pm 0.6 \%$ 1 1 $\infty$ $\pm 4.7 \%$ R $\sqrt{3}$ $\pm 2.7\%$ $\pm 2.7 \%$ Linearity 1 1 $\infty$ System Detection Limits $\pm 1.0 \%$ R $\sqrt{3}$ 1 1 $\pm 0.6\%$ $\pm 0.6 \%$ $\infty$ Readout Electronics ±1.0 % N 1 1 1 $\pm 1.0\%$ $\pm 1.0 \%$ $\infty$ Response Time $\pm 0.8 \%$ R $\sqrt{3}$ $\pm 0.5 \%$ $\pm 0.5\%$ 1 1 $\infty$ Integration Time $\pm 2.6 \%$ R $\sqrt{3}$ 1 1 $\pm 1.5\%$ $\pm 1.5\%$ $\infty$ RF Ambient Conditions $\pm 3.0 \%$ R $\sqrt{3}$ 1 $\pm 1.7\%$ $\pm 1.7\%$ 1 $\infty$ Probe Positioner $\pm 0.4\%$ R $\sqrt{3}$ 1 $\pm 0.2\%$ $\pm 0.2\%$ 1 $\infty$ Probe Positioning $\pm 2.9 \%$ R $\sqrt{3}$ 1 $\pm 1.7\%$ $\pm 1.7 \%$ 1 $\infty$ $\pm 1.0 \, \%$ Max. SAR Eval. R $\sqrt{3}$ 1 1 $\pm 0.6\%$ $\pm 0.6 \%$ $\infty$ Test Sample Related Device Positioning $\pm 2.9 \%$ N 1 1 1 $\pm 2.9 \%$ $\pm 2.9 \%$ 875 Device Holder $\pm 3.6 \%$ N $\pm 3.6 \%$ $\pm 3.6 \%$ 1 1 5 1 Power Drift $\pm 5.0 \%$ R $\sqrt{3}$ 1 $\pm 2.9\%$ $\pm 2.9 \%$

1

1

0.43

0.43

0.49

0.49

 $\pm 2.3\%$ 

 $\pm 1.8\%$ 

 $\pm 1.6 \%$ 

 $\pm 1.7\%$ 

 $\pm 1.5\%$ 

 $\pm 10.3 \%$ 

 $\pm 20.6\,\%$ 

Page: 21 of 23

## 6. Phantom Description

Schmid & Partner Engineering AG

e a

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 1 245 9700, Fax +41 1 245 9779 info@speag.com, http://www.speag.com

#### Certificate of Conformity / First Article Inspection

Item	SAM Twin Phantom V4.0	
Type No	QD 000 P40 C	
Series No	TP-1150 and higher	
Manufacturer	SPEAG Zeughausstrasse 43 CH-8004 Zürich Switzerland	

#### Tests

The series production process used allows the limitation to test of first articles. Complete tests were made on the pre-series Type No. QD 000 P40 AA, Serial No. TP-1001 and on the series first article Type No. QD 000 P40 BA, Serial No. TP-1006. Certain parameters have been retested using further series items (called samples) or are tested at each item.

Test	Requirement	Details	Units tested
Dimensions	Compliant with the geometry according to the CAD model.	IT'IS CAD File (*)	First article, Samples
Material thickness of shell	Compliant with the requirements according to the standards	2mm +/- 0.2mm in flat and specific areas of head section	First article, Samples, TP-1314 ff.
Material thickness at ERP	Compliant with the requirements according to the standards	6mm +/- 0.2mm at ERP	First article, All items
Material parameters	Dielectric parameters for required frequencies	300 MHz – 6 GHz: Relative permittivity < 5, Loss tangent < 0.05	Material samples
Material resistivity	The material has been tested to be compatible with the liquids defined in the standards if handled and cleaned according to the instructions.  Observe technical Note for material compatibility.	DEGMBE based simulating liquids	Pre-series, First article, Material samples
Sagging	Compliant with the requirements according to the standards. Sagging of the flat section when filled with tissue simulating liquid.	< 1% typical < 0.8% if filled with 155mm of HSL900 and without DUT below	Prototypes, Sample testing

- Standards [1] CENELEC EN 50361
- IEEE Std 1528-2003
- IEC 62209 Part I FCC OET Bulletin 65, Supplement C, Edition 01-01
- The IT'IS CAD file is derived from [2] and is also within the tolerance requirements of the shapes of the other documents.

Conformity
Based on the sample tests above, we certify that this item is in compliance with the uncertainty requirements of SAR measurements specified in standards [1] to [4].

07.07.2005

Schmid & Pariner Engineering AG
250ghausstesse 43, 8004 Zurich Switzerla
Phone 1411-245 8700 Fax 441/245 9779
Info@apeag.com, http://www.speag.com

Signature / Stamp

Report No.: ES/2007/50009 Page: 22 of 23

## 7. System Validation from Original equipment supplier

#### **DASY4 Validation Report for Body TSL**

Date/Time: 12.02.2007 14:24:23

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN:178

Communication System: CW; Frequency: 900 MHz; Duty Cycle: 1:1

Medium: MSL900;

Medium parameters used: f = 900 MHz;  $\sigma$  = 1.04 mho/m;  $\varepsilon_r$  = 52.4;  $\rho$  = 1000 kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

#### DASY4 Configuration:

Probe: ET3DV6 - SN1507 (HF); ConvF(5.8, 5.8, 5.8); Calibrated: 19.10.2006

Sensor-Surface: 4mm (Mechanical Surface Detection)


- Electronics: DAE4 Sn601; Calibrated: 30.01.2007
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA;;
- Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

#### Pin=250mW/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 55.9 V/m; Power Drift = 0.018 dB

Peak SAR (extrapolated) = 3.80 W/kg

SAR(1 g) = 2.69 mW/g; SAR(10 g) = 1.76 mW/g Maximum value of SAR (measured) = 2.92 mW/g



0 dB = 2.92 mW/g

Report No.: ES/2007/50009 Page: 23 of 23

#### **DASY4 Validation Report for Body TSL**

Date/Time: 20.03.2007 15:34:07

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:5d027

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: MSL U10 BB;

Medium parameters used: f = 1900 MHz;  $\sigma = 1.58 \text{ mho/m}$ ;  $\varepsilon_r = 51.9$ ;  $\rho = 1000 \text{ kg/m}^3$ 

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

#### DASY4 Configuration:

• Probe: ET3DV6 - SN1507 (HF); ConvF(4.43, 4.43, 4.43); Calibrated: 19.10.2006

- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 30.01.2007
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA;;
- Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

#### Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 90.5 V/m; Power Drift = -0.010 dB

Peak SAR (extrapolated) = 16.2 W/kg

SAR(1 g) = 9.67 mW/g; SAR(10 g) = 5.16 mW/g

Maximum value of SAR (measured) = 10.9 mW/g



0 dB = 10.9 mW/g