



Date: 01/09/2025

Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, MD 21046

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID: URA-MS40251  
Original Grant date: 09/10/2024  
Applicant: Draeger Medical Systems, Inc.

Dear Examiner,

This is to request a Class II permissive change for FCC ID: URA-MS40251

The major change filed under this application is:

- The WIFI module is integrated by us (Grantee) into our own mobile host device with product identification (HMN): Infinity® M540 patient monitor
- New integrated PCB antenna
  - Antenova CU5006 2.4 GHz: 3.1 dBi; 5 GHz: 3.78 dBi
- Deactivating Bluetooth via software
- 2.4 + 5 GHz HT 40 MHz Mode deactivated
- Reduced power settings for UNII HT20 – Channel 1 from 10dBm to 8dBm

If you have any questions regarding this application, please feel free to contact me.

Sincerely,

Signature

Georgios Kokovidis  
Wireless Manager  
Draeger Medical Systems Inc.