



American Telecommunications Certification Body Inc.
6731 Whittier Ave, McLean, VA 22101

October 24, 2007

RE: VivoMetrics

FCC ID: UWVLS300RX

After a review of the submitted information, I have a few comments on the above referenced Application.

- 1) For IC, the model Certified must match the device labeling, or somewhere on the device. Currently the IC form is asking for LS300RS-US to be Certified while the label only shows LS300. Models must match what is being Certified. Note that
- 2) FYI....Users Manual still references 15.247. This should be corrected.
- 3) FYI....Certain parts of the system would require 15.105(b) statements to be placed in the manual. The manufacturer should correct this.



Timothy R. Johnson
Examining Engineer

[mailto: tjohnson@AmericanTCB.com](mailto:tjohnson@AmericanTCB.com)

The items indicated above must be submitted before processing can continue on the above referenced application. Failure to provide the requested information may result in application termination. Correspondence should be considered part of the permanent submission and may be viewed from the Internet after a Grant of Equipment Authorization is issued.

Please do not respond to this correspondence using the email reply button. In order for your response to be processed expeditiously, you must submit your documents through the AmericanTCB.com website. Also, please note that partial responses increase processing time and should not be submitted.

Any questions about the content of this correspondence should be directed to the sender.