

Telecommunication Certification Body
UL VS Ltd
Unit 3, Horizon
Wade Road
Kingsland Business Park
Basingstoke
Hampshire
RG24 8AH
United Kingdom

Subject: FCC Single-Modular Approval Letter

FCC ID: V5A-NOXBLEMOD

To whom it may concern

We, Nox Medical, hereby declare that the product, Nox Bluetooth LE Module FCC ID: V5A-NOXBLEMOD, has met the single-modular approval requirements of FCC rule part §15.212(a)(1) and this is shown in the table below.

Requirement	Compliance: Yes or No along with a justification
The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly	Yes. The Nox Bluetooth LE Module has a dedicated shielding implemented according to FCC module approval requirements.
The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal	Yes. The information is located in the Operational Description exhibit
The module must contain power supply regulation on the module	Yes. The information is located in the Operational Description exhibit
The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b)	Yes. See the Photos exhibits.
The module must demonstrate compliance in a stand-alone configuration	Yes. The module is a complete BLE communication system and is capable of FCC compliance testing as a stand-alone device.

The module must be labelled with its permanently affixed FCC ID label, or use an electronic display (See KDB Publication 784748 about labelling requirements)	Yes. The module will be marked with a permanent FCC label attached on the shielding box cover.
The module must comply with all specific rules applicable to the transmitter including all the conditions provided in the integration instructions by the grantee	Yes. See the User Installation Guide
The module must comply with RF exposure requirements	Yes. See the RF Exposure exhibit.

Yours faithfully,



Kolbrún Eyðís Ottósdóttir

Chief Compliance Officer, Nox Medical