

Annex H. Calibration of Test Equipment List

Calibration of Test Equipment List are shown as below.

Equipments used for HAC RF-Emission Testing					
Equipment	Manufacturer	Model	SN	Cal. Date	Cal. Interval
835MHz Calibration Dipole	SPEAG	CD835V3	1041	Jan. 22, 2020	3 Years
1880MHz Calibration Dipole	SPEAG	CD1880V3	1032	Jan. 22, 2020	3 Years
3500MHz Calibration Dipole	SPEAG	CD3500V3	1004	Sep. 15, 2020	3 Years
Isotropic E-Field Probe	SPEAG	EF3DV3	4049	Jan. 25, 2021	1 Year
Data Acquisition Electronics	SPEAG	DAE4	1589	Sep. 15, 2020	1 Year
Universal Radio Communication Tester	Anritsu	MT8821C	6261786083	Jun. 27, 2019	1 Year