

Correspondence Reference # 37247

Federal Communication Commission
Equipment Authorization Division,
Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

March 30, 2009

To Whom It May Concern

ID Label/ Label Location – Unitron Passport BTE

Pursuant to paragraph §2.925 of 47 C.F.R. the hearing instrument Unitron Passport BTE, which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of the 47 C.F.R. Due to the specific use of the hearing instrument (as shown in figure 1) its external surface is in contact with the user's skin behind the ear.



Figure 1: Position of the Hearing Instrument on the User

Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations. The only alternative place for the FCC ID and serial number of the device would be in the battery compartment as shown in figure 02.



Figure 2: Battery Compartment Passport BTE


In this case the size of the FCC ID will be smaller than 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925 (d)(2) and §2.925(g).

Due to the small size of the devices and pursuant to paragraph § 15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the user manual as shown figure 3.

Notices

Notice 1

This instrument is certified under:



This device complies with Part 15 of the FCC Rules and with RSS-240 of Industry Canada. Operation is subject to the following two conditions:

- 1) this device may not cause harmful interference, and
- 2) this device must accept any interference received, including interference that may cause undesired operation. Changes or modifications made to this equipment not expressly approved by Phonak may void the FCC authorization to operate this equipment.

Notice 2

This Class B digital apparatus complies with Canadian ICES-003.

Notice 3

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna
- Increase the separation between the equipment and receiver
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected
- Consult the dealer or an experienced radio/TV technician for help

Figure 3: FCC ID and Compliance Statement – User Manual

Ric Castle
Manager, Quality Assurance.
Unitron Hearing
20 Beasley Drive, P.O. Box 9017
Kitchener, Ontario,
Canada N2G 4X1

Phone: 519 895 0100
Fax: 519 895 0108
Email: ric.castle@unitronhearing.com