

01 July 2008

AmericanTCB
6731 Whittier Avenue
McLean VA 22101
Attn: Director of Certification
RE: Declaration of Compliance to RF Exposure Limits for Humans
Healthsense MSR-154-200

RF Exposure Statement

1. Standard Applicable


According to 1.1307 (b)(1), systems operating under the provisions of this section shall be operated in a manner that ensure that the public is not exposed to radio frequency energy level in excess of the Commission's guideline.

This is a portable device operating in 2450 MHz range.

2. Measurement Result:

This is a mobile device and the max peak output power is 0.0076 W. Lower than low threshold 60/f GHz mW (24 mW), d<2.5 cm, general population category.

The SAR measurement is not necessary.



Joel T. Schneider
Senior Engineer
Minnesota Wild River Lab
TÜV SUD America
Product Service
Tel: 651 604 4582
Fax: 651 638 0298
jschneider@tuvam.com