



May 19, 2009

Applications Examiner
American TCB, Inc.
6731 Whittier Avenue, Suite C110
McLean, VA 22101

RE: Permissive Change Application for Lecat's Ventriloscope, LLC.

**FCC ID: WNT-VENT-R
IC: 7880A-VENTR**

Dear Applications Examiner:

The Lecat's Ventriloscope ARX was modified to allow the transmitter microcontroller to communicate directly with the receiver microcontroller. Upon powering up a transceiver pair (Transmitter and Receiver) they connect with each other with a preprogrammed address. The transmitter microcontroller then calculates a random address and transmits it to the receiver and they connect again at this new address and remain at that address until powered down. Circuitry around R26 – R28 was changed to allow this communication to occur. Also the battery for the Receiver, a LPP 402934 was changed.

Based on the equipment modifications, re-testing was performed for Fundamental and Harmonics and Radiated Spurious Emissions. The re-measured data was within FCC Limits, but higher than the original data, so this Permissive Change 2 Application is being submitted to report the change and the new data.

Sincerely,

A handwritten signature in black ink that reads 'Sandi McEnergy'.

Sandi McEnergy
Manager – US Tech
Agent for Lecat's Ventriloscope, LLC.