


Dipole Intermediate Check


2600 MHz

Dipole D2600V2 SN: 1162				
Head Liquid				
Test By: Allen Sun		Checked By: Bard Liu		
Date of Measurement	Return Loss(dB)	$\Delta\%$	Impedance (Ω)	$\Delta\Omega$
2022.08.22	-23.2	N/A	47.7	N/A
2023.08.21	-23.8	2.586	45.5	2.2
2024.08.20	-20.3	12.5	46.5	1.2

D2600V2 - SN: 1162 (Date of Measurement: 2023/08/21)



D2600V2 - SN: 1162 (Date of Measurement: 2024/08/20)




Dipole Intermediate Check


5250 MHz

Dipole D5GH2V2 SN: 1296				
Head Liquid				
Test By: Allen Sun		Checked By: Brad Liu		
Date of Measurement	Return Loss(dB)	$\Delta\%$	Impedance (Ω)	$\Delta\Omega$
2022.08.17	-30.1	N/A	48.5	N/A
2023.08.16	-32.3	0.664	47.1	1.4
2024.08.15	-27.2	9.635	47.5	1

D5GH2V2 - SN: 1296 (Date of Measurement: 2023/08/16)



D5GH2V2 - SN: 1296 (Date of Measurement: 2024/08/15)




Dipole Intermediate Check

5750 MHz

Dipole D5GH2V2 SN: 1296				
Head Liquid				
Test By: <i>Kellen Sun</i>		Checked By: <i>Brent Lin</i>		
Date of Measurement	Return Loss(dB)	$\Delta\%$	Impedance (Ω)	$\Delta\Omega$
2022.08.17	-30.1	N/A	49.2	N/A
2023.08.16	-30.7	1.993	52.8	3.6
2024.08.15	-29.3	2.658	47.7	1.5

D5GH2V2 - SN: 1296 (Date of Measurement: 2023/08/16)



D5GH2V2 - SN: 1296 (Date of Measurement: 2024/08/15)

