



Ref: US.000269

BABT
Balfour House, Churchfield Road
Walton on Thames
Surrey, KT12 4TD
United Kingdom
Telephone: +44 (0)1932 251200
Fax: +44 (0)1932 251201

5th May 2009

To FCC

Direct Dial: +44 (0)1932 251230
E-mail: Tom.Twynam@babt.com
Website: www.babt.com

Evaluation of Beijing Choice Electronic Company Pulse Oximeter

FCC id: WWIMD300C318

I have assessed the above referenced submission for Beijing Choice Electronic Company and found it to be compliant. The product is a pulse oximeter model MD300C318 with Bluetooth..

The following points were noted and resolved during the evaluation.

1. User Manual

As originally submitted the regulatory warnings were incorrect and missing some statements regarding interference. A revised manual was provided by the applicant.

2. Confidentiality

Short term confidentiality was originally requested for some items that should have been long term. The letter was amended to reflect this and to remove references to the bill of materials and tune up procedure.

Yours sincerely

A handwritten signature in blue ink that reads "Tom Twynam". The signature is fluid and cursive, with "Tom" and "Twynam" clearly legible.

Tom Twynam
Technical Certifier
For BABT TCB