



Manufactured for:
ClearCount Medical Solutions, Inc.
101 Bellevue Road, #300
Pittsburgh, PA 15229

ClearCount SmartSponge® System
Model A02

120-240 VAC, 50/60 Hz, 60 W
0.65 Amps at 120 VAC
Class 1, Type B Equipment

window



FCC ID: WWQCCMS001 IC: 7864A-CCMS001

This device complies with Part 15 of FCC rules. Operation is subject to the following two conditions: 1) this device may not cause harmful interference and, 2) this device must accept any interference received, including interference that may cause undesired operation.

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

The System contains no user replaceable fuses. Any attempt to service this device by the user will result in voiding of any and all warranties.

This Class 1 digital apparatus complies with Canadian ICES-003.

Cet appareil numérique de la classe 1 est conforme à la norme NMB-003 du Canada



(100278) Label Location

