

Information about the Applicant	
Grantee (Company Name)	ClearCount Medical Solutions, Inc.
FCC Contact Person	Jeff Wolfgang
Address	101 Bellevue Road, Suite 300
City, State, Zip	Pittsburg, PA 15229
Job Number	KEYS0009
Model	SmartSponge Flex A04
FCC ID	WWQCCMS004
Agent	Franck D'Annunzio
Approval Type	Original
Equipment Class	DXX Part 15, Low Power Communication Device Transmitter
Rule Part	15C

Overview

The SmartSponge Flex Model A04 is a component of the ClearCount's SmartSponge system, which includes surgical sponges, laparotomy pads and surgical towels. Each of these components contains unique RFID tags. The Model A04 employs a radio, operating at 13.56 MHz to manage the various components of the system in the surgical suite.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Short term confidentiality requested.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature


Dave Tolman, TCB Committee
2/16/2012