



Nevro Corp. • 1800 Bridge Parkway • Redwood City, CA 94065 USA • 650-251-0005

XKYPR1D3000 Labeling



Materials Specifications

Material	Polycarbonate Lexan 8010V (Thickness = 0.010") Satin finish
Adhesive	3M 9471 or equivalent (Thickness = 0.002")

Visual Requirements

Label must be in good condition, not presenting any bends.

Label is permanently attached to the back of the device with adhesive and is resistant against wear.