

## DECLARATION OF SIMILARITY

March 24, 2016

To:  
Bay Area Compliance Laboratories Corp.  
1274 Anvilwood Ave.  
Sunnyvale, CA 94089  
Phone: 408-732-9162, Fax: 408-732-9164  
<http://www.baclcorp.com>

Dear Sir or Madam:

We *Siemens Medical Solutions USA Inc.* hereby declare that product: *Handheld Probe*, model(s): *L8-3, C5-2* is/are electrically identical with the same electromagnetic emissions and electromagnetic compatibility characteristics as model: *L13-5* tested by BACL, the results of which are featured in BACL project: *R1602221*

A description of the differences between the tested model and those that are declared similar are as follows:

*The ACUSON Freestyle handheld probes, Models L8-3, L13-5 and C5-2 are all equivalent electrically and use the same internal circuitry. The schematics and materials for all three models are identical. The difference is that each model uses a different ultrasonic transducer array (the nosepiece), which is driven by the internal circuitry. The different nosepieces are optimized for different procedures within diagnostic ultrasound.*

Please contact me should there be need for any additional clarification or information.

Best Regards,



Larry Engle  
Director, Quality  
Phone: 610.834.1220 x102  
Email: [larry.engle@siemens.com](mailto:larry.engle@siemens.com)  
Siemens Medical Solutions USA Inc.  
5168 Campus Drive, Plymouth Meeting, PA 19462, USA