

Appendix C for KSCR221100220901

Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29
	<input type="checkbox"/>	4	D835V2	4d114	2022/03/31
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07
	<input type="checkbox"/>	6	D1800V2	2d170	2022/03/31
	<input type="checkbox"/>	7	D1900V2	5d136	2022/06/07
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01
	<input type="checkbox"/>	11	D2600V2	1158	2022/03/31
	<input type="checkbox"/>	12	D5GHzV2	1095	2022/06/01
DAE	<input checked="" type="checkbox"/>	13	DAE4	1245	2022/05/30
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7767	2022/10/28

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1 Dipole

1.1 CLA150 - SN 4025

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 45, 8001 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No: CLA150-4025_Apr21</p> <p>CALIBRATION CERTIFICATE</p> <table border="1"> <tr> <td>Object</td> <td>CLA150 - SN: 4025</td> </tr> <tr> <td>Calibration procedure</td> <td>QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz</td> </tr> <tr> <td>Calibration date</td> <td>April 26, 2021</td> </tr> <tr> <td colspan="2">This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</td> </tr> <tr> <td colspan="2">All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.</td> </tr> <tr> <td colspan="2">Calibration Equipment used (NOTE: critical for calibration)</td> </tr> <tr> <td colspan="2"> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 104/78</td> <td>09-Apr-21 (No. 217-03291/03292)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/44</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/45</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C23662 (20)</td> <td>09-Apr-21 (No. 217-03293)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch compensation</td> <td>SN: 310982 / 00377</td> <td>09-Apr-21 (No. 217-03294)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3084</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-0877, Dec-20)</td> <td>Dec-21</td> </tr> <tr> <td>D414</td> <td>SN: 654</td> <td>26-Jun-20 (No. D414-654, Jun-20)</td> <td>Jun-21</td> </tr> </tbody> </table> </td> </tr> <tr> <td colspan="2"> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4419B</td> <td>SN: GB41209874</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419B</td> <td>SN: MY41490607</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419A</td> <td>SN: 103/10</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>EUT generator HP 8548C</td> <td>SN: US28420-001700</td> <td>31-Aug-08 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8380A</td> <td>EN: US41002477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check: Oct-21</td> </tr> </tbody> </table> </td> </tr> <tr> <td colspan="2"> <p>Calibrated by: Name: Jeffrey Katzman Function: Laboratory Technician Signature: </p> <p>Approved by: Name: Katica Polovic Function: Technical Manager Signature: </p> <p>Issued: April 26, 2021</p> </td> </tr> <tr> <td colspan="2"> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 6</p> </td> </tr> </table> <table border="1"> <tr> <td colspan="4">Measurement Conditions DASY system configuration, as far as not given on page 1.</td> </tr> <tr> <td>DASY Version</td> <td colspan="3">DASY5 V32.10.4</td> </tr> <tr> <td>Extrapolation</td> <td colspan="3">Advanced Extrapolation</td> </tr> <tr> <td>Phantom</td> <td colspan="3">EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td colspan="3">Touch Position</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Grated Ratio = 1.4 (Z direction)</td> <td></td> </tr> <tr> <td>Frequency</td> <td colspan="3">150 MHz ± 1 MHz</td> </tr> </table> <table border="1"> <tr> <td colspan="4">Head TSL parameters The following parameters and calculations were applied.</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>22.0 °C</td> <td>62.3</td> <td>0.78 mho/m</td> <td></td> </tr> <tr> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.78 mho/m ± 6 %</td> <td></td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>< 0.5 °C</td> <td>----</td> <td>----</td> </tr> </table> <table border="1"> <tr> <td colspan="4">SAR result with Head TSL</td> </tr> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td colspan="2"></td> </tr> <tr> <td>SAR measured</td> <td>1 W input power</td> <td colspan="2">3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td colspan="2">3.88 W/kg ± 18.4 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td colspan="2"></td> </tr> <tr> <td>SAR measured</td> <td>1 W input power</td> <td colspan="2">2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td colspan="2">2.59 W/kg ± 18.0 % (k=2)</td> </tr> </table>		Object	CLA150 - SN: 4025	Calibration procedure	QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz	Calibration date	April 26, 2021	This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.		All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.		Calibration Equipment used (NOTE: critical for calibration)		<table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 104/78</td> <td>09-Apr-21 (No. 217-03291/03292)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/44</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/45</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C23662 (20)</td> <td>09-Apr-21 (No. 217-03293)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch compensation</td> <td>SN: 310982 / 00377</td> <td>09-Apr-21 (No. 217-03294)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3084</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-0877, Dec-20)</td> <td>Dec-21</td> </tr> <tr> <td>D414</td> <td>SN: 654</td> <td>26-Jun-20 (No. D414-654, Jun-20)</td> <td>Jun-21</td> </tr> </tbody> </table>		Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 104/78	09-Apr-21 (No. 217-03291/03292)	Apr-22	Power sensor NRP 291	SN: 103/44	09-Apr-21 (No. 217-03291)	Apr-22	Power sensor NRP 291	SN: 103/45	09-Apr-21 (No. 217-03291)	Apr-22	Reference 20 dB Attenuator	SN: C23662 (20)	09-Apr-21 (No. 217-03293)	Apr-22	Type-N mismatch compensation	SN: 310982 / 00377	09-Apr-21 (No. 217-03294)	Apr-22	Reference Probe EX3084	SN: 3877	30-Dec-20 (No. EX3-0877, Dec-20)	Dec-21	D414	SN: 654	26-Jun-20 (No. D414-654, Jun-20)	Jun-21	<table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4419B</td> <td>SN: GB41209874</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419B</td> <td>SN: MY41490607</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419A</td> <td>SN: 103/10</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>EUT generator HP 8548C</td> <td>SN: US28420-001700</td> <td>31-Aug-08 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8380A</td> <td>EN: US41002477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check: Oct-21</td> </tr> </tbody> </table>		Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E4419B	SN: GB41209874	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	Power sensor E4419B	SN: MY41490607	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	Power sensor E4419A	SN: 103/10	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	EUT generator HP 8548C	SN: US28420-001700	31-Aug-08 (in house check Jun-20)	In house check: Jun-22	Network Analyzer Agilent E8380A	EN: US41002477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21	<p>Calibrated by: Name: Jeffrey Katzman Function: Laboratory Technician Signature: </p> <p>Approved by: Name: Katica Polovic Function: Technical Manager Signature: </p> <p>Issued: April 26, 2021</p>		<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 6</p>		Measurement Conditions DASY system configuration, as far as not given on page 1.				DASY Version	DASY5 V32.10.4			Extrapolation	Advanced Extrapolation			Phantom	EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm			EUT Positioning	Touch Position			Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Grated Ratio = 1.4 (Z direction)		Frequency	150 MHz ± 1 MHz			Head TSL parameters The following parameters and calculations were applied.				Nominal Head TSL parameters	Temperature	Permittivity	Conductivity	22.0 °C	62.3	0.78 mho/m		(22.0 ± 0.2) °C	51.1 ± 6 %	0.78 mho/m ± 6 %		Head TSL temperature change during test	< 0.5 °C	----	----	SAR result with Head TSL				SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition			SAR measured	1 W input power	3.90 W/kg		SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)		SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition			SAR measured	1 W input power	2.60 W/kg		SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)		<p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 45, 8001 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Glossary:</p> <p>TSL tissue simulating liquid Conv/F sensitivity in TSL / NORM x,y,z N/A not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEEE Std 2009-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013</p> <p>b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016</p> <p>c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 300 MHz to 6 GHz)", March 2010</p> <p>d) KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation:</p> <p>e) DASY4/5 System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom. Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low required power. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: CLA150-4025_Apr21 Page 2 of 6</p>
Object	CLA150 - SN: 4025																																																																																																																																																									
Calibration procedure	QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz																																																																																																																																																									
Calibration date	April 26, 2021																																																																																																																																																									
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																																																																																																																																																										
All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.																																																																																																																																																										
Calibration Equipment used (NOTE: critical for calibration)																																																																																																																																																										
<table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 104/78</td> <td>09-Apr-21 (No. 217-03291/03292)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/44</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP 291</td> <td>SN: 103/45</td> <td>09-Apr-21 (No. 217-03291)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C23662 (20)</td> <td>09-Apr-21 (No. 217-03293)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch compensation</td> <td>SN: 310982 / 00377</td> <td>09-Apr-21 (No. 217-03294)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3084</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-0877, Dec-20)</td> <td>Dec-21</td> </tr> <tr> <td>D414</td> <td>SN: 654</td> <td>26-Jun-20 (No. D414-654, Jun-20)</td> <td>Jun-21</td> </tr> </tbody> </table>		Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 104/78	09-Apr-21 (No. 217-03291/03292)	Apr-22	Power sensor NRP 291	SN: 103/44	09-Apr-21 (No. 217-03291)	Apr-22	Power sensor NRP 291	SN: 103/45	09-Apr-21 (No. 217-03291)	Apr-22	Reference 20 dB Attenuator	SN: C23662 (20)	09-Apr-21 (No. 217-03293)	Apr-22	Type-N mismatch compensation	SN: 310982 / 00377	09-Apr-21 (No. 217-03294)	Apr-22	Reference Probe EX3084	SN: 3877	30-Dec-20 (No. EX3-0877, Dec-20)	Dec-21	D414	SN: 654	26-Jun-20 (No. D414-654, Jun-20)	Jun-21																																																																																																																									
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																																																																																																																																																							
Power meter NRP	SN: 104/78	09-Apr-21 (No. 217-03291/03292)	Apr-22																																																																																																																																																							
Power sensor NRP 291	SN: 103/44	09-Apr-21 (No. 217-03291)	Apr-22																																																																																																																																																							
Power sensor NRP 291	SN: 103/45	09-Apr-21 (No. 217-03291)	Apr-22																																																																																																																																																							
Reference 20 dB Attenuator	SN: C23662 (20)	09-Apr-21 (No. 217-03293)	Apr-22																																																																																																																																																							
Type-N mismatch compensation	SN: 310982 / 00377	09-Apr-21 (No. 217-03294)	Apr-22																																																																																																																																																							
Reference Probe EX3084	SN: 3877	30-Dec-20 (No. EX3-0877, Dec-20)	Dec-21																																																																																																																																																							
D414	SN: 654	26-Jun-20 (No. D414-654, Jun-20)	Jun-21																																																																																																																																																							
<table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4419B</td> <td>SN: GB41209874</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419B</td> <td>SN: MY41490607</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4419A</td> <td>SN: 103/10</td> <td>09-Apr-10 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>EUT generator HP 8548C</td> <td>SN: US28420-001700</td> <td>31-Aug-08 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8380A</td> <td>EN: US41002477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check: Oct-21</td> </tr> </tbody> </table>		Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E4419B	SN: GB41209874	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	Power sensor E4419B	SN: MY41490607	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	Power sensor E4419A	SN: 103/10	09-Apr-10 (in house check Jun-20)	In house check: Jun-22	EUT generator HP 8548C	SN: US28420-001700	31-Aug-08 (in house check Jun-20)	In house check: Jun-22	Network Analyzer Agilent E8380A	EN: US41002477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21																																																																																																																																	
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																																																																																																																																																							
Power meter E4419B	SN: GB41209874	09-Apr-10 (in house check Jun-20)	In house check: Jun-22																																																																																																																																																							
Power sensor E4419B	SN: MY41490607	09-Apr-10 (in house check Jun-20)	In house check: Jun-22																																																																																																																																																							
Power sensor E4419A	SN: 103/10	09-Apr-10 (in house check Jun-20)	In house check: Jun-22																																																																																																																																																							
EUT generator HP 8548C	SN: US28420-001700	31-Aug-08 (in house check Jun-20)	In house check: Jun-22																																																																																																																																																							
Network Analyzer Agilent E8380A	EN: US41002477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21																																																																																																																																																							
<p>Calibrated by: Name: Jeffrey Katzman Function: Laboratory Technician Signature: </p> <p>Approved by: Name: Katica Polovic Function: Technical Manager Signature: </p> <p>Issued: April 26, 2021</p>																																																																																																																																																										
<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 6</p>																																																																																																																																																										
Measurement Conditions DASY system configuration, as far as not given on page 1.																																																																																																																																																										
DASY Version	DASY5 V32.10.4																																																																																																																																																									
Extrapolation	Advanced Extrapolation																																																																																																																																																									
Phantom	EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm																																																																																																																																																									
EUT Positioning	Touch Position																																																																																																																																																									
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Grated Ratio = 1.4 (Z direction)																																																																																																																																																								
Frequency	150 MHz ± 1 MHz																																																																																																																																																									
Head TSL parameters The following parameters and calculations were applied.																																																																																																																																																										
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																																																																																																																							
22.0 °C	62.3	0.78 mho/m																																																																																																																																																								
(22.0 ± 0.2) °C	51.1 ± 6 %	0.78 mho/m ± 6 %																																																																																																																																																								
Head TSL temperature change during test	< 0.5 °C	----	----																																																																																																																																																							
SAR result with Head TSL																																																																																																																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																																																																																																																									
SAR measured	1 W input power	3.90 W/kg																																																																																																																																																								
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)																																																																																																																																																								
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																																																																																																																									
SAR measured	1 W input power	2.60 W/kg																																																																																																																																																								
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)																																																																																																																																																								
<p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr> <td>DASY Version</td> <td colspan="3">DASY5 V32.10.4</td> </tr> <tr> <td>Extrapolation</td> <td colspan="3">Advanced Extrapolation</td> </tr> <tr> <td>Phantom</td> <td colspan="3">EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td colspan="3">Touch Position</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Grated Ratio = 1.4 (Z direction)</td> <td></td> </tr> <tr> <td>Frequency</td> <td colspan="3">150 MHz ± 1 MHz</td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr> <td>Nominal Head TSL parameters</td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>22.0 °C</td> <td>62.3</td> <td>0.78 mho/m</td> <td></td> </tr> <tr> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.78 mho/m ± 6 %</td> <td></td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>< 0.5 °C</td> <td>----</td> <td>----</td> </tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td colspan="2"></td> </tr> <tr> <td>SAR measured</td> <td>1 W input power</td> <td colspan="2">3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td colspan="2">3.88 W/kg ± 18.4 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td colspan="2"></td> </tr> <tr> <td>SAR measured</td> <td>1 W input power</td> <td colspan="2">2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td colspan="2">2.59 W/kg ± 18.0 % (k=2)</td> </tr> </table>		DASY Version	DASY5 V32.10.4			Extrapolation	Advanced Extrapolation			Phantom	EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm			EUT Positioning	Touch Position			Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Grated Ratio = 1.4 (Z direction)		Frequency	150 MHz ± 1 MHz			Nominal Head TSL parameters	Temperature	Permittivity	Conductivity	22.0 °C	62.3	0.78 mho/m		(22.0 ± 0.2) °C	51.1 ± 6 %	0.78 mho/m ± 6 %		Head TSL temperature change during test	< 0.5 °C	----	----	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition			SAR measured	1 W input power	3.90 W/kg		SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)		SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition			SAR measured	1 W input power	2.60 W/kg		SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)		<p>Appendix (Additional assessments outside the scope of SCS 0108)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point:</td> <td>47.8 Ω ± 1.5 jΩ</td> </tr> <tr> <td>Return Loss:</td> <td>-31.4 dB</td> </tr> </table> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> <p>Certificate No: CLA150-4025_Apr21 Page 4 of 6</p>	Impedance, transformed to feed point:	47.8 Ω ± 1.5 jΩ	Return Loss:	-31.4 dB	Manufactured by	SPEAG																																																																																		
DASY Version	DASY5 V32.10.4																																																																																																																																																									
Extrapolation	Advanced Extrapolation																																																																																																																																																									
Phantom	EL14 Flat Phantom Shell thickness: 2 ± 0.2 mm																																																																																																																																																									
EUT Positioning	Touch Position																																																																																																																																																									
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Grated Ratio = 1.4 (Z direction)																																																																																																																																																								
Frequency	150 MHz ± 1 MHz																																																																																																																																																									
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																																																																																																																							
22.0 °C	62.3	0.78 mho/m																																																																																																																																																								
(22.0 ± 0.2) °C	51.1 ± 6 %	0.78 mho/m ± 6 %																																																																																																																																																								
Head TSL temperature change during test	< 0.5 °C	----	----																																																																																																																																																							
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																																																																																																																									
SAR measured	1 W input power	3.90 W/kg																																																																																																																																																								
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)																																																																																																																																																								
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																																																																																																																									
SAR measured	1 W input power	2.60 W/kg																																																																																																																																																								
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)																																																																																																																																																								
Impedance, transformed to feed point:	47.8 Ω ± 1.5 jΩ																																																																																																																																																									
Return Loss:	-31.4 dB																																																																																																																																																									
Manufactured by	SPEAG																																																																																																																																																									

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn

t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



<p>DASY5 Validation Report for Head TSL</p> <p>Date: 26.04.2021</p> <p>Test Laboratory: SPEAG, Zurich, Switzerland</p> <p>DUT: CLA150; Type: CLA150; Serial: CLA150 - SN: 4025</p> <p>Communication System: UUD 0 - CW; Frequency: 150 MHz</p> <p>Medium parameters used: $\epsilon_r = 1.0$; $\sigma = 0.76 \text{ S/m}$; $\epsilon_0 = 51.1$; $\rho = 1000 \text{ kg/m}^3$</p> <p>Phantom section: Flat Section</p> <p>Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)</p> <p>DASY52 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN3877; ConvF(12.51, 12.51, 12.51) @ 150 MHz; Calibrated: 30.12.2020 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAB4 Suf654; Calibrated: 26.06.2020 Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP:1003 DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483) <p>CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x8) /Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</p> <p>Reference Value = 85.93 V/m; Power Drift = -0.02 dB</p> <p>Peak SAR (extrapolated) = 7.36 W/kg</p> <p>SAR(1 g) = 3.90 W/kg; SAR(10 g) = 2.00 W/kg</p> <p>Small distances tend to result in amounts 3 dB below; Larger than measurement grid (>30mm)</p> <p>Ratio of SAR at M2 to SAR at M1 = 80.4%</p> <p>Maximum value of SAR (measured) = 5.48 W/kg</p> <p>0 dB = 5.48 W/kg = 7.39 dBW/kg</p>  <p>Certificates No: CLA150-4025_Apr21 Page 5 of 6</p>	 <p>Certificates No: CLA150-4025_Apr21 Page 6 of 6</p>
--	--

1.2 D450V3 - SN 1103

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zughestrasse 43, 8004 Zurich, Switzerland</p> <p> </p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No: D450V3-1103_Apr21</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D450V3 - SN:1103</p> <p>Calibration procedure(s): QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz</p> <p>Calibration date: April 21, 2021</p> <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature ($20 \pm 3^\circ\text{C}$) and humidity $\approx 75\%$.</p> <p>Calibration Equipment used (M/TE critical for calibration):</p> <table border="1"> <tr> <td>Primary Standards</td> <td>ID #</td> <td>Cal Date (Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>Power meter NIP-291</td> <td>SN: 100778</td> <td>09-Apr-21 (No. 217-0021-0320)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NIP-291</td> <td>SN: 100284</td> <td>09-Apr-21 (No. 217-0021)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NIP-291</td> <td>SN: 103219</td> <td>09-Apr-21 (No. 217-0021)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: CC2652 (200)</td> <td>09-Apr-21 (No. 217-0043)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N minmax combination</td> <td>SN: 310982 / 06327</td> <td>09-Apr-21 (No. 217-0044)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-3077- Dec20)</td> <td>Dec-21</td> </tr> <tr> <td>DAE4</td> <td>SN: 954</td> <td>28-Jun-20 (No. DAE4-954_Jun20)</td> <td>Jun-21</td> </tr> </table> <table border="1"> <tr> <td>Secondary Standards</td> <td>ID #</td> <td>Check Date (in house)</td> <td>Scheduled Check</td> </tr> <tr> <td>Power meter E41198</td> <td>SN: G411290874</td> <td>09-Apr-16 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power meter E41198</td> <td>SN: 9411290874</td> <td>09-Apr-16 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Power sensor E4412A</td> <td>SN: 200116019</td> <td>09-Apr-19 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>RF generator HP 8648C</td> <td>EN: URS440-011700</td> <td>09-Aug-19 (in house check Jun-20)</td> <td>In house check: Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E3036A</td> <td>SN: US410960477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check: Oct-21</td> </tr> </table> <p>Calibrated by: Name: Claudio Leutler Function: Laboratory Technician Signature: </p> <p>Approved by: Name: Katja Povacic Function: Technical Manager Signature: </p> <p>Issued: April 23, 2021</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: D450V3-1103_Apr21 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NIP-291	SN: 100778	09-Apr-21 (No. 217-0021-0320)	Apr-22	Power sensor NIP-291	SN: 100284	09-Apr-21 (No. 217-0021)	Apr-22	Power sensor NIP-291	SN: 103219	09-Apr-21 (No. 217-0021)	Apr-22	Reference 20 dB Attenuator	SN: CC2652 (200)	09-Apr-21 (No. 217-0043)	Apr-22	Type-N minmax combination	SN: 310982 / 06327	09-Apr-21 (No. 217-0044)	Apr-22	Reference Probe EX3DV4	SN: 3877	30-Dec-20 (No. EX3-3077- Dec20)	Dec-21	DAE4	SN: 954	28-Jun-20 (No. DAE4-954_Jun20)	Jun-21	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E41198	SN: G411290874	09-Apr-16 (in house check Jun-20)	In house check: Jun-22	Power meter E41198	SN: 9411290874	09-Apr-16 (in house check Jun-20)	In house check: Jun-22	Power sensor E4412A	SN: 200116019	09-Apr-19 (in house check Jun-20)	In house check: Jun-22	RF generator HP 8648C	EN: URS440-011700	09-Aug-19 (in house check Jun-20)	In house check: Jun-22	Network Analyzer Agilent E3036A	SN: US410960477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zughestrasse 43, 8004 Zurich, Switzerland</p> <p> </p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Glossary:</p> <ul style="list-style-type: none"> TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORM x,y,z N/A: not applicable or not measured <p>Calibration is Performed According to the Following Standards:</p> <ol style="list-style-type: none"> IEEE Std 1628-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013 IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010 KDB 605864, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p>Additional Documentation:</p> <ol style="list-style-type: none"> DASY4/5 System Handbook <p>Method Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Peak Power Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: D450V3-1103_Apr21 Page 2 of 6</p>
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																																																						
Power meter NIP-291	SN: 100778	09-Apr-21 (No. 217-0021-0320)	Apr-22																																																						
Power sensor NIP-291	SN: 100284	09-Apr-21 (No. 217-0021)	Apr-22																																																						
Power sensor NIP-291	SN: 103219	09-Apr-21 (No. 217-0021)	Apr-22																																																						
Reference 20 dB Attenuator	SN: CC2652 (200)	09-Apr-21 (No. 217-0043)	Apr-22																																																						
Type-N minmax combination	SN: 310982 / 06327	09-Apr-21 (No. 217-0044)	Apr-22																																																						
Reference Probe EX3DV4	SN: 3877	30-Dec-20 (No. EX3-3077- Dec20)	Dec-21																																																						
DAE4	SN: 954	28-Jun-20 (No. DAE4-954_Jun20)	Jun-21																																																						
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																																																						
Power meter E41198	SN: G411290874	09-Apr-16 (in house check Jun-20)	In house check: Jun-22																																																						
Power meter E41198	SN: 9411290874	09-Apr-16 (in house check Jun-20)	In house check: Jun-22																																																						
Power sensor E4412A	SN: 200116019	09-Apr-19 (in house check Jun-20)	In house check: Jun-22																																																						
RF generator HP 8648C	EN: URS440-011700	09-Aug-19 (in house check Jun-20)	In house check: Jun-22																																																						
Network Analyzer Agilent E3036A	SN: US410960477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21																																																						


Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



Measurement Conditions DASY system configuration, as far as not given on page 1. <table border="1"> <tr><td>DASY Version</td><td>DASY5</td><td>V62.10.4</td></tr> <tr><td>Extrapolation</td><td colspan="2">Advanced Extrapolation</td></tr> <tr><td>Phantom</td><td>ELI Flat Phantom</td><td>Shell thickness: 2 ± 0.2 mm</td></tr> <tr><td>Distance Dipole Center - TSL</td><td>15 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>$\Delta x, \Delta y, \Delta z = 6 \text{ mm}$</td><td></td></tr> <tr><td>Frequency</td><td>450 MHz ± 1 MHz</td><td></td></tr> </table>		DASY Version	DASY5	V62.10.4	Extrapolation	Advanced Extrapolation		Phantom	ELI Flat Phantom	Shell thickness: 2 ± 0.2 mm	Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	$\Delta x, \Delta y, \Delta z = 6 \text{ mm}$		Frequency	450 MHz ± 1 MHz		Appendix (Additional assessments outside the scope of SCS 0108) Antenna Parameters with Head TSL <table border="1"> <tr><td>Impedance, transformed to feed point</td><td>57.1 Ω - 2.8 jΩ</td></tr> <tr><td>Return Loss</td><td>-23.0 dB</td></tr> </table> General Antenna Parameters and Design <table border="1"> <tr><td>Electrical Delay (one direction)</td><td>1.346 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured. The dipole is made of standard semi-glossy copper cable. The outer conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the arms, the small end caps are added to the dipole arms in order to improve matching when loaded according to the test and explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.</p>		Impedance, transformed to feed point	57.1 Ω - 2.8 jΩ	Return Loss	-23.0 dB	Electrical Delay (one direction)	1.346 ns
DASY Version	DASY5	V62.10.4																									
Extrapolation	Advanced Extrapolation																										
Phantom	ELI Flat Phantom	Shell thickness: 2 ± 0.2 mm																									
Distance Dipole Center - TSL	15 mm	with Spacer																									
Zoom Scan Resolution	$\Delta x, \Delta y, \Delta z = 6 \text{ mm}$																										
Frequency	450 MHz ± 1 MHz																										
Impedance, transformed to feed point	57.1 Ω - 2.8 jΩ																										
Return Loss	-23.0 dB																										
Electrical Delay (one direction)	1.346 ns																										
Head TSL parameters The following parameters and calculations were applied. <table border="1"> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>43.5</td><td>0.57 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2 °C)</td><td>43.1 ± 6 %</td><td>0.67 mho/m ± 8 %</td></tr> <tr><td>Head TSL temperature change during test</td><td>< 0.5 °C</td><td>----</td><td>-----</td></tr> </table>		Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m	Measured Head TSL parameters	(22.0 ± 0.2 °C)	43.1 ± 6 %	0.67 mho/m ± 8 %	Head TSL temperature change during test	< 0.5 °C	----	-----	Additional EUT Data <table border="1"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table>		Manufactured by	SPEAG										
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m																								
Measured Head TSL parameters	(22.0 ± 0.2 °C)	43.1 ± 6 %	0.67 mho/m ± 8 %																								
Head TSL temperature change during test	< 0.5 °C	----	-----																								
Manufactured by	SPEAG																										
SAR result with Head TSL <table border="1"> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>Condition</td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>1.14 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>4.86 W/kg ± 18.1 % (Ko2)</td></tr> </table> <table border="1"> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>0.757 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>3.06 W/kg ± 17.6 % (Ko2)</td></tr> </table>		SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured	250 mW input power	1.14 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	4.86 W/kg ± 18.1 % (Ko2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	SAR measured	250 mW input power	0.757 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg ± 17.6 % (Ko2)	Certificate No: D450V3-1103_Apr21 Page 3 of 6									
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																										
SAR measured	250 mW input power	1.14 W/kg																									
SAR for nominal Head TSL parameters	normalized to 1W	4.86 W/kg ± 18.1 % (Ko2)																									
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																										
SAR measured	250 mW input power	0.757 W/kg																									
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg ± 17.6 % (Ko2)																									
DASY5 Validation Report for Head TSL Test Laboratory: SPEAG, Zurich, Switzerland DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103 Communication System: UID 0 - CW; Frequency: 450 MHz Medium parameters used: $\epsilon' = 450 \text{ MHz}$; $\sigma = 0.87 \text{ S/m}$; $\epsilon_r = 43.1$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011) DASY52 Configuration: <ul style="list-style-type: none"> Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn654; Calibrated: 26.06.2020 Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP:1003 DASY52.52.10.4(1527); SEMCAD X 14.6.14(7483) Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7) /Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 39.18 W/kg; Power Drift = -0.08 dB Peak SAR (extrapolated) = 1.76 W/kg SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg Smallest distance from peaks to all points 3 dB below: Larger than measurement grid Ratio of SAR at M2 to SAR at M1 = 64.9% Maximum value of SAR (measured) = 1.53 W/kg		Certificate No: D450V3-1103_Apr21 Page 4 of 6 																									
Certificate No: D450V3-1103_Apr21 Page 5 of 6		Certificate No: D450V3-1103_Apr21 Page 6 of 6																									

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com



1.3 D750V3 - SN 1188

    <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-62305012 Fax: +86-10-62304633-2504 E-mail: cttf@chinaitt.com http://www.chinaitt.com</p> <p>Client: SGS-CN Certificate No: Z22-60103</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D750V3 - SN: 1188</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 29, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibration have been conducted in the closed laboratory facility: environment temperature (22±3)°C; and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Power Meter NRP2</td> <td>105277</td> <td>24-Sep-21 (CTTL, No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL, No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4 DAE4</td> <td>SN 7307 SN 1556</td> <td>26-May-21 (SPEAG No EX3-7307, May21) 12-Jan-22 (CTTL-SPEAG, No.Z22-60007)</td> <td>May-22 Jan-23</td> </tr> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL, No.J22X04049)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL, No.J22X04046)</td> <td>Jan-23</td> </tr> </table> <p>Calibrated by: Zhao Jing Reviewed by: Lin Hao Approved by: Qi Dianyuan</p> <p>Issued: April 3, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>		Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	105277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22	Reference Probe EX3DV4 DAE4	SN 7307 SN 1556	26-May-21 (SPEAG No EX3-7307, May21) 12-Jan-22 (CTTL-SPEAG, No.Z22-60007)	May-22 Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X04049)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X04046)	Jan-23	  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: cttf@chinaitt.com http://www.chinaitt.com</p> <p>Glossary: TSL tissue simulating liquid ConvF sensitivity in TSL / NORMx,y,z N/A not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for the Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices - Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 655564, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measured value at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</p> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</p>																																
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																											
Power Meter NRP2	105277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22																																																											
Power sensor NRP8S	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22																																																											
Reference Probe EX3DV4 DAE4	SN 7307 SN 1556	26-May-21 (SPEAG No EX3-7307, May21) 12-Jan-22 (CTTL-SPEAG, No.Z22-60007)	May-22 Jan-23																																																											
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																											
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X04049)	Jan-23																																																											
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X04046)	Jan-23																																																											
<p>Certificate No: Z22-60103 Page 1 of 6</p> <p>    <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: cttf@chinaitt.com http://www.chinaitt.com</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY2</td> <td>VS2.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>750 MHz ± 1 MHz</td> <td></td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr> <td></td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>42.0</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.4 ± 6 %</td> <td>0.89 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW Input power</td> <td>2.07 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>8.27 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW Input power</td> <td>1.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>5.48 W/kg ± 18.7 % (k=2)</td> </tr> </table> </p>		DASY Version	DASY2	VS2.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	750 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW Input power	2.07 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW Input power	1.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)	<p>Certificate No: Z22-60103 Page 2 of 6</p> <p>  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: cttf@chinaitt.com http://www.chinaitt.com</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>53.6Ω-1.13jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-28.7dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>0.947 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by the use of end caps, as long as the dipole is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> </p>	Impedance, transformed to feed point	53.6Ω-1.13jΩ	Return Loss	-28.7dB	Electrical Delay (one direction)	0.947 ns	Manufactured by	SPEAG
DASY Version	DASY2	VS2.10.4																																																												
Extrapolation	Advanced Extrapolation																																																													
Phantom	Triple Flat Phantom 5.1C																																																													
Distance Dipole Center - TSL	15 mm	with Spacer																																																												
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																													
Frequency	750 MHz ± 1 MHz																																																													
	Temperature	Permittivity	Conductivity																																																											
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m																																																											
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %																																																											
Head TSL temperature change during test	<1.0 °C	---	---																																																											
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																													
SAR measured	250 mW Input power	2.07 W/kg																																																												
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)																																																												
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																													
SAR measured	250 mW Input power	1.37 W/kg																																																												
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)																																																												
Impedance, transformed to feed point	53.6Ω-1.13jΩ																																																													
Return Loss	-28.7dB																																																													
Electrical Delay (one direction)	0.947 ns																																																													
Manufactured by	SPEAG																																																													
<p>Certificate No: Z22-60103 Page 3 of 6</p>		<p>Certificate No: Z22-60103 Page 4 of 6</p>																																																												


Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: to check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com


No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com






1.4 D835V2 - SN 4d114



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com




<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctitl@chinaitl.com http://www.chinaitl.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>DASY Version</td><td>DASY52</td><td>V52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>15 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>835 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Nominal Head TSL parameters</td><td>Temperature</td><td>Permittivity</td><td>Conductivity</td></tr> <tr><td>22.0 °C</td><td>41.5</td><td>0.90 mho/m</td><td></td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>41.0 ± 6 %</td><td>0.91 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>2.37 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>9.49 W/kg ± 18.8 % (n=2)</td></tr> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>1.54 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>6.12 W/kg ± 18.7 % (n=2)</td></tr> </table> </div>	DASY Version	DASY52	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	835 MHz ± 1 MHz		Nominal Head TSL parameters	Temperature	Permittivity	Conductivity	22.0 °C	41.5	0.90 mho/m		Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 6 %	0.91 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	9.49 W/kg ± 18.8 % (n=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.54 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (n=2)	<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctitl@chinaitl.com http://www.chinaitl.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Impedance, transformed to feed point</td><td>48.7Ω - 5.22jΩ</td></tr> <tr><td>Return Loss</td><td>-25.3dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Electrical Delay (one direction)</td><td>1.307 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the standard. The antenna is not loaded according to the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table> </div>	Impedance, transformed to feed point	48.7Ω - 5.22jΩ	Return Loss	-25.3dB	Electrical Delay (one direction)	1.307 ns	Manufactured by	SPEAG
DASY Version	DASY52	V52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	835 MHz ± 1 MHz																																																												
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																										
22.0 °C	41.5	0.90 mho/m																																																											
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 6 %	0.91 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	—	—																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.37 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	9.49 W/kg ± 18.8 % (n=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.54 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (n=2)																																																											
Impedance, transformed to feed point	48.7Ω - 5.22jΩ																																																												
Return Loss	-25.3dB																																																												
Electrical Delay (one direction)	1.307 ns																																																												
Manufactured by	SPEAG																																																												
Certificate No: Z22-60104 Page 2 of 6																																																													
Certificate No: Z22-60104 Date: 2022-03-31																																																													
DASY5's Variations Report for Head TSL Test Laboratory: CAICT, Beijing China DUT: Dipole 835 MHz; Type: DR35V2; Serial: 4d114 Communication System: UID 0; CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium parameters used: $\epsilon' = 835 \text{ MHz}$; $\sigma = 0.907 \text{ S/m}$; $\tau_s = 40.98$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: <ul style="list-style-type: none"> • Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26 • Sensor-Surface: 1.4mm (Mechanical Surface Detection) • Electronics: DAE4.5n1556; Calibrated: 2022-01-12 • Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 • DASY5 52.1.0.4(1535); SEMICAD X 14.6.14(7501) <p>Diagrams: Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm. Reference Value = 57.88 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg Smallest distance from peaks to all points 3 dB below = 15.8 mm Ratio of SAR at M2 to SAR at M1 = 66.2% Maximum value of SAR (measured) = 3.17 W/kg  <p>0 dB = 3.17 W/kg = 5.01 dBW/kg</p> </p>																																																													
Certificate No: Z22-60104 Page 5 of 6																																																													
Certificate No: Z22-60104 Page 6 of 6																																																													

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: to check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com



1.5 D900V2 - SN 1d079

   <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-62704531-2117 E-mail: offi@caict.ac.cn http://www.caict.ac.cn</p> <p>Client: SGS-CN Certificate No: Z22-60184</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D900V2 - SN: 1d079</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility, environment temperature (22±5)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration):</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter: NIP82</td> <td>10277</td> <td>24-Jan-21 (CTTL, No. J22X00260)</td> <td>Sept-22</td> </tr> <tr> <td>Power sensor: NIP85</td> <td>10491</td> <td>24-Sep-21 (CTTL, No. J21X03262)</td> <td>Sept-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7464</td> <td>26-Jan-22(SPEAG No EX3-7464, Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1558</td> <td>12-Jan-22(CTTL-SPEAG No Z22-80007)</td> <td>Jan-23</td> </tr> <tr> <td>Secondary Standards</td> <td>ID #</td> <td>Cal Date (Calibrated by, Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL, No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL, No. J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Qi Dianyuan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p>Issued: June 13, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60184 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter: NIP82	10277	24-Jan-21 (CTTL, No. J22X00260)	Sept-22	Power sensor: NIP85	10491	24-Sep-21 (CTTL, No. J21X03262)	Sept-22	Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23	DAE4	SN 1558	12-Jan-22(CTTL-SPEAG No Z22-80007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No. J22X00409)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No. J22X00406)	Jan-23	Name	Function	Signature	Zhao Jing	SAR Test Engineer		Lin Hao	SAR Test Engineer		Qi Dianyuan	SAR Project Leader		  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62704531-3217 E-mail: offi@caict.ac.cn http://www.caict.ac.cn</p> <p>Glossary:</p> <p>TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMxyz N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for the Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 895994, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation:</p> <p>o DASY4.5 System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> • Antenna Configuration: Further details are available from the Validation Report at the end of the certificate. All parameters in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance value is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • SAR measured: One day delay between the SMA connector and the antenna feed point. • No uncertainty required: No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: Z22-60184 Page 2 of 6</p>																
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter: NIP82	10277	24-Jan-21 (CTTL, No. J22X00260)	Sept-22																																																										
Power sensor: NIP85	10491	24-Sep-21 (CTTL, No. J21X03262)	Sept-22																																																										
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23																																																										
DAE4	SN 1558	12-Jan-22(CTTL-SPEAG No Z22-80007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No. J22X00409)	Jan-23																																																										
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No. J22X00406)	Jan-23																																																										
Name	Function	Signature																																																											
Zhao Jing	SAR Test Engineer																																																												
Lin Hao	SAR Test Engineer																																																												
Qi Dianyuan	SAR Project Leader																																																												
  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62704531-2117 E-mail: offi@caict.ac.cn http://www.caict.ac.cn</p> <p>Measurement System Configuration: DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY92</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters: The following parameters and calculations were applied.</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>41.5</td> <td>0.97 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>42.1 ± 6 %</td> <td>0.98 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL:</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.76 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>11.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.78 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>7.09 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: Z22-60184 Page 3 of 6</p>	DASY Version	DASY92	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	900 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.76 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.78 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)	  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62704531-3217 E-mail: offi@caict.ac.cn http://www.caict.ac.cn</p> <p>Appendix: Additional assessments outside the scope of CNAS L0570:</p> <p>Antenna Parameters with Head TSL:</p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>48.10 ± 8.49Ω</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-23.3 dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design:</p> <table border="1"> <thead> <tr> <th>Electrical (one direction)</th> <th>1.312 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard enameled copper cable. The center conductor of the testing line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded with the liquid phantom explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The dipole impedance is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data:</p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> <p>Certificate No: Z22-60184 Page 4 of 6</p>	Impedance, transformed to feed point	48.10 ± 8.49Ω	Return Loss	-23.3 dB	Electrical (one direction)	1.312 ns	Manufactured by	SPEAG
DASY Version	DASY92	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	900 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.76 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.78 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	48.10 ± 8.49Ω																																																												
Return Loss	-23.3 dB																																																												
Electrical (one direction)	1.312 ns																																																												
Manufactured by	SPEAG																																																												


Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





1.6 D1800V2 - SN 2d170

<p>Client: SGS-CN Certificate No: Z22-60105</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D1800V2 - SN: 2d170</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 31, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity >70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (C TTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP85</td> <td>104291</td> <td>24-Sep-21 (C TTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-22 (SPEAG No. EX-7307, May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22 (C TTL, SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> </table> <table border="1"> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (C TTL, No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY48110673</td> <td>14-Jan-22 (C TTL, No. J22X00406)</td> <td>Jan-23</td> </tr> </table> <p>Calibrated by: Zhao Jing Reviewed by: Lin Hao Approved by: Qi Dianyuan</p> <p>Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60105 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (C TTL, No. J21X08326)	Sep-22	Power sensor NRP85	104291	24-Sep-21 (C TTL, No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-22 (SPEAG No. EX-7307, May21)	May-22	DAE4	SN 1556	12-Jan-22 (C TTL, SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (C TTL, No. J22X00409)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (C TTL, No. J22X00406)	Jan-23	<p>Client: SGS-CN Certificate No: Z22-60105</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D1800V2 - SN: 2d170</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 31, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity >70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (C TTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP85</td> <td>104291</td> <td>24-Sep-21 (C TTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-22 (SPEAG No. EX-7307, May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22 (C TTL, SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> </table> <table border="1"> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (C TTL, No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY48110673</td> <td>14-Jan-22 (C TTL, No. J22X00406)</td> <td>Jan-23</td> </tr> </table> <p>Calibrated by: Zhao Jing Reviewed by: Lin Hao Approved by: Qi Dianyuan</p> <p>Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60105 Page 2 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (C TTL, No. J21X08326)	Sep-22	Power sensor NRP85	104291	24-Sep-21 (C TTL, No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-22 (SPEAG No. EX-7307, May21)	May-22	DAE4	SN 1556	12-Jan-22 (C TTL, SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (C TTL, No. J22X00409)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (C TTL, No. J22X00406)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																														
Power Meter NRP2	106277	24-Sep-21 (C TTL, No. J21X08326)	Sep-22																																																														
Power sensor NRP85	104291	24-Sep-21 (C TTL, No. J21X08326)	Sep-22																																																														
Reference Probe EX3DV4	SN 7307	26-May-22 (SPEAG No. EX-7307, May21)	May-22																																																														
DAE4	SN 1556	12-Jan-22 (C TTL, SPEAG No. Z22-60007)	Jan-23																																																														
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																														
Signal Generator E4438C	MY49071430	13-Jan-22 (C TTL, No. J22X00409)	Jan-23																																																														
Network Analyzer E5071C	MY48110673	14-Jan-22 (C TTL, No. J22X00406)	Jan-23																																																														
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																														
Power Meter NRP2	106277	24-Sep-21 (C TTL, No. J21X08326)	Sep-22																																																														
Power sensor NRP85	104291	24-Sep-21 (C TTL, No. J21X08326)	Sep-22																																																														
Reference Probe EX3DV4	SN 7307	26-May-22 (SPEAG No. EX-7307, May21)	May-22																																																														
DAE4	SN 1556	12-Jan-22 (C TTL, SPEAG No. Z22-60007)	Jan-23																																																														
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																														
Signal Generator E4438C	MY49071430	13-Jan-22 (C TTL, No. J22X00409)	Jan-23																																																														
Network Analyzer E5071C	MY48110673	14-Jan-22 (C TTL, No. J22X00406)	Jan-23																																																														



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)5735888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)5735888 f(86-512)57370818 sgs.china@sgs.com

<p>TTL <small>In Collaboration with</small> SPeAG CALIBRATION LABORATORY</p> <p>Add: No.53 HuaYanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304613-2079 Fax: +86-10-62304613-2504 E-mail: ctt@chinattt.cn http://www.chinattt.cn</p> <p>CAICT</p> <p>Measurement Conditions DASY5 system configuration, as far as not given on page 1.</p> <table border="1"> <tr><td>DASY Version</td><td>DASY52</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>1800 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>40.0</td><td>1.40 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>40.6 ± 6 %</td><td>1.41 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr><th>SAR averaged over 1 cm³ (1 g) of Head TSL</th><th>Condition</th></tr> <tr><td>SAR measured</td><td>250 mW input power 9.73 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W 38.9 W/kg ± 18.8 % (k=2)</td></tr> <tr><th>SAR averaged over 10 cm³ (10 g) of Head TSL</th><th>Condition</th></tr> <tr><td>SAR measured</td><td>250 mW input power 5.11 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W 20.4 W/kg ± 18.7 % (k=2)</td></tr> </table> <p>Certificate No: Z22-60105 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1800 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	40.6 ± 6 %	1.41 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	SAR measured	250 mW input power 9.73 W/kg	SAR for nominal Head TSL parameters	normalized to 1W 38.9 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	SAR measured	250 mW input power 5.11 W/kg	SAR for nominal Head TSL parameters	normalized to 1W 20.4 W/kg ± 18.7 % (k=2)	<p>TTL <small>In Collaboration with</small> SPeAG CALIBRATION LABORATORY</p> <p>Add: No.53 HuaYanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304613-2079 Fax: +86-10-62304613-2504 E-mail: ctt@chinattt.cn http://www.chinattt.cn</p> <p>CAICT</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr><td>Impedance, transformed to feed point</td><td>47.90- 2.54jΩ</td></tr> <tr><td>Return Loss</td><td>-29.4dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr><td>Electrical Delay (one direction)</td><td>1.116 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the standard. The overall dipole length is still according to the Standard. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr><td>Manufactured by</td><td>SPeAG</td></tr> </table> <p>Certificate No: Z22-60105 Page 4 of 6</p>	Impedance, transformed to feed point	47.90- 2.54jΩ	Return Loss	-29.4dB	Electrical Delay (one direction)	1.116 ns	Manufactured by	SPeAG
DASY Version	DASY52	52.10.4																																																					
Extrapolation	Advanced Extrapolation																																																						
Phantom	Triple Flat Phantom 5.1C																																																						
Distance Dipole Center - TSL	10 mm	with Spacer																																																					
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																						
Frequency	1800 MHz ± 1 MHz																																																						
	Temperature	Permittivity	Conductivity																																																				
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m																																																				
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.6 ± 6 %	1.41 mho/m ± 6 %																																																				
Head TSL temperature change during test	<1.0 °C	—	—																																																				
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																						
SAR measured	250 mW input power 9.73 W/kg																																																						
SAR for nominal Head TSL parameters	normalized to 1W 38.9 W/kg ± 18.8 % (k=2)																																																						
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																						
SAR measured	250 mW input power 5.11 W/kg																																																						
SAR for nominal Head TSL parameters	normalized to 1W 20.4 W/kg ± 18.7 % (k=2)																																																						
Impedance, transformed to feed point	47.90- 2.54jΩ																																																						
Return Loss	-29.4dB																																																						
Electrical Delay (one direction)	1.116 ns																																																						
Manufactured by	SPeAG																																																						
<p>TTL <small>In Collaboration with</small> SPeAG CALIBRATION LABORATORY</p> <p>Add: No.53 HuaYanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304613-2079 Fax: +86-10-62304613-2504 E-mail: ctt@chinattt.cn http://www.chinattt.cn</p> <p>CAICT</p> <p>DASV5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China Date: 2022-03-31</p> <p>DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 2d170 Communication System: UUD 0; CW; Frequency: 1800 MHz; Duty Cycle: 1:1 Medium parameters used: $\epsilon = 1800 \text{ MHz}$; $\sigma = 1.411 \text{ S/m}$; $\epsilon_r = 40.62$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Right Section Measurement Standard: DASV5 (IEEE/IEC/ANSI C63.19-2007)</p> <p>DASV5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SNT307; ConvF(8.34, 8.34, 8.34) @ 1800 MHz; Calibrated: 2021-05-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V3.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASV52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Probe Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 98.14 V/m; Power Drift = 0.03 dB Peak SAR (1 g) = 9.73 W/kg; SAR(10 g) = 38.9 W/kg Smallest distance from peaks to all points 3 dB below = 10 mm Ratio of SAR at M2 to SAR at M1 = 54% Maximum value of SAR (measured) = 15.2 W/kg</p> <p>Figure: A 3D heatmap showing SAR distribution. The color scale ranges from -17.12 dB to 0 dB. The text at the bottom indicates $0 \text{ dB} = 15.2 \text{ W/kg} = 11.82 \text{ dBW/kg}$.</p> <p>Certificate No: Z22-60105 Page 5 of 6</p>	<p>TTL <small>In Collaboration with</small> SPeAG CALIBRATION LABORATORY</p> <p>Add: No.53 HuaYanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304613-2079 Fax: +86-10-62304613-2504 E-mail: ctt@chinattt.cn http://www.chinattt.cn</p> <p>CAICT</p> <p>Impedance Measurement Plot for Head TSL</p> <p>Certificate No: Z22-60105 Page 6 of 6</p>																																																						

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: to check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.7 D1900V2 - SN 5d136

<div style="text-align: center;">   <p>In Collaboration with SGS CALIBRATION LABORATORY Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-62904532-2117 E-mail: sgs.cn@kcaict.com Client: SGS-CN Certificate No: Z22-60185</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: D1900V2 - SN: 5d136</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility, environment temperature (22±5)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE criteria for calibration)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power sensor NRP2</td> <td>102277</td> <td>24-Jan-22 (CTTL No.J22X00409)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP85</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X00326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7464</td> <td>26-Jan-22(SPEAG No EX3-7464, Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAEA</td> <td>SN 1558</td> <td>12-Jan-22(CTTL-SPEAG No 222-6007)</td> <td>Jan-23</td> </tr> <tr> <td>Secondary Standards</td> <td>ID #</td> <td>Cal Date (Calibrated by, Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY48110673</td> <td>14-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing SAR Test Engineer</p> <p>Reviewed by: Lin Hao SAR Test Engineer</p> <p>Approved by: Qi Danyuan SAR Project Leader</p> <p>Issued: June 13, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory</p> <p>Certificate No: Z22-60185 Page 1 of 6</p> </div>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power sensor NRP2	102277	24-Jan-22 (CTTL No.J22X00409)	Sep-22	Power sensor NRP85	104291	24-Sep-21 (CTTL No.J21X00326)	Sep-22	Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23	DAEA	SN 1558	12-Jan-22(CTTL-SPEAG No 222-6007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23	<div style="text-align: center;">   <p>In Collaboration with SGS CALIBRATION LABORATORY Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62904532-2117 E-mail: sgs.cn@kcaict.com http://www.caict.ac.cn</p> <p>Glossary:</p> <p>TSL tissue simulating liquid Conv/F sensitivity in TSL / NORMLxyz N/A not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for the Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Personal Communication Devices- Part 1528: Human Models, Instrumentation and Procedure (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation:</p> <p>DASY4.3 System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of this document. All figures and tables are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transferred from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: The electrical delay between the SMA connector and the antenna feed point. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> </div> <p>Certificate No: Z22-60185 Page 2 of 6</p> </div>																														
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Power sensor NRP2	102277	24-Jan-22 (CTTL No.J22X00409)	Sep-22																																																												
Power sensor NRP85	104291	24-Sep-21 (CTTL No.J21X00326)	Sep-22																																																												
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23																																																												
DAEA	SN 1558	12-Jan-22(CTTL-SPEAG No 222-6007)	Jan-23																																																												
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																																																												
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23																																																												
<div style="text-align: center;">   <p>In Collaboration with SGS CALIBRATION LABORATORY Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62904532-2117 E-mail: sgs.cn@kcaict.com http://www.caict.ac.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>DASY Version</td> <td>DASY2</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>1900 MHz ± 1 MHz</td> <td></td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Temperature</td> <td>23.0 °C</td> <td>Condition</td> <td>52.10.4</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>23.0 °C</td> <td>Permittivity</td> <td>40.0</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>Conductivity</td> <td>1.40 mho/m</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>41.0 °C</td> <td>—</td> <td>—</td> </tr> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>9.95 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>40.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.18 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>20.8 W/kg ± 18.7 % (k=2)</td> </tr> </table> <p>Certificate No: Z22-60185 Page 3 of 6</p> </div>		DASY Version	DASY2	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1900 MHz ± 1 MHz		Temperature	23.0 °C	Condition	52.10.4	Nominal Head TSL parameters	23.0 °C	Permittivity	40.0	Measured Head TSL parameters	(22.0 ± 0.2) °C	Conductivity	1.40 mho/m	Head TSL temperature change during test	41.0 °C	—	—	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	9.95 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.18 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)	<div style="text-align: center;">   <p>In Collaboration with SGS CALIBRATION LABORATORY Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62904532-2117 E-mail: sgs.cn@kcaict.com http://www.caict.ac.cn</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Impedance, transformed to feed point</td> <td>51.20 ± 7.98 jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-22.4 dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Electrical Delay (one direction)</td> <td>1.109 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-biases. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded. The antenna position is explained in the "Measurement Conditions" paragraph. The SAR data are not affected by these changes, because the antenna length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> <p>Certificate No: Z22-60185 Page 4 of 6</p> </div>		Impedance, transformed to feed point	51.20 ± 7.98 jΩ	Return Loss	-22.4 dB	Electrical Delay (one direction)	1.109 ns	Manufactured by	SPEAG
DASY Version	DASY2	52.10.4																																																													
Extrapolation	Advanced Extrapolation																																																														
Phantom	Triple Flat Phantom 5.1C																																																														
Distance Dipole Center - TSL	10 mm	with Spacer																																																													
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																														
Frequency	1900 MHz ± 1 MHz																																																														
Temperature	23.0 °C	Condition	52.10.4																																																												
Nominal Head TSL parameters	23.0 °C	Permittivity	40.0																																																												
Measured Head TSL parameters	(22.0 ± 0.2) °C	Conductivity	1.40 mho/m																																																												
Head TSL temperature change during test	41.0 °C	—	—																																																												
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																														
SAR measured	250 mW input power	9.95 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)																																																													
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																														
SAR measured	250 mW input power	5.18 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)																																																													
Impedance, transformed to feed point	51.20 ± 7.98 jΩ																																																														
Return Loss	-22.4 dB																																																														
Electrical Delay (one direction)	1.109 ns																																																														
Manufactured by	SPEAG																																																														


Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com


No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





1.8 D2000V2 - SN 1041





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_DocCheck@sgs.com



No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttf@caict.ac.cn http://www.caict.ac.cn</p> <p>Measurement Conditions DASY system configuration, as far as given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>DASY Version</td><td>DASY52</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>2000 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>40.0</td><td>1.40 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>40.2 ± 6 %</td><td>1.39 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><th></th><th>Condition</th></tr> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>250 mW input power 10.4 W/kg</td></tr> <tr><td>SAR measured</td><td>normalized to 1W</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>41.8 W/kg ± 18.6 % (n=2)</td></tr> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td></tr> <tr><td>SAR measured</td><td>250 mW input power 5.30 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W 21.3 W/kg ± 18.7 % (n=2)</td></tr> </table> </div>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2000 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 6 %	1.39 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		Condition	SAR averaged over 1 cm ³ (1 g) of Head TSL	250 mW input power 10.4 W/kg	SAR measured	normalized to 1W	SAR for nominal Head TSL parameters	41.8 W/kg ± 18.6 % (n=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	SAR measured	250 mW input power 5.30 W/kg	SAR for nominal Head TSL parameters	normalized to 1W 21.3 W/kg ± 18.7 % (n=2)	<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttf@caict.ac.cn http://www.caict.ac.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Impedance, transformed to feed port</td><td>48.40+ 0.74jΩ</td></tr> <tr><td>Return Loss</td><td>-34.9dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Electric Delay (one direction)</td><td>1.088 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the testing line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded with the probe. Measurement conditions: Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still adapted to the phantom.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table> </div>	Impedance, transformed to feed port	48.40+ 0.74jΩ	Return Loss	-34.9dB	Electric Delay (one direction)	1.088 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																							
Extrapolation	Advanced Extrapolation																																																								
Phantom	Triple Flat Phantom 5.1C																																																								
Distance Dipole Center - TSL	10 mm	with Spacer																																																							
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																								
Frequency	2000 MHz ± 1 MHz																																																								
	Temperature	Permittivity	Conductivity																																																						
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m																																																						
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 6 %	1.39 mho/m ± 6 %																																																						
Head TSL temperature change during test	<1.0 °C	—	—																																																						
	Condition																																																								
SAR averaged over 1 cm ³ (1 g) of Head TSL	250 mW input power 10.4 W/kg																																																								
SAR measured	normalized to 1W																																																								
SAR for nominal Head TSL parameters	41.8 W/kg ± 18.6 % (n=2)																																																								
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																								
SAR measured	250 mW input power 5.30 W/kg																																																								
SAR for nominal Head TSL parameters	normalized to 1W 21.3 W/kg ± 18.7 % (n=2)																																																								
Impedance, transformed to feed port	48.40+ 0.74jΩ																																																								
Return Loss	-34.9dB																																																								
Electric Delay (one direction)	1.088 ns																																																								
Manufactured by	SPEAG																																																								
<p>Certificate No: Z22-60186 Page 3 of 6</p> <p>Certificate No: Z22-60186 Page 4 of 6</p>																																																									
<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttf@caict.ac.cn http://www.caict.ac.cn</p> <p>DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China Date: 2022-06-06 DUT: Dipole 2000 MHz, Type: D2000V2; Serial: D2000V2 - SN: 1041 Customer Selection: UUD 0, CW, Frequency: 2000 MHz, Duty Cycle: 1:1 Medium parameters used: f = 2000 MHz, ε = 1.392 Sim, σ = 40.21; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> • Probe: EX304V4 - SN7464; ConvF(8.2, 8.2, 8.2) @ 2000 MHz; Calibrated: 2022-01-26 • Sensor-Surface: 1.4mm (Mechanical Surface Detection) • Electronics: DAE4 Snt156; Calibrated: 2022-01-12 • Phantom: MPP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 • DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 103.4 V/m Power Drift = 0.03 dB Peak SAR (extrapolated) = 19.6 W/kg SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg Smallest distance from peaks to all points 3 dB below = 9.1 mm Ratio of SAR at M2 to SAR at M1 = 53.6% Maximum value of SAR (measured) = 16.3 W/kg</p>  </div>	<div style="text-align: center;">  <p>In Collaboration with CAICT</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttf@caict.ac.cn http://www.caict.ac.cn</p> <p>Impedance Measurement Plot for Head TSL</p>  </div>																																																								
<p>Certificate No: Z22-60186 Page 5 of 6</p> <p>Certificate No: Z22-60186 Page 6 of 6</p>																																																									

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com




1.9 D2300V2 - SN 1096

  <p>Object: D2300V2 - SN: 1096</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 31, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.</p> <p>Calibration Equipment used (MATE critical for calibration)</p> <table border="1"> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL, No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-21 (SPEAG No. EX3-7307, May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22 (CTTL-SPEAG No. 222-60007)</td> <td>Jan-23</td> </tr> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL, No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL, No. J22X00406)</td> <td>Jan-23</td> </tr> </table> <p>Calibrated by: Zhao Jing Reviewed by: Lin Hao Approved by: Qi Dianyuan</p> <p>Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>		Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL, No. J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL, No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No. EX3-7307, May21)	May-22	DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. 222-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No. J22X00409)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No. J22X00406)	Jan-23	  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctli@chinaitt.com http://www.chinaitt.cn</p> <p>Glossary: TSL tissue simulating liquid ConvF sensitivity in TSL / NORMx,y,z N/A not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for the Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 655664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measured value at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</p> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</p>																												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																											
Power Meter NRP2	106277	24-Sep-21 (CTTL, No. J21X08326)	Sep-22																																																											
Power sensor NRP8S	104291	24-Sep-21 (CTTL, No. J21X08326)	Sep-22																																																											
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No. EX3-7307, May21)	May-22																																																											
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. 222-60007)	Jan-23																																																											
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																											
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No. J22X00409)	Jan-23																																																											
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No. J22X00406)	Jan-23																																																											
<p>Certificate No: Z22-60106 Page 1 of 6</p> <p>   <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctli@chinaitt.com http://www.chinaitt.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY2</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz < 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2300 MHz ± 1 MHz</td> <td></td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr> <td></td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.5</td> <td>1.67 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.8 ± 6 %</td> <td>1.70 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </table> <p>SAR result with Head TSL</p> <table border="1"> <tr> <td>SAR averaged over 1 cm² (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>12.4 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>49.2 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm² (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.88 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>23.4 W/kg ± 18.7 % (k=2)</td> </tr> </table> </p>		DASY Version	DASY2	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz < 5 mm		Frequency	2300 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—	SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	12.4 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.88 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)	<p>Certificate No: Z22-60106 Page 2 of 6</p> <p>  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctli@chinaitt.com http://www.chinaitt.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>49.2Ω-4.5Ω</td> </tr> <tr> <td>Return Loss</td> <td>-26.6dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.083 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the standard. The antenna is not loaded in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> </p>	Impedance, transformed to feed point	49.2Ω-4.5Ω	Return Loss	-26.6dB	Electrical Delay (one direction)	1.083 ns	Manufactured by	SPEAG
DASY Version	DASY2	52.10.4																																																												
Extrapolation	Advanced Extrapolation																																																													
Phantom	Triple Flat Phantom 5.1C																																																													
Distance Dipole Center - TSL	10 mm	with Spacer																																																												
Zoom Scan Resolution	dx, dy, dz < 5 mm																																																													
Frequency	2300 MHz ± 1 MHz																																																													
	Temperature	Permittivity	Conductivity																																																											
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m																																																											
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %																																																											
Head TSL temperature change during test	<1.0 °C	—	—																																																											
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition																																																													
SAR measured	250 mW input power	12.4 W/kg																																																												
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)																																																												
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition																																																													
SAR measured	250 mW input power	5.88 W/kg																																																												
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)																																																												
Impedance, transformed to feed point	49.2Ω-4.5Ω																																																													
Return Loss	-26.6dB																																																													
Electrical Delay (one direction)	1.083 ns																																																													
Manufactured by	SPEAG																																																													

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com





1.10 D2450V2 - SN 817





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



<div style="text-align: center;">  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctfl@chinattl.com http://www.chinattl.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>DASY Version</td><td>DASY52</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr> <tr><td>Frequency</td><td>2450 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>39.2</td><td>1.80 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>39.5 ± 6 %</td><td>1.79 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>13.2 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>83.0 W/kg ± 18.8 % (k=2)</td></tr> <tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>6.15 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>24.7 W/kg ± 18.7 % (k=2)</td></tr> </table> </div>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2450 MHz ± 1 MHz		Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	13.2 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	83.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	6.15 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	<div style="text-align: center;">  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctfl@chinattl.com http://www.chinattl.cn</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Impedance, transformed to feed point</td><td>52.10 ± 3.20Ω</td></tr> <tr><td>Return Loss</td><td>-28.5dB</td></tr> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Electrical Delay (one direction)</td><td>1.066 ns</td></tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the standard. The overall dipole length is still according to the Standard. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>Manufactured by</td><td>SPEAG</td></tr> </table> </div>	Impedance, transformed to feed point	52.10 ± 3.20Ω	Return Loss	-28.5dB	Electrical Delay (one direction)	1.066 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																							
Extrapolation	Advanced Extrapolation																																																								
Phantom	Triple Flat Phantom 5.1C																																																								
Distance Dipole Center - TSL	10 mm	with Spacer																																																							
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																								
Frequency	2450 MHz ± 1 MHz																																																								
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m																																																						
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %																																																						
Head TSL temperature change during test	<1.0 °C	—	—																																																						
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																								
SAR measured	250 mW input power	13.2 W/kg																																																							
SAR for nominal Head TSL parameters	normalized to 1W	83.0 W/kg ± 18.8 % (k=2)																																																							
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																								
SAR measured	250 mW input power	6.15 W/kg																																																							
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)																																																							
Impedance, transformed to feed point	52.10 ± 3.20Ω																																																								
Return Loss	-28.5dB																																																								
Electrical Delay (one direction)	1.066 ns																																																								
Manufactured by	SPEAG																																																								
Certificate No: Z22-60107 Page 3 of 6																																																									
<div style="text-align: center;">  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctfl@chinattl.com http://www.chinattl.cn</p> <p>DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing China DUT: Dipole 2450 MHz; Type: D2450V2 - SN: 817 Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium parameters used: $\epsilon' = 2450$ MHz, $\sigma = 1.79$ S/m, $\epsilon_r = 39.52$, $\rho = 1000$ kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY3 Configuration:</p> <ul style="list-style-type: none"> • Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26 • Sensor-Surface: 1.4mm (Mechanical Surface Detection) • Electronics: DAE4 Sn1556; Calibrated: 2022-01-12 • Phantom: MFP_V5.1C (2ddeg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 • DASY52 52.10.4(535); SEMCAD X 14.6.14(550) <p>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)Cu6 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.6 V/m; Power Drift = -0.03 dB Peak SAR (1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg Smallest distance from peaks to all points 3 dB below = 8.9 mm Ratio of SAR at M2 to SAR at M1 = 49.2% Maximum value of SAR (measured) = 22.1 W/kg</p>  </div>	<div style="text-align: center;">  <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctfl@chinattl.com http://www.chinattl.cn</p> <p>Impedance Measurement Plot for Head TSL</p>  </div>																																																								
Certificate No: Z22-60107 Page 5 of 6																																																									
Certificate No: Z22-60107 Page 6 of 6																																																									

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com



1.11 D2600V2 - SN 1158

 <p>In Collaboration with SGS CALIBRATION LABORATORY</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctli@chinatl.cn http://www.chinatl.cn</p> <p>Client SGS-CN Certificate No: Z22-60108</p> <p>CALIBRATION CERTIFICATE</p> <p>Object D2600V2 - SN: 1158</p> <p>Calibration Procedure(s) FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: March 31, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL, No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8</td> <td>104291</td> <td>24-Sep-21 (CTTL, No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-21 (SPEAG No. EX3-7307, May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> <tr> <td>Secondary Standards</td> <td>ID #</td> <td>Cal Date (Calibrated by, Certificate No.)</td> <td>Scheduled Calibration</td> </tr> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL, No.J22X04049)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL, No.J22X04046)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Calibrated by:</th> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyuan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p>Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22	Power sensor NRP8	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No. EX3-7307, May21)	May-22	DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X04049)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X04046)	Jan-23	Calibrated by:	Name	Function	Signature	Zhao Jing	SAR Test Engineer		Reviewed by:	Lin Hao	SAR Test Engineer		Approved by:	Qi Dianyuan	SAR Project Leader		 <p>In Collaboration with CAICT CALIBRATION LABORATORY</p> <p>Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ctli@chinatl.cn http://www.chinatl.cn</p> <p>Glossary: TSL tissue simulating liquid ConvF sensitivity in TSL / NORMx,y,z N/A not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices - Part 1528: Human Models, Instrumentation and Procedure (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 655664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the fat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measured at the SMA connector to the feed point. The Return Loss ensures low reflected power. No absorption required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</p> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</p>													
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	106277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22																																																										
Power sensor NRP8	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22																																																										
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No. EX3-7307, May21)	May-22																																																										
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X04049)	Jan-23																																																										
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X04046)	Jan-23																																																										
Calibrated by:	Name	Function	Signature																																																										
Zhao Jing	SAR Test Engineer																																																												
Reviewed by:	Lin Hao	SAR Test Engineer																																																											
Approved by:	Qi Dianyuan	SAR Project Leader																																																											
<p>Certificate No: Z22-60108 Page 1 of 6</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz < 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2600 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.0</td> <td>1.96 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>38.7 ± 6 %</td> <td>1.96 mho/m ± 8 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>54.8 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>24.5 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz < 5 mm		Frequency	2600 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 8 %	Head TSL temperature change during test	<1.0 °C	—	—	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	13.7 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	6.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.5 W/kg ± 18.7 % (k=2)	<p>Certificate No: Z22-60108 Page 2 of 6</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>49.9Ω-6.4Ω</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-23.8dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>1.053 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the standard. The antenna is mounted on the "Measurement Conditions" paragraph. The SAR data are not affected by the change. The overall dipole is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table>	Impedance, transformed to feed point	49.9Ω-6.4Ω	Return Loss	-23.8dB	Electrical Delay (one direction)	1.053 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz < 5 mm																																																												
Frequency	2600 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 8 %																																																										
Head TSL temperature change during test	<1.0 °C	—	—																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	13.7 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	6.12 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	24.5 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	49.9Ω-6.4Ω																																																												
Return Loss	-23.8dB																																																												
Electrical Delay (one direction)	1.053 ns																																																												
Manufactured by	SPEAG																																																												


Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: to check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com


No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





1.12 D5GHzV2 - SN 1095



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



<p>SGS</p> <p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Add: No.52 HuaYuanBei Road, HuaLian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: ctttg@sohu.com http://www.caict.ac.cn</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr><td>DASY Version</td><td>DASY32</td><td>52.10.4</td></tr> <tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr> <tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr> <tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr> <tr><td>Zoom Scan Resolution</td><td>dx, dy = 4 mm, dz = 1.4 mm</td><td>Graded Ratio = 1.4 (Z direction)</td></tr> <tr><td>Frequency</td><td>4000 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz</td><td></td></tr> </table> <p>Head TSL parameters at 5200MHz The following parameters and calculations were applied.</p> <table border="1"> <tr><td></td><td>Temperature</td><td>Permittivity</td><td>Conductivity</td></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>36.0</td><td>4.66 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>35.4 ± 6 %</td><td>4.62 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL at 5200MHz SAR averaged over 1 cm³ (1 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>7.79 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>77.6 W/kg ± 24.4 % (k=2)</td></tr> </table> <p>SAR result with Head TSL at 5200MHz SAR averaged over 10 cm³ (10 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>250 mW input power</td><td>2.22 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>22.1 W/kg ± 24.2 % (k=2)</td></tr> </table>	DASY Version	DASY32	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	Frequency	4000 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.62 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		Condition		SAR measured	250 mW input power	7.79 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	77.6 W/kg ± 24.4 % (k=2)		Condition		SAR measured	250 mW input power	2.22 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)	<p>CAICT</p> <p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Add: No.52 HuaYuanBei Road, HuaLian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: ctttg@sohu.com http://www.caict.ac.cn</p> <p>Head TSL parameters at 5600MHz The following parameters and calculations were applied.</p> <table border="1"> <tr><td></td><td>Temperature</td><td>Permittivity</td><td>Conductivity</td></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>35.5</td><td>5.07 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>34.7 ± 6 %</td><td>5.05 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL at 5600MHz SAR averaged over 1 cm³ (1 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>100 mW input power</td><td>8.12 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>80.8 W/kg ± 24.4 % (k=2)</td></tr> </table> <p>SAR result with Head TSL at 5600MHz SAR averaged over 10 cm³ (10 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>100 mW input power</td><td>2.30 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>22.8 W/kg ± 24.2 % (k=2)</td></tr> </table> <p>Head TSL parameters at 5800MHz The following parameters and calculations were applied.</p> <table border="1"> <tr><td></td><td>Temperature</td><td>Permittivity</td><td>Conductivity</td></tr> <tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>35.3</td><td>5.27 mho/m</td></tr> <tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>34.4 ± 6 %</td><td>5.25 mho/m ± 6 %</td></tr> <tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>—</td><td>—</td></tr> </table> <p>SAR result with Head TSL at 5800MHz SAR averaged over 1 cm³ (1 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>100 mW input power</td><td>7.71 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>76.7 W/kg ± 24.4 % (k=2)</td></tr> </table> <p>SAR result with Head TSL at 5800MHz SAR averaged over 10 cm³ (10 g) of Head TSL</p> <table border="1"> <tr><td></td><td>Condition</td><td></td></tr> <tr><td>SAR measured</td><td>100 mW input power</td><td>2.16 W/kg</td></tr> <tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>21.5 W/kg ± 24.2 % (k=2)</td></tr> </table>		Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		Condition		SAR measured	100 mW input power	8.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)		Condition		SAR measured	100 mW input power	2.30 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	22.8 W/kg ± 24.2 % (k=2)		Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		Condition		SAR measured	100 mW input power	7.71 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)		Condition		SAR measured	100 mW input power	2.16 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	21.5 W/kg ± 24.2 % (k=2)
DASY Version	DASY32	52.10.4																																																																																																																							
Extrapolation	Advanced Extrapolation																																																																																																																								
Phantom	Triple Flat Phantom 5.1C																																																																																																																								
Distance Dipole Center - TSL	10 mm	with Spacer																																																																																																																							
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)																																																																																																																							
Frequency	4000 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz																																																																																																																								
	Temperature	Permittivity	Conductivity																																																																																																																						
Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m																																																																																																																						
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.62 mho/m ± 6 %																																																																																																																						
Head TSL temperature change during test	<1.0 °C	—	—																																																																																																																						
	Condition																																																																																																																								
SAR measured	250 mW input power	7.79 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	77.6 W/kg ± 24.4 % (k=2)																																																																																																																							
	Condition																																																																																																																								
SAR measured	250 mW input power	2.22 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)																																																																																																																							
	Temperature	Permittivity	Conductivity																																																																																																																						
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m																																																																																																																						
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %																																																																																																																						
Head TSL temperature change during test	<1.0 °C	—	—																																																																																																																						
	Condition																																																																																																																								
SAR measured	100 mW input power	8.12 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)																																																																																																																							
	Condition																																																																																																																								
SAR measured	100 mW input power	2.30 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	22.8 W/kg ± 24.2 % (k=2)																																																																																																																							
	Temperature	Permittivity	Conductivity																																																																																																																						
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m																																																																																																																						
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %																																																																																																																						
Head TSL temperature change during test	<1.0 °C	—	—																																																																																																																						
	Condition																																																																																																																								
SAR measured	100 mW input power	7.71 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)																																																																																																																							
	Condition																																																																																																																								
SAR measured	100 mW input power	2.16 W/kg																																																																																																																							
SAR for nominal Head TSL parameters	normalized to 1W	21.5 W/kg ± 24.2 % (k=2)																																																																																																																							
<p>Certificate No: Z22-60187</p> <p>Page 3 of 10</p>	<p>Certificate No: Z22-60187</p> <p>Page 4 of 10</p>																																																																																																																								




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DocCheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)5735888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)5735888 f(86-512)57370818 sgs.china@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



2 DAE4 - SN 1245

<div style="border: 1px solid black; padding: 10px;"> <p>IMPORTANT NOTICE</p> <p>USAGE OF THE DAE4</p> <p>The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:</p> <p>Battery Exchange: The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.</p> <p>Shipping of the DAE: Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an airtight bag. The airtight bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.</p> <p>E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulation on the E-stop. Therefore, the E-stop shall always mount the probe to the DAE carefully and never touch the DAE unit in a non-dust-free environment. It is not recommended to use the E-stop for safety applications.</p> <p>Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.</p> <p>DASY Configuration Files: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.</p> <p>Important Note: Warranty and calibration is valid if the DAE unit is disassembled partly or fully by the Customer.</p> <p>Important Note: Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.</p> <p>Important Note: To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.</p> <p>TN_EH190306AE DAE4.docx 07.03.2019</p> </div>	<div style="border: 1px solid black; padding: 10px;"> <p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.ch, info@speag.ch</p> <p>Accredited by: The Swiss Accreditation Service (SCS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No.: DAE4-1245_May22</p> <p>CALIBRATION CERTIFICATE</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">Object</td> <td colspan="3">DAE4 - SD 000 D04 BM - SN: 1245</td> </tr> <tr> <td>Calibration procedure(s)</td> <td colspan="3">QA CAL-06.v30 Calibration procedure for the data acquisition electronics (DAE)</td> </tr> <tr> <td>Calibration date:</td> <td colspan="3">May 30, 2022</td> </tr> <tr> <td colspan="4"> This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%. </td> </tr> <tr> <td>Calibration Equipment used (M&TE criteria for calibration)</td> <td colspan="3"> Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration Keithley Multimeter Type 2301 SN: 9810278 31-Aug-21 (No.31968) Aug-22 Secondary Standards ID # Check Date (in house) Scheduled Check Auto DAE Calibration Unit SE UWB 050 AA 1001 24-Jan-22 (in house check) In house check: Jan-23 Calibrator Box V2.1 SE UWB 006 AA 1002 24-Jan-22 (in house check) In house check: Jan-23 </td> </tr> <tr> <td>Calibrated by:</td> <td>Name: Dominique Steffen</td> <td>Function: Laboratory Technician</td> <td>Signature: </td> </tr> <tr> <td>Approved by:</td> <td>Sven Kohn</td> <td>Technical Manager</td> <td></td> </tr> <tr> <td colspan="4">Issued: May 30, 2022</td> </tr> </table> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: DAE4-1245_May22 Page 1 of 5</p> </div>	Object	DAE4 - SD 000 D04 BM - SN: 1245			Calibration procedure(s)	QA CAL-06.v30 Calibration procedure for the data acquisition electronics (DAE)			Calibration date:	May 30, 2022			This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.				Calibration Equipment used (M&TE criteria for calibration)	Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration Keithley Multimeter Type 2301 SN: 9810278 31-Aug-21 (No.31968) Aug-22 Secondary Standards ID # Check Date (in house) Scheduled Check Auto DAE Calibration Unit SE UWB 050 AA 1001 24-Jan-22 (in house check) In house check: Jan-23 Calibrator Box V2.1 SE UWB 006 AA 1002 24-Jan-22 (in house check) In house check: Jan-23			Calibrated by:	Name: Dominique Steffen	Function: Laboratory Technician	Signature: 	Approved by:	Sven Kohn	Technical Manager		Issued: May 30, 2022			
Object	DAE4 - SD 000 D04 BM - SN: 1245																																
Calibration procedure(s)	QA CAL-06.v30 Calibration procedure for the data acquisition electronics (DAE)																																
Calibration date:	May 30, 2022																																
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.																																	
Calibration Equipment used (M&TE criteria for calibration)	Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration Keithley Multimeter Type 2301 SN: 9810278 31-Aug-21 (No.31968) Aug-22 Secondary Standards ID # Check Date (in house) Scheduled Check Auto DAE Calibration Unit SE UWB 050 AA 1001 24-Jan-22 (in house check) In house check: Jan-23 Calibrator Box V2.1 SE UWB 006 AA 1002 24-Jan-22 (in house check) In house check: Jan-23																																
Calibrated by:	Name: Dominique Steffen	Function: Laboratory Technician	Signature: 																														
Approved by:	Sven Kohn	Technical Manager																															
Issued: May 30, 2022																																	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: to check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com


3 EX3DV4 - SN 7767

Calibration Laboratory of Schmid & Partner Engineering AG Zueggstrasse 43, 8004 Zurich, Switzerland		  S Schweizerischer Kalibrierdienst S Service suisse d'etonnage C Servizio svizzero di taratura S Servizio calibrazione Service																																																
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Mutual Recognition Agreement on the recognition of calibration certificates		  S Schweizerischer Kalibrierdienst S Service suisse d'etonnage C Servizio svizzero di taratura S Servizio calibrazione Service																																																
Accredited No.: SCS 0108		Accredited No.: SC 0108																																																
Client: SGS-CN (Auden)		Certificate No: EX-7787_Oct22																																																
CALIBRATION CERTIFICATE																																																		
Object: EX3DV4 - SN:7767																																																		
Calibration procedure(s): QA CAL-01.v9, QA CAL-12.v9, QA CAL-14.v6, QA CAL-23.v5, QA CAL-25.v7 Calibration procedure for dosimetric E-field probes																																																		
Calibration date: October 28, 2022																																																		
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 0.5°C) and humidity < 70%.																																																		
Calibration Equipment used (MATE critical for calibration): <table border="1"> <thead> <tr> <th>Power meter M9P</th> <th>ID: SN:104779</th> <th>Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03584)</th> <th>Scheduled Calibration: Apr-23</th> </tr> </thead> <tbody> <tr> <td>Power meter M9P-ZH</td> <td>ID: SN:104780</td> <td>Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03585)</td> <td>Scheduled Calibration: Apr-23</td> </tr> <tr> <td>DOOP DAK-3.5 (used)</td> <td>ID: SN:12468</td> <td>Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK3.5-12468_Oct22)</td> <td>Scheduled Calibration: Oct-23</td> </tr> <tr> <td>DOOP DAK-12</td> <td>ID: SN:1018</td> <td>Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK12-1018_Oct22)</td> <td>Scheduled Calibration: Oct-23</td> </tr> <tr> <td>DOPEK B2B (used)</td> <td>ID: SN:66052</td> <td>Cal Date (Certificate No.): 19-Oct-22 (DOPEK-B2B-66052_Oct22)</td> <td>Scheduled Calibration: Oct-23</td> </tr> <tr> <td>DOPEK</td> <td>ID: SN:66053</td> <td>Cal Date (Certificate No.): 19-Oct-22 (No. DOPEK-66053_Oct22)</td> <td>Scheduled Calibration: Oct-23</td> </tr> <tr> <td>Reference Probe E3303N</td> <td>ID: SN:3013</td> <td>Cal Date (Certificate No.): 27-Oct-21 (No. E33-3013_Oct21)</td> <td>Scheduled Calibration: Dec-22</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Boundary Sensors</th> <th>ID: SN:104793814</th> <th>Check Date (In house): 06-Apr-18 (In house check Jun-22)</th> <th>Scheduled Check: 01-May-23 (In house check Jun-24)</th> </tr> </thead> <tbody> <tr> <td>Power sensor E5413A</td> <td>ID: SN:1047940607</td> <td>Check Date (In house): 06-Apr-18 (In house check Jun-22)</td> <td>Scheduled Check: 01-May-23 (In house check Jun-24)</td> </tr> <tr> <td>Power sensor E5414A</td> <td>ID: SN:00010210</td> <td>Check Date (In house): 06-Apr-18 (In house check Jun-22)</td> <td>Scheduled Check: 01-May-23 (In house check Jun-24)</td> </tr> <tr> <td>Network Analyzer E8356A</td> <td>ID: SN:10479379</td> <td>Check Date (In house): 31-Mar-14 (In house check Oct-22)</td> <td>Scheduled Check: 01-May-23 (In house check Jun-24)</td> </tr> <tr> <td>Network Analyzer E8356A</td> <td>ID: SN:10480477</td> <td>Check Date (In house): 31-Mar-14 (In house check Oct-22)</td> <td>Scheduled Check: 01-May-23 (In house check Jun-24)</td> </tr> </tbody> </table>			Power meter M9P	ID: SN:104779	Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03584)	Scheduled Calibration: Apr-23	Power meter M9P-ZH	ID: SN:104780	Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03585)	Scheduled Calibration: Apr-23	DOOP DAK-3.5 (used)	ID: SN:12468	Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK3.5-12468_Oct22)	Scheduled Calibration: Oct-23	DOOP DAK-12	ID: SN:1018	Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK12-1018_Oct22)	Scheduled Calibration: Oct-23	DOPEK B2B (used)	ID: SN:66052	Cal Date (Certificate No.): 19-Oct-22 (DOPEK-B2B-66052_Oct22)	Scheduled Calibration: Oct-23	DOPEK	ID: SN:66053	Cal Date (Certificate No.): 19-Oct-22 (No. DOPEK-66053_Oct22)	Scheduled Calibration: Oct-23	Reference Probe E3303N	ID: SN:3013	Cal Date (Certificate No.): 27-Oct-21 (No. E33-3013_Oct21)	Scheduled Calibration: Dec-22	Boundary Sensors	ID: SN:104793814	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)	Power sensor E5413A	ID: SN:1047940607	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)	Power sensor E5414A	ID: SN:00010210	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)	Network Analyzer E8356A	ID: SN:10479379	Check Date (In house): 31-Mar-14 (In house check Oct-22)	Scheduled Check: 01-May-23 (In house check Jun-24)	Network Analyzer E8356A	ID: SN:10480477	Check Date (In house): 31-Mar-14 (In house check Oct-22)	Scheduled Check: 01-May-23 (In house check Jun-24)
Power meter M9P	ID: SN:104779	Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03584)	Scheduled Calibration: Apr-23																																															
Power meter M9P-ZH	ID: SN:104780	Cal Date (Certificate No.): 04-Apr-22 (No. 217-0358/03585)	Scheduled Calibration: Apr-23																																															
DOOP DAK-3.5 (used)	ID: SN:12468	Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK3.5-12468_Oct22)	Scheduled Calibration: Oct-23																																															
DOOP DAK-12	ID: SN:1018	Cal Date (Certificate No.): 26-Oct-22 (DOOP-DAK12-1018_Oct22)	Scheduled Calibration: Oct-23																																															
DOPEK B2B (used)	ID: SN:66052	Cal Date (Certificate No.): 19-Oct-22 (DOPEK-B2B-66052_Oct22)	Scheduled Calibration: Oct-23																																															
DOPEK	ID: SN:66053	Cal Date (Certificate No.): 19-Oct-22 (No. DOPEK-66053_Oct22)	Scheduled Calibration: Oct-23																																															
Reference Probe E3303N	ID: SN:3013	Cal Date (Certificate No.): 27-Oct-21 (No. E33-3013_Oct21)	Scheduled Calibration: Dec-22																																															
Boundary Sensors	ID: SN:104793814	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)																																															
Power sensor E5413A	ID: SN:1047940607	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)																																															
Power sensor E5414A	ID: SN:00010210	Check Date (In house): 06-Apr-18 (In house check Jun-22)	Scheduled Check: 01-May-23 (In house check Jun-24)																																															
Network Analyzer E8356A	ID: SN:10479379	Check Date (In house): 31-Mar-14 (In house check Oct-22)	Scheduled Check: 01-May-23 (In house check Jun-24)																																															
Network Analyzer E8356A	ID: SN:10480477	Check Date (In house): 31-Mar-14 (In house check Oct-22)	Scheduled Check: 01-May-23 (In house check Jun-24)																																															
Name: Aliona Georgiadou Function: Laboratory Technician Signature: 																																																		
Approved by: Sven Kohn Technical Manager Signature: 																																																		
Issued: October 31, 2022																																																		
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																																																		
Certificate No: EX-7787_Oct22		Certificate No: EX-7787_Oct22																																																
Page 1 of 9		Page 2 of 8																																																

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>, and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-For-Document.aspx>. It is the responsibility of the Client to read and understand these documents. Any type of intervention by the Company is advised that the information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



<p>EX3DV4 - SN:7767</p> <p>October 29, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Basic Calibration Parameters</p> <table border="1"> <tr> <th></th> <th>Sensor X</th> <th>Sensor Y</th> <th>Sensor Z</th> <th>Unc (k = 2)</th> </tr> <tr> <td>Norm (A/m²)^a</td> <td>0.67</td> <td>0.69</td> <td>0.66</td> <td>$\pm 10.1\%$</td> </tr> <tr> <td>DCP (mV)^b</td> <td>103.4</td> <td>107.3</td> <td>105.7</td> <td>$\pm 4.7\%$</td> </tr> </table> <p>Calibration Results for Modulation Response</p> <table border="1"> <thead> <tr> <th>UID / Communication System Name</th> <th>A dB</th> <th>B dB/mV</th> <th>C</th> <th>D dB</th> <th>VR mV</th> <th>Max dev.</th> <th>Max Unc^c k = 2</th> </tr> </thead> <tbody> <tr> <td>0 CW</td> <td>X: 0.00</td> <td>0.00</td> <td>1.00</td> <td>0.00</td> <td>164.7</td> <td>$\pm 3.9\%$</td> <td>$\pm 4.7\%$</td> </tr> <tr> <td></td> <td>Y: 0.00</td> <td>0.00</td> <td>1.00</td> <td>0.00</td> <td>166.7</td> <td></td> <td></td> </tr> <tr> <td></td> <td>Z: 0.00</td> <td>0.00</td> <td>1.00</td> <td>0.00</td> <td>176.3</td> <td></td> <td></td> </tr> </tbody> </table> <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>^a The uncertainty of Norm is 3.2% as per the 95% total uncertainty per Table 3B, (see Page 5). ^b Uncertainties per sensor uncertainty for maximum specified field strength. ^c Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.</p> <p>Certificate No: EX-7767_Oct22 Page 5 of 9</p>		Sensor X	Sensor Y	Sensor Z	Unc (k = 2)	Norm (A/m ²) ^a	0.67	0.69	0.66	$\pm 10.1\%$	DCP (mV) ^b	103.4	107.3	105.7	$\pm 4.7\%$	UID / Communication System Name	A dB	B dB/mV	C	D dB	VR mV	Max dev.	Max Unc ^c k = 2	0 CW	X: 0.00	0.00	1.00	0.00	164.7	$\pm 3.9\%$	$\pm 4.7\%$		Y: 0.00	0.00	1.00	0.00	166.7				Z: 0.00	0.00	1.00	0.00	176.3			<p>EX3DV4 - SN:7767</p> <p>October 29, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Other Probe Parameters</p> <table border="1"> <tr> <td>Sensor Arrangement</td> <td>Triangular</td> </tr> <tr> <td>Orientation</td> <td>144.8°</td> </tr> <tr> <td>Mechanical Surface Detection Mode</td> <td>enabled</td> </tr> <tr> <td>Optical Surface Detection Mode</td> <td>disabled</td> </tr> <tr> <td>Probe Overall Length</td> <td>337 mm</td> </tr> <tr> <td>Probe Body Diameter</td> <td>10 mm</td> </tr> <tr> <td>Tip Length</td> <td>9 mm</td> </tr> <tr> <td>Tip Diameter</td> <td>2.5 mm</td> </tr> <tr> <td>Probe Tip to Sensor X Calibration Point</td> <td>1 mm</td> </tr> <tr> <td>Probe Tip to Sensor Y Calibration Point</td> <td>1 mm</td> </tr> <tr> <td>Probe Tip to Sensor Z Calibration Point</td> <td>1 mm</td> </tr> <tr> <td>Recommended Measurement Distance from Surface</td> <td>1.4 mm</td> </tr> </table> <p>Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.</p> <p>Certificate No: EX-7767_Oct22 Page 4 of 9</p>	Sensor Arrangement	Triangular	Orientation	144.8°	Mechanical Surface Detection Mode	enabled	Optical Surface Detection Mode	disabled	Probe Overall Length	337 mm	Probe Body Diameter	10 mm	Tip Length	9 mm	Tip Diameter	2.5 mm	Probe Tip to Sensor X Calibration Point	1 mm	Probe Tip to Sensor Y Calibration Point	1 mm	Probe Tip to Sensor Z Calibration Point	1 mm	Recommended Measurement Distance from Surface	1.4 mm																																																																																																																																																																		
	Sensor X	Sensor Y	Sensor Z	Unc (k = 2)																																																																																																																																																																																																																																						
Norm (A/m ²) ^a	0.67	0.69	0.66	$\pm 10.1\%$																																																																																																																																																																																																																																						
DCP (mV) ^b	103.4	107.3	105.7	$\pm 4.7\%$																																																																																																																																																																																																																																						
UID / Communication System Name	A dB	B dB/mV	C	D dB	VR mV	Max dev.	Max Unc ^c k = 2																																																																																																																																																																																																																																			
0 CW	X: 0.00	0.00	1.00	0.00	164.7	$\pm 3.9\%$	$\pm 4.7\%$																																																																																																																																																																																																																																			
	Y: 0.00	0.00	1.00	0.00	166.7																																																																																																																																																																																																																																					
	Z: 0.00	0.00	1.00	0.00	176.3																																																																																																																																																																																																																																					
Sensor Arrangement	Triangular																																																																																																																																																																																																																																									
Orientation	144.8°																																																																																																																																																																																																																																									
Mechanical Surface Detection Mode	enabled																																																																																																																																																																																																																																									
Optical Surface Detection Mode	disabled																																																																																																																																																																																																																																									
Probe Overall Length	337 mm																																																																																																																																																																																																																																									
Probe Body Diameter	10 mm																																																																																																																																																																																																																																									
Tip Length	9 mm																																																																																																																																																																																																																																									
Tip Diameter	2.5 mm																																																																																																																																																																																																																																									
Probe Tip to Sensor X Calibration Point	1 mm																																																																																																																																																																																																																																									
Probe Tip to Sensor Y Calibration Point	1 mm																																																																																																																																																																																																																																									
Probe Tip to Sensor Z Calibration Point	1 mm																																																																																																																																																																																																																																									
Recommended Measurement Distance from Surface	1.4 mm																																																																																																																																																																																																																																									
<p>EX3DV4 - SN:7767</p> <p>October 28, 2022</p> <p>Parameters of Probe: EX3DV4 - SN:7767</p> <p>Calibration Parameter Determined in Head Tissue Simulating Media</p> <table border="1"> <thead> <tr> <th>T (GHz)^d</th> <th>Relative Permittivity^e</th> <th>Conductivity^f (S/m)</th> <th>Conv^g X</th> <th>Conv^g Y</th> <th>Conv^g Z</th> <th>Alpha^h</th> <th>Degⁱ (mm)</th> <th>Unc (k = 2)</th> </tr> </thead> <tbody> <tr><td>150</td><td>52.3</td><td>0.76</td><td>14.06</td><td>14.06</td><td>0.00</td><td>1.00</td><td>$\pm 13.3\%$</td><td></td></tr> <tr><td>450</td><td>43.5</td><td>0.87</td><td>11.90</td><td>11.90</td><td>0.16</td><td>1.30</td><td>$\pm 13.3\%$</td><td></td></tr> <tr><td>750</td><td>41.9</td><td>0.89</td><td>10.26</td><td>10.26</td><td>0.65</td><td>0.98</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>835</td><td>41.5</td><td>0.90</td><td>10.00</td><td>10.00</td><td>0.45</td><td>0.89</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>1750</td><td>40.1</td><td>1.37</td><td>9.32</td><td>9.32</td><td>0.36</td><td>0.86</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>1900</td><td>40.0</td><td>1.40</td><td>8.91</td><td>8.91</td><td>0.33</td><td>0.86</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>2100</td><td>39.8</td><td>1.49</td><td>8.60</td><td>8.60</td><td>0.30</td><td>0.86</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>2300</td><td>39.5</td><td>1.57</td><td>8.44</td><td>8.44</td><td>0.33</td><td>0.90</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>2450</td><td>39.2</td><td>1.80</td><td>8.24</td><td>8.24</td><td>0.24</td><td>0.92</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>2600</td><td>39.0</td><td>1.96</td><td>7.99</td><td>7.99</td><td>0.27</td><td>0.90</td><td>$\pm 12.0\%$</td><td></td></tr> <tr><td>3300</td><td>38.2</td><td>2.71</td><td>7.55</td><td>7.55</td><td>0.30</td><td>1.25</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>3500</td><td>37.9</td><td>2.91</td><td>7.45</td><td>7.45</td><td>0.30</td><td>1.25</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>3700</td><td>37.7</td><td>3.12</td><td>7.20</td><td>7.20</td><td>0.30</td><td>1.25</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>3900</td><td>37.6</td><td>3.32</td><td>6.84</td><td>6.84</td><td>0.40</td><td>1.60</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4100</td><td>37.2</td><td>3.53</td><td>6.53</td><td>6.53</td><td>0.40</td><td>1.60</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4200</td><td>37.1</td><td>3.63</td><td>6.30</td><td>6.30</td><td>0.40</td><td>1.70</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4400</td><td>36.9</td><td>3.84</td><td>6.17</td><td>6.17</td><td>0.40</td><td>1.70</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4600</td><td>36.7</td><td>4.04</td><td>6.15</td><td>6.15</td><td>0.40</td><td>1.70</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4800</td><td>36.4</td><td>4.26</td><td>6.13</td><td>6.13</td><td>0.40</td><td>1.90</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>4900</td><td>36.3</td><td>4.40</td><td>6.07</td><td>6.07</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>5000</td><td>36.0</td><td>4.96</td><td>5.65</td><td>5.65</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>5300</td><td>35.9</td><td>4.76</td><td>5.48</td><td>5.48</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>5500</td><td>35.6</td><td>4.96</td><td>5.30</td><td>5.30</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>5600</td><td>35.5</td><td>5.07</td><td>5.14</td><td>5.14</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> <tr><td>5800</td><td>35.3</td><td>5.27</td><td>5.10</td><td>5.10</td><td>0.40</td><td>1.80</td><td>$\pm 13.1\%$</td><td></td></tr> </tbody> </table> <p>^d Frequency validity above 300 MHz at ±10% by weight for EX3DV4.4 and higher class. Page 5, also it is mentioned at 450 MHz. The uncertainty is the RSS of the Conv^g uncertainty of calibration frequency and the uncertainty for the indicated frequency range. Frequency validity below 300 MHz is ±10.2%. ^e At 450 MHz for Conv^g assessments at 150, 450, 750, 835, 1750, 1900, 2100, 2300, 2450, 2600, 3300, 3500, 3700, 3900, 4100, 4200, 4400, 4600, 4800, 4900, 5000, 5300, 5500, 5600, 5800, 6000, 6300, 6500, 6700, 7000, 7500, 7800, 8000, 8300, 8500, 8800, 9000, 9300, 9500, 9800, 10000, 10300, 10500, 10800, 11000, 11300, 11500, 11800, 12000, 12300, 12500, 12800, 13000, 13300, 13500, 13800, 14000, 14300, 14500, 14800, 15000, 15300, 15500, 15800, 16000, 16300, 16500, 16800, 17000, 17300, 17500, 17800, 18000, 18300, 18500, 18800, 19000, 19300, 19500, 19800, 20000, 20300, 20500, 20800, 21000, 21300, 21500, 21800, 22000, 22300, 22500, 22800, 23000, 23300, 23500, 23800, 24000, 24300, 24500, 24800, 25000, 25300, 25500, 25800, 26000, 26300, 26500, 26800, 27000, 27300, 27500, 27800, 28000, 28300, 28500, 28800, 29000, 29300, 29500, 29800, 30000, 30300, 30500, 30800, 31000, 31300, 31500, 31800, 32000, 32300, 32500, 32800, 33000, 33300, 33500, 33800, 34000, 34300, 34500, 34800, 35000, 35300, 35500, 35800, 36000, 36300, 36500, 36800, 37000, 37300, 37500, 37800, 38000, 38300, 38500, 38800, 39000, 39300, 39500, 39800, 40000, 40300, 40500, 40800, 41000, 41300, 41500, 41800, 42000, 42300, 42500, 42800, 43000, 43300, 43500, 43800, 44000, 44300, 44500, 44800, 45000, 45300, 45500, 45800, 46000, 46300, 46500, 46800, 47000, 47300, 47500, 47800, 48000, 48300, 48500, 48800, 49000, 49300, 49500, 49800, 50000, 50300, 50500, 50800, 51000, 51300, 51500, 51800, 52000, 52300, 52500, 52800, 53000, 53300, 53500, 53800, 54000, 54300, 54500, 54800, 55000, 55300, 55500, 55800, 56000, 56300, 56500, 56800, 57000, 57300, 57500, 57800, 58000, 58300, 58500, 58800, 59000, 59300, 59500, 59800, 60000, 60300, 60500, 60800, 61000, 61300, 61500, 61800, 62000, 62300, 62500, 62800, 63000, 63300, 63500, 63800, 64000, 64300, 64500, 64800, 65000, 65300, 65500, 65800, 66000, 66300, 66500, 66800, 67000, 67300, 67500, 67800, 68000, 68300, 68500, 68800, 69000, 69300, 69500, 69800, 70000, 70300, 70500, 70800, 71000, 71300, 71500, 71800, 72000, 72300, 72500, 72800, 73000, 73300, 73500, 73800, 74000, 74300, 74500, 74800, 75000, 75300, 75500, 75800, 76000, 76300, 76500, 76800, 77000, 77300, 77500, 77800, 78000, 78300, 78500, 78800, 79000, 79300, 79500, 79800, 80000, 80300, 80500, 80800, 81000, 81300, 81500, 81800, 82000, 82300, 82500, 82800, 83000, 83300, 83500, 83800, 84000, 84300, 84500, 84800, 85000, 85300, 85500, 85800, 86000, 86300, 86500, 86800, 87000, 87300, 87500, 87800, 88000, 88300, 88500, 88800, 89000, 89300, 89500, 89800, 90000, 90300, 90500, 90800, 91000, 91300, 91500, 91800, 92000, 92300, 92500, 92800, 93000, 93300, 93500, 93800, 94000, 94300, 94500, 94800, 95000, 95300, 95500, 95800, 96000, 96300, 96500, 96800, 97000, 97300, 97500, 97800, 98000, 98300, 98500, 98800, 99000, 99300, 99500, 99800, 100000, 100300, 100500, 100800, 101000, 101300, 101500, 101800, 102000, 102300, 102500, 102800, 103000, 103300, 103500, 103800, 104000, 104300, 104500, 104800, 105000, 105300, 105500, 105800, 106000, 106300, 106500, 106800, 107000, 107300, 107500, 107800, 108000, 108300, 108500, 108800, 109000, 109300, 109500, 109800, 110000, 110300, 110500, 110800, 111000, 111300, 111500, 111800, 112000, 112300, 112500, 112800, 113000, 113300, 113500, 113800, 114000, 114300, 114500, 114800, 115000, 115300, 115500, 115800, 116000, 116300, 116500, 116800, 117000, 117300, 117500, 117800, 118000, 118300, 118500, 118800, 119000, 119300, 119500, 119800, 120000, 120300, 120500, 120800, 121000, 121300, 121500, 121800, 122000, 122300, 122500, 122800, 123000, 123300, 123500, 123800, 124000, 124300, 124500, 124800, 125000, 125300, 125500, 125800, 126000, 126300, 126500, 126800, 127000, 127300, 127500, 127800, 128000, 128300, 128500, 128800, 129000, 129300, 129500, 129800, 130000, 130300, 130500, 130800, 131000, 131300, 131500, 131800, 132000, 132300, 132500, 132800, 133000, 133300, 133500, 133800, 134000, 134300, 134500, 134800, 135000, 135300, 135500, 135800, 136000, 136300, 136500, 136800, 137000, 137300, 137500, 137800, 138000, 138300, 138500, 138800, 139000, 139300, 139500, 139800, 140000, 140300, 140500, 140800, 141000, 141300, 141500, 141800, 142000, 142300, 142500, 142800, 143000, 143300, 143500, 143800, 144000, 144300, 144500, 144800, 145000, 145300, 145500, 145800, 146000, 146300, 146500, 146800, 147000, 147300, 147500, 147800, 148000, 148300, 148500, 148800, 149000, 149300, 149500, 149800, 150000, 150300, 150500, 150800, 151000, 151300, 151500, 151800, 152000, 152300, 152500, 152800, 153000, 153300, 153500, 153800, 154000, 154300, 154500, 154800, 155000, 155300, 155500, 155800, 156000, 156300, 156500, 156800, 157000, 157300, 157500, 157800, 158000, 158300, 158500, 158800, 159000, 159300, 159500, 159800, 160000, 160300, 160500, 160800, 161000, 161300, 161500, 161800, 162000, 162300, 162500, 162800, 163000, 163300, 163500, 163800, 164000, 164300, 164500, 164800, 165000, 165300, 165500, 165800, 166000, 166300, 166500, 166800, 167000, 167300, 167500, 167800, 168000, 168300, 168500, 168800, 169000, 169300, 169500, 169800, 170000, 170300, 170500, 170800, 171000, 171300, 171500, 171800, 172000, 172300, 172500, 172800, 173000, 173300, 173500, 173800, 174000, 174300, 174500, 174800, 175000, 175300, 175500, 175800, 176000, 176300, 176500, 176800, 177000, 177300, 177500, 177800, 178000, 178300, 178500, 178800, 179000, 179300, 179500, 179800, 180000, 180300, 180500, 180800, 181000, 181300, 181500, 181800, 182000, 182300, 182500, 182800, 183000, 183300, 183500, 183800, 184000, 184300, 184500, 184800, 185000, 185300, 185500, 185800, 186000, 186300, 186500, 186800, 187000, 187300, 187500, 187800, 188000, 188300, 188500, 188800, 189000, 189300, 189500, 189800, 190000, 190300, 190500, 190800, 191000, 191300, 191500, 191800, 192000, 192300, 192500, 192800, 193000, 193300, 193500, 193800, 194000, 194300, 194500, 194800, 195000, 195300, 195500, 195800, 196000, 196300, 196500, 196800, 197000, 197300, 197500, 197800, 198000, 198300, 198500, 198800, 199000, 199300, 199500, 199800, 200000, 200300, 200500, 200800, 201000, 201300, 201500, 201800, 202000, 202300, 202500, 202800, 203000, 203300, 203500, 203800, 204000, 204300, 204500, 204800, 205000, 205300, 205500, 205800, 206000, 206300, 206500, 206800, 207000, 207300, 207500, 207800, 208000, 208300, 208500, 208800, 209000, 209300, 209500, 209800, 210000, 210300, 210500, 210800, 211000, 211300, 211500, 211800, 212000, 212300, 212500, 212800, 213000, 213300, 213500, 213800, 214000, 214300, 214500, 214800, 215000, 215300, 215500, 215800, 216000, 216300, 216500, 216800, 217000, 217300, 217500, 217800, 218000, 218300, 218500, 218800, 219000, 219300, 219500, 219800, 220000, 220300, 220500, 220800, 221000, 221300, 221500, 221800, 222000, 222300, 222500, 222800, 223000, 223300, 223500, 223800, 224000, 224300, 224500, 224800, 225000, 225300, 225500, 225800, 226000, 226300, 226500, 226800, 227000, 227300, 227500, 227800, 228000, 228300, 228500, 228800, 229000, 229300, 229500, 229800, 230000, 230300, 230500, 230800, 231000, 231300, 231500, 231800, 232000, 232300, 232500, 232800, 233000, 233300, 233500, 233800, 234000, 234300, 234500, 234800, 235000, 235300, 235500, 235800, 236000, 236300, 236500, 236800, 237000, 237300, 237500, 237800, 238000, 238300, 238500, 238800, 239000, 239300, 239500, 239800, 240000, 240300, 240500, 240800, 241000, 241300, 241500, 241800, 242000, 242300, 242500, 242800, 243000, 243300, 243500, 243800, 2</p>	T (GHz) ^d	Relative Permittivity ^e	Conductivity ^f (S/m)	Conv ^g X	Conv ^g Y	Conv ^g Z	Alpha ^h	Deg ⁱ (mm)	Unc (k = 2)	150	52.3	0.76	14.06	14.06	0.00	1.00	$\pm 13.3\%$		450	43.5	0.87	11.90	11.90	0.16	1.30	$\pm 13.3\%$		750	41.9	0.89	10.26	10.26	0.65	0.98	$\pm 12.0\%$		835	41.5	0.90	10.00	10.00	0.45	0.89	$\pm 12.0\%$		1750	40.1	1.37	9.32	9.32	0.36	0.86	$\pm 12.0\%$		1900	40.0	1.40	8.91	8.91	0.33	0.86	$\pm 12.0\%$		2100	39.8	1.49	8.60	8.60	0.30	0.86	$\pm 12.0\%$		2300	39.5	1.57	8.44	8.44	0.33	0.90	$\pm 12.0\%$		2450	39.2	1.80	8.24	8.24	0.24	0.92	$\pm 12.0\%$		2600	39.0	1.96	7.99	7.99	0.27	0.90	$\pm 12.0\%$		3300	38.2	2.71	7.55	7.55	0.30	1.25	$\pm 13.1\%$		3500	37.9	2.91	7.45	7.45	0.30	1.25	$\pm 13.1\%$		3700	37.7	3.12	7.20	7.20	0.30	1.25	$\pm 13.1\%$		3900	37.6	3.32	6.84	6.84	0.40	1.60	$\pm 13.1\%$		4100	37.2	3.53	6.53	6.53	0.40	1.60	$\pm 13.1\%$		4200	37.1	3.63	6.30	6.30	0.40	1.70	$\pm 13.1\%$		4400	36.9	3.84	6.17	6.17	0.40	1.70	$\pm 13.1\%$		4600	36.7	4.04	6.15	6.15	0.40	1.70	$\pm 13.1\%$		4800	36.4	4.26	6.13	6.13	0.40	1.90	$\pm 13.1\%$		4900	36.3	4.40	6.07	6.07	0.40	1.80	$\pm 13.1\%$		5000	36.0	4.96	5.65	5.65	0.40	1.80	$\pm 13.1\%$		5300	35.9	4.76	5.48	5.48	0.40	1.80	$\pm 13.1\%$		5500	35.6	4.96	5.30	5.30	0.40	1.80	$\pm 13.1\%$		5600	35.5	5.07	5.14	5.14	0.40	1.80	$\pm 13.1\%$		5800	35.3	5.27	5.10	5.10	0.40	1.80	$\pm 13.1\%$	
T (GHz) ^d	Relative Permittivity ^e	Conductivity ^f (S/m)	Conv ^g X	Conv ^g Y	Conv ^g Z	Alpha ^h	Deg ⁱ (mm)	Unc (k = 2)																																																																																																																																																																																																																																		
150	52.3	0.76	14.06	14.06	0.00	1.00	$\pm 13.3\%$																																																																																																																																																																																																																																			
450	43.5	0.87	11.90	11.90	0.16	1.30	$\pm 13.3\%$																																																																																																																																																																																																																																			
750	41.9	0.89	10.26	10.26	0.65	0.98	$\pm 12.0\%$																																																																																																																																																																																																																																			
835	41.5	0.90	10.00	10.00	0.45	0.89	$\pm 12.0\%$																																																																																																																																																																																																																																			
1750	40.1	1.37	9.32	9.32	0.36	0.86	$\pm 12.0\%$																																																																																																																																																																																																																																			
1900	40.0	1.40	8.91	8.91	0.33	0.86	$\pm 12.0\%$																																																																																																																																																																																																																																			
2100	39.8	1.49	8.60	8.60	0.30	0.86	$\pm 12.0\%$																																																																																																																																																																																																																																			
2300	39.5	1.57	8.44	8.44	0.33	0.90	$\pm 12.0\%$																																																																																																																																																																																																																																			
2450	39.2	1.80	8.24	8.24	0.24	0.92	$\pm 12.0\%$																																																																																																																																																																																																																																			
2600	39.0	1.96	7.99	7.99	0.27	0.90	$\pm 12.0\%$																																																																																																																																																																																																																																			
3300	38.2	2.71	7.55	7.55	0.30	1.25	$\pm 13.1\%$																																																																																																																																																																																																																																			
3500	37.9	2.91	7.45	7.45	0.30	1.25	$\pm 13.1\%$																																																																																																																																																																																																																																			
3700	37.7	3.12	7.20	7.20	0.30	1.25	$\pm 13.1\%$																																																																																																																																																																																																																																			
3900	37.6	3.32	6.84	6.84	0.40	1.60	$\pm 13.1\%$																																																																																																																																																																																																																																			
4100	37.2	3.53	6.53	6.53	0.40	1.60	$\pm 13.1\%$																																																																																																																																																																																																																																			
4200	37.1	3.63	6.30	6.30	0.40	1.70	$\pm 13.1\%$																																																																																																																																																																																																																																			
4400	36.9	3.84	6.17	6.17	0.40	1.70	$\pm 13.1\%$																																																																																																																																																																																																																																			
4600	36.7	4.04	6.15	6.15	0.40	1.70	$\pm 13.1\%$																																																																																																																																																																																																																																			
4800	36.4	4.26	6.13	6.13	0.40	1.90	$\pm 13.1\%$																																																																																																																																																																																																																																			
4900	36.3	4.40	6.07	6.07	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			
5000	36.0	4.96	5.65	5.65	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			
5300	35.9	4.76	5.48	5.48	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			
5500	35.6	4.96	5.30	5.30	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			
5600	35.5	5.07	5.14	5.14	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			
5800	35.3	5.27	5.10	5.10	0.40	1.80	$\pm 13.1\%$																																																																																																																																																																																																																																			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyi Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	ΔΩ
2021/4/26	-31.4	/	47.8	/
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	ΔΩ
2021/4/21	-23	/	57.1	/

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention is drawn to the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com



No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com