

In Accordance With:

Nemko USA, Inc.

11696 Sorrento Valley Rd., Suite F San Diego, CA 92121-1024 Phone (858) 755-5525 Fax (858) 452-1810

CERTIFICATION TEST REPORT

FCC Part 95 Subpart I

m / toos danes vviii.	r do r art do daspart r
Applicant:	Spinal Modulation Inc.
	1135 O'Brien Dr.
	Menlo Park, CA 94025

Equipment Under Test (EUT): Trial Neurostimulator

Model: MN0100

FCC ID: Y8L-MN0100

Tested By: Nemko USA Inc.

11696 Sorrento Valley Road, Suite F

San Diego, CA 92121

PREPARED ON January 17, 2011
REPORT NUMBER: 2011 01160678 FCC3

PROJECT NUMBER: 53259 NEx Number: 160678

Total Number of Pages: 19

Report Number: 2011 01160678 FCC3

www.nemko.com

FCC ID: Y8L-MN0100

1. Section1: Summary of Test Results

1.1 General

All measurements are traceable to national standards

These tests were conducted on a sample of the equipment for the purpose of demonstrating compliance with FCC Part 95 Subpart I. Radiated tests were conducted is accordance with ANSI C63.4-2003. Radiated emissions are made on an open area test site. A description of the test facility is on file with the FCC.

The assessment summary is as follows:

Apparatus Assessed: Trial Neurostimulator

Model: MN0100

Serial: See Data Sheets for Specific Serial Numbers

Specifications: FCC Part 95 Subpart I

Date Received in Laboratory: January 4, 2011

Compliance Status: Complies

Exclusions: None

Non-compliances: None

FCC ID: Y8L-MN0100 Report Number: **2011 01160678 FCC3**

1.2 Document History

REVISION	DATE	COMMENTS	
-	January 17, 2011	Prepared By:	Alan Laudani
-	January 17, 2011	Initial Release:	Alan Laudani

Note that the results contained in this report relate only to the items tested and were obtained in the period between the date of initial receipt of samples and the date of issue of the report.

This test report has been completed in accordance with the requirements of ISO/IEC 17025.

Nemko USA Inc. authorizes the applicant to reproduce this report provided it is reproduced in its entirety and for use by the company's employees only.

Any use which a third party makes of this report, or any reliance on or decisions to be made based on it, are the responsibility of such third parties. Nemko USA Inc. accepts no responsibility for damages, if any, suffered by any third party as a result of decisions made or actions based on this report.

TESTED BY:

Date: January 17, 2011

Alan Laudani, RF/EMC Engineer



www.nemko.com

TABLE OF CONTENTS

1.	Section1: Summary of Test Results	. 2
1.1	General	
1.2	Document History	
2.	Section 2: Equipment Under Test	
1.3	Product Identification	
1.4	Technical Specifications of the EUT	. 5
3.	Section 3: Test Conditions	. 6
3.1	Test Environment	. 6
3.2	Test Equipment	. 6
4.	Section 4: Observations	. 7
4.1	Modifications Performed During Assessment	
4.2	Record Of Technical Judgments	
4.3	EUT Parameters Affecting Compliance	. 7
4.4	Test Deleted	. 7
4.5	Additional Observations	. 7
5.	Section 5: Results Summary	. 8
5.1.	Test Result summary table	
6.	Appendix A: Test Results	. 9
A1.	Frequency Monitoring - Not Required	
A2.	Frequency vs Temperature	
A3.	Emission Bandwidth	12
A4.	Unwanted Radiation	
A5.	Maximum Transmitter Power	
A6.	Emission Types	19

Report Number: 2011 01160678 FCC3



www.nemko.com

2. Section 2: Equipment Under Test

1.3 Product Identification

The Equipment Under Test was identified as follows:

Description	Serial No.
Trial Neurostimulator Model: MN0100	See Data Sheets for Specific Serial Numbers

1.4 Technical Specifications of the EUT

Manufacturer: Spinal Modulation Inc.

Transmit Frequency: 402.150 MHz to 404.850 MHz

Rated Power: 0.133 µW

Modulation: 2FSK

Emission Designator: 244KFID

Antenna: helix type antenna

Antenna Connector: Integral to circuitry

Power Source: BATTERY

Report Number: 2011 01160678 FCC3

www.nemko.com

3. Section 3: Test Conditions

3.1 Test Environment

All tests were performed under the following environmental conditions:

Temperature range : 21-31 °C Humidity range : 18-70 % Pressure range : 101.2 kPa

Power supply range : 102-132 Vac 60 Hz

3.2 Test Equipment

Nemko						Cal Due
ID	Device	Manufacturer	Model	Serial Number	Cal Date	Date
111	Antenna, LPA	EMCO	3146	1382	11/29/2010	11/29/2012
128	Antenna, Bicon	EMCO	3104	2882	2/9/2009	2/9/2011
317	Preamplifier	HP	8449A	2749A00167	5/7/2010	5/7/2011
752	Antenna, DRWG	EMCO	3115	4943	12/2/2010	12/2/2012
835	Spectrum Analyzer	Rohde & Schwarz	RHDFSEK	829058/005	7/12/2010	7/12/2011
836	Signal Generator	Agilent	E8254A	US41140229	2/5/2010	2/5/2011
815	Multimeter	Fluke	111	78130066	8/4/2010	8/4/2011
877	Antenna, DRG Horn, .7- 18GHz	AH Systems	SAS-571	688	8/16/2010	8/16/2011
919	Preamplifier	Spacek Labs MM-Wave Technology	100MHz to 40GHz	3M12 (SLK-35- 3) and 3M13 (SLKa-35-4)	12/14/2010	12/14/2011
N149	Environmental Chamber	Cincinnati Sub- Zero	ZPHS-32-2-2- H/AC	ZP0552665	6/22/2010	6/22/2011
926	UWave Freq Counter	Anritsu	MF2512B	6200229301	2-Mar-10	2-Mar-11
E1013	DRG Horn (Small)	EMCO	3116	00119488	12/23/2009	12/23/2011
E1018	9kHz to 7GHz Spectrum Analyzer	Rohde & Schwarz	FSP7	835363/0003	1/22/2010	1/22/2011
911	Spectrum Analyzer	Agilent	E4440A	US41421266	10/26/2010	10/26/2011
client	DC Power Supply	Gwinstek	GPS-30300	NA	NCR	NCR
NA	20 dB Attenuator	Winschel	24-20-234	NA	Verified	Verified

NVLAP LAB CODE: 200116-0.

Registration of the OATS are on file with the Federal Communications Commission, under the VCCI under registration number R-3027, and are also registered with Industry Canada under Site Numbers 2040B-1 and 2040B-2.



FCC ID: Y8L-MN0100 Report Number: **2011 01160678 FCC3**

www.nemko.com

4. Section 4: Observations

4.1 Modifications Performed During Assessment None

4.2 Record Of Technical JudgmentsNo technical judgments were made during the assessment.

4.3 EUT Parameters Affecting Compliance The user of the apparatus could not alter parameters that would affect compliance.

4.4 Test DeletedNo Tests were deleted from this assessment.

4.5 Additional Observations There were no additional observations made during this assessment.

Report Number: 2011 01160678 FCC3

www.nemko.com

FCC ID: Y8L-MN0100

5. Section 5: Results Summary

5.1. Test Result summary table

FCC Part 95 Subpart I:

The column headed "Required" indicates whether the associated clauses were invoked for the apparatus under test. The following abbreviations are used:

- No: not applicable / not relevant
- Y Yes: Mandatory i.e. the apparatus shall conform to these tests.
- N/T Not Tested, mandatory but not assessed. (See section 4.4 Test deleted)

The results contained in this section are representative of the operation of the apparatus as originally submitted.

FCC	Test/Requirement Description	Required	Result
95.628 (a)	Frequency Monitoring	N	* NR
95.628 (e)(2)	Frequency vs Temperature	Υ	Complies
95.628 (a) (6) (i); 95.633 (e) (3)	Emission Bandwidth	Υ	Complies
95.635 (d)	Unwanted Radiation	Υ	Complies
95.628 (c)(4)(iii)	Maximum Transmitter Power	Υ	Complies
95.631 (h)	Emission Types	Y	Complies

^{*} Not Required

FCC ID: Y8L-MN0100 Report Number: **2011 01160678 FCC3**

6. Appendix A: Test Results

A1. Frequency Monitoring - Not Required

95.628 (a) (a) Frequency monitoring. Except as provided in (b) of this section, all MedRadio programmer/control transmitters operating in the 401–406 MHz band must operate under the control of a monitoring system that incorporates a mechanism for monitoring the channel or channels that the MedRadio system devices intend to occupy. The monitoring system antenna shall be the antenna normally used by the programmer/control transmitter for a communications session. Before the monitoring system of a MedRadio programmer/control transmitter initiates a MedRadio communications session, the following access criteria must be met:

- (1) The monitoring system bandwidth measured at its 20 dB down points must be equal to or greater than the emission bandwidth of the intended transmission.
- (2) Within 5 seconds prior to initiating a communications session, circuitry associated with a MedRadio programmer/control transmitter must monitor the channel or channels the system devices intend to occupy for a minimum of 10 milliseconds per channel.
- (3) Based on use of an isotropic monitoring system antenna, the monitoring threshold power level must not be more than 10logB(Hz) –150 (dBm/Hz) + G(dBi), where B is the emission bandwidth of the MedRadio communications session transmitter having the widest emission and G is the MedRadio programmer/control transmitter monitoring system antenna gain relative to an isotropic antenna. For purposes of showing compliance with the above provision, the above calculated threshold power level must be increased or decreased by an amount equal to the monitoring system antenna gain above or below the gain of an isotropic antenna, respectively.
- (4) If no signal in a MedRadio channel above the monitoring threshold power level is detected, the MedRadio programmer/control transmitter may initiate a MedRadio-communications session involving transmissions to and from a medical implant or medical body-worn device on that channel. The MedRadio communications session may continue as long as any silent period between consecutive data transmission bursts does not exceed 5 seconds. If a channel meeting the criteria in paragraph (a)(3) of this section is unavailable, MedRadio transmitters that are capable of operating on multiple channels may transmit on the alternate channel accessible by the device with the lowest monitored ambient power level. Except as provided in paragraph (b) of this section, MedRadio transmitters that operate on a single channel and thus do not have the capability of operating on alternate channels may not transmit unless no signal on the single channel of operation exceeds the monitoring threshold power level.
- (5) When a channel is selected prior to a MedRadio communications session, it is permissible to select an alternate channel for use if communications are interrupted, provided that the alternate channel selected is the next best choice using the above criteria. The alternate channel may be accessed in the event a communications session is interrupted by interference. The following criteria must be met:
- (i) Before transmitting on the alternate channel, the channel must be monitored for a period of at least 10 milliseconds.
- (ii) The detected power level during this 10 millisecond or greater monitoring period must be no higher than 6dB above the power level detected when the channel was chosen as the alternate channel.
- (iii) In the event that this alternate channel provision is not used by the MedRadio system or if the criteria in paragraphs (a)(5)(i) and (ii) are not met, a channel must be selected using the access criteria specified in paragraphs (a)(1) through (a)(4) of this section.
- (6) As used in this section, the following definitions apply:
- (i) *Emission bandwidth* Measured as the width of the signal between the points on either side of carrier center frequency that are 20 dB down relative to the maximum level of the modulated carrier. Compliance will be determined using instrumentation employing a peak detector function and a resolution bandwidth approximately equal to 1% of the emission bandwidth of the device under test.
- (ii) MedRadio channel—Any continuous segment of spectrum in the MedRadio band that is equal to the emission bandwidth of the device with the largest bandwidth that is to participate in a MedRadio communications session. Note to paragraph (a)(6)(ii): The rules do not specify a channeling scheme for use by MedRadio systems.
- (iii) *MedRadio communications session* —A collection of transmissions, that may or may not be continuous, between MedRadio system devices.

Report Number: 2011 01160678 FCC3



www.nemko.com

A2. Frequency vs Temperature

95.628 (e) (2)

(e) Frequency stability. Each transmitter in the MedRadio service must maintain a frequency stability of ±100 ppm of the operating frequency over the range:

(2) 0 °C to 55 °C in the case of MedRadio programmer/control transmitters and MedRadio body-worn transmitters.

Conditions:

Model:	MN0100	Temperature:	19°C
Date:	1-11-2011	Humidity:	33%
Modification State:	None	Tester:	Alan Laudani
		Laboratory:	Nemko

Observations:

Method of Measurement:

Modulation: CW.

Spectrum Analyzer settings: 3 kHz RBW, 10 kHz VBW and/or use of frequency counter.

Direct connection to the spectrum analyzer (or frequency counter) was used for measurements using the test boards.

The Neurostimulator Circuit Board, model AD1518 Rev 1, SN 200849 was used as a representative sample of the Neurostimulator: Implant MNO200, Trial MN0100.

Test Conditions: Ambient Temperature: 19°C

Relative Humidity: 33%

Measurement Data: Table below.

Limits: The frequency shall remain within 100 ppm of the channel frequency.

 \pm 100ppm x 405 MHz = \pm 40,500 Hz

EUT complies



FCC ID: Y8L-MN0100

Report Number: 2011 01160678 FCC3

www.nemko.com

Neurostimulator

Voltage Input	Frequency (MHz)	Frequency Delta (Hz)
2.72 VDC	402.14236	-7640
2.52 VDC	402.14235	-7650
2.31 VDC	402.14235	-7650
2.13 VDC	402.14237	-7630
2.03 VDC	402.14235	-7650
1.92 VDC	402.14236	-7640
1.81 VDC	402.14235	-7650
1.70 VDC	402.14236	-7640
1.61 VDC	402.14235	-7650
1.51 VDC	402.14237	-7630
1.41 VDC	402.14237	-7630
1.30 VDC	402.14237	-7630
1.20 VDC	402.14235	-7650
1.15 VDC	402.14235	-7650
1.11 VDC	402.14236	-7640
1.10 VDC	OFF	

Test Condition	Frequency (MHz)	Frequency Delta (Hz)
20°C	402.141600	-8400
30°C	402.138800	-11200
40°C	402.135543	-14457
50°C	402.131670	-18330
55°C	402.129690	-20310
-10°C	402.148370	-1630
0°C	402.147670	-2330
10°C	402.145190	-4810
20°C	402.142580	-7420

Report Number: 2011 01160678 FCC3

www.nemko.com

A3. **Emission Bandwidth**

95.628 (a) (6) (i) Emission bandwidth — Measured as the width of the signal between the points on either side of carrier center frequency that are 20 dB down relative to the maximum level of the modulated carrier. Compliance will be determined using instrumentation employing a peak detector function and a resolution bandwidth approximately equal to 1% of the emission bandwidth of the device under test.

95.633 (e) (3) Emission bandwidth will be determined by measuring the width of the signal between points, one below the carrier center frequency and one above the carrier center frequency, that are 20 dB down relative to the maximum level of the modulated carrier. Compliance with the emission bandwidth limit is based on the use of measurement instrumentation employing a peak detector function with an instrument resolution bandwidth approximately equal to 1.0 percent of the emission bandwidth of the device under measurement.

Conditions:

Model:	MN0100	Temperature:	19°C
Date:	1-11-2011	Humidity:	33%
Modification State:	Normal	Tester:	Alan Laudani
		Laboratory:	Nemko

Observations:

Peak, max hold emission, continuous test mode.

The Neurostimulator Circuit Board, model AD1518 Rev 1, SN 200849 was used as a representative sample of the Neurostimulator: Implant MNO200, Trial MN0100.


Test Results: Complies

Test Data: See attached plots.


FCC ID: Y8L-MN0100

Report Number: 2011 01160678 FCC3

TNS



Date: 1.JAN.1997 05:37:01



www.nemko.com

Report Number: 2011 01160678 FCC3



www.nemko.com

A4. Unwanted Radiation

Para. No.: 95.635 (d)

(d) For transmitters designed to operate in the MedRadio service, emissions shall be attenuated in accordance with the following: (paragraphs (d)(1) through (d)(5) pertain to MedRadio transmitters operating in the 402–405 MHz band; paragraphs (d)(6) through (d)(10) pertain to MedRadio transmitters operating in the 401–402 MHz or 405–406 MHz bands).

(1) Emissions from a MedRadio transmitter more than 250 kHz outside of the 402–405 MHz band shall be

attenuated to a level no greater than the following field strength limits:

Frequency (MHz)	Field strength (μV/m)	Measurement distance (m)
30–88	100	3
88–216	150	3
216–960	200	3
960 and above	500	3

Note—At band edges, the tighter limit applies.

(2) The emission limits shown in the table of paragraph (d)(1) are based on measurements employing a CISPR quasi-peak detector except that above 1 GHz, the limit is based on measurements employing an average detector. Measurements above 1 GHz shall be performed using a minimum resolution bandwidth of 1 MHz. See also §95.605.

(3) The emissions from a MedRadio transmitter must be measured to at least the tenth harmonic of the highest fundamental frequency designed to be emitted by the transmitter.

Conditions:

Model:	MN0100	Temperature:	13°C
Date:	1-12-2011	Humidity:	36%
Modification State:	Normal	Tester:	Alan Laudani
		Laboratory:	Nemko

Observations: Peak hold detection worst case over quasi-peak detector.

Test Results: Complies

Test Data: See Data Sheet Below

Report Number: 2011 01160678 FCC3

www.nemko.com

Trial NeuroStimulator transmitter



NEMKO USA, Inc.

San Diego Headquarters:

11696 Sorrento Valley Rd. San Diego, CA 92121 Tel: (858) 755-5525 Fax: (858) 452-1810

Radiated Emissions Data

Job # : 43259-1 Date : 1-12-2011 Page 1 of 1
NEX #: 160678 Time : 1130 NOATS

Staff: aal SOATS X

Client Name : Spinal Modulation, Inc. Distance < 1000 MHz: 3 m

EUT Name : Trial NeuroStimulator Distance > 1000 MHz: 3 m

EUT Model #: MN0100
EUT Serial #: CB0731

EUT Config. : Continuous Transmit

NA Loop Ant. #: 114_3m Bicon Ant.#: Temp. (°C): 16 Log Ant.#: 111 3m Humidity (%): DRG Ant. # Spec Analyzer #: NA E1017 Cable LF#: Analyzer Display #: E1017 soats

Cable LF#: soats Analyzer Display #: E1017
Cable HF#: 877 Quasi-Peak Detector #: E1017
Preamp LF#: 902 Preselector #: E1017
Preamp HF# NA

Quasi-Peak RBW: 120 kHz
Video Bandw idth 300 kHz
Peak RBW: 300 kHz
Video Bandw idth 1 MHz

Meas.	Meter	Meter	Det.	EUT	Ant.	Max.	Corrected	Spec.	CR/SL	Pass	
Freq.	Reading	Reading		Side	Height	Reading	Reading	limit	Diff.	Fail	
(MHz)	Vertical	Horizontal		F/L/R/B	m	(dBµV)	(dBµV/m)	(dBµV/m)	(dB)		Comment
401.75	15.8	17.0	Q		2.5	17.0	36.1	46.0	-9.9	Pass	low er band edge
402.15	34.9	37.7	Р		2.5	37.7	56.8	58.3	-1.5	Pass	Channel 0
404.85	35.2	38.5	Р		2.4	38.5	57.6	58.3	-0.7	Pass	Channel 9
405.25	16.1	16.8	Q		2.4	14.6	33.7	46.0	-12.3	Pass	upper band edge

		Substitu	tion Metho						
Tar	get	Dipole	Cable	Signal	Total	Total	Spec	Margin	1
Frequency	Level	ant.	loss	Generator	(EIRP)	(EIRP)			
MHz	dBuV/m	gain	dB	dBm	dBm	μW	μW	dB	
402.15	37.7	2.15	5.0	-37.5	-40.4	0.092	0.200	-3.4	
404.85	38.5	2.15	5.0	-35.9	-38.8	0.133	0.200	-1.8	

NOTE: No other emissions detected over frequency range for 30 MHz to 4.1 GHz




Report Number: 2011 01160678 FCC3

Radiated Band Edge


Trial NeuroStimulator

Channel 0, 250 kHz below 402 MHz



Date: 1.JAN.1997 04:02:57

Channel 9, 250 kHz above 405 MHz



Date: 1.JAN.1997 04:18:35

Report Number: 2011 01160678 FCC3



www.nemko.com

A5. Maximum Transmitter Power

Para. No.: 95.628 (c)(4)(iii)

(c) Operating frequency. MedRadio stations authorized under this part may operate on frequencies in the 401–406 MHz band as follows provided that the out-of-band emissions are attenuated in accordance with §95.635:

(4) MedRadio stations that are used externally to evaluate the efficacy of a more permanent medical implant device, regardless of whether a frequency monitoring system as set forth in paragraph (a) of this section is employed, may operate on any of the frequencies in the 402–405 MHz band, provided that:

(iii) The maximum output power of the temporary body-worn device shall not exceed 200 nW EIRP;

Conditions:

Model:	MN0100	Temperature:	13°C
Date:	1/12/2011	Humidity:	36%
Modification State:	Normal	Tester:	Alan Laudani
		Laboratory:	Nemko

Observations:

Field strength was substituted for to result in output power EIRP.

Test Results: Passed

The maximum field strength is 0.133 µW (133nW).

Test Data: See attached tables

www.nemko.com

FCC ID: Y8L-MN0100

Report Number: 2011 01160678 FCC3



San Diego Headquarters:

11696 Sorrento Valley Rd. San Diego, CA 92121 Tel: (858) 755-5525 Fax: (858) 452-1810

NEMKO USA, Inc.

Radiated Emissions Data

Job #: 43259-1 Date: 1-12-2011 Page of Time: 1130 NEX#: 160678 **NOATS**

Staff: aal SOATS Χ Client Name: Distance < 1000 MHz: Spinal Modulation, Inc. 3 m Trial NeuroStimulator EUT Name: Distance > 1000 MHz: 3 m

EUT Model #: MN0100 EUT Serial #: CB0731

EUT Config. : Continuous Transmit

Loop Ant. #: NA 114 3m Bicon Ant.#: Temp. (°C): Humidity (%): 36
Spec Analyzer #: E1017 111_3m Log Ant.#: DRG Ant. # NA Analyzer Display #: _ Cable LF#: soats E1017 Cable HF#: 877 Quasi-Peak Detector #: E1017

902 Preamp LF#: Preselector #: E1017 Preamp HF# NA

Quasi-Peak RBW: 120 kHz Video Bandw idth 300 kHz RBW: 300 kHz Peak Video Bandwidth 1 MHz

Meas.	Meter	Meter	Det.	EUT	Ant.	Max.	Corrected	Spec.	CR/SL	Pass	
Freq.	Reading	Reading		Side	Height	Reading	Reading	limit	Diff.	Fail	
(MHz)	Vertical	Horizontal		F/L/R/B	m	(dBµV)	(dBµV/m)	(dBµV/m)	(dB)		Comment
401.75	15.8	17.0	Q		2.5	17.0	36.1	46.0	-9.9	Pass	low er band edge
402.15	34.9	37.7	Р		2.5	37.7	56.8	58.3	-1.5	Pass	Channel 0
404.85	35.2	38.5	Р		2.4	38.5	57.6	58.3	-0.7	Pass	Channel 9
405.25	16.1	16.8	Q		2.4	14.6	33.7	46.0	-12.3	Pass	upper band edge

	Substitution Method For Radiated Emissions							
Tar	get	Dipole	Cable	Signal	Total	Total	Spec	Margin
Frequency		ant.	loss	Generator	(EIRP)	(EIRP)		. 3
MHz	dBuV/m	gain	dB	dBm	dBm	μW	μW	dB
402.15	37.7	2.15	5.0	-37.5	-40.4	0.092	0.200	-3.4
404.85	38.5	2.15	5.0	-35.9	-38.8	0.133	0.200	-1.8
	_		_					



Report Number: 2011 01160678 FCC3



www.nemko.com

A6. Emission Types

Para. No.: 95.631 (h)


(h) A MedRadio station may transmit any emission type appropriate for communications in this service. Voice communications, however, are prohibited.

Conditions:

Model:	MN0100	Temperature:	13°C
Date:	10/11/2010	Humidity:	35%
Modification State:	Normal	Tester:	Alan Laudani
		Laboratory:	Nemko

Observations: Not a voice communication.

Test Results: Passed



Date: 1.JAN.1997 04:18:35